




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032021\
Data File : BN013991.D
Acq On : 19 Mar 2021 09:46
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 19 11:10:48 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 05 13:12:06 2021
Response via : Initial Calibration



TIC: BN013991.D\data.ms

(77) Benzidine




19.328min (-0.017) 0.00 ng

response 1545872

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.47
183.00	13.20	11.52
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032021\
Data File : BN013991.D
Acq On : 19 Mar 2021 09:46
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 19 11:10:48 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 05 13:12:06 2021
Response via : Initial Calibration



TIC: BN013991.D\data.ms

(70) Pentachlorophenol (C)

16.728min (-0.023) 19721.95 ng

response 229683

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	55.85
264.00	61.60	62.94
0.00	0.00	0.00