

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032021\
 Data File : BN013999.D
 Acq On : 19 Mar 2021 14:52
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN032021

Quant Time: Mar 19 15:28:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	110	0.00
2	1,4-Dioxane	0.400	0.430	-7.5	116	0.00
3	n-Nitrosodimethylamine	0.400	0.416	-4.0	128	0.00
4 S	2-Fluorophenol	0.400	0.435	-8.7	112	0.00
5 S	Phenol-d6	0.400	0.427	-6.7	111	0.00
6	bis(2-Chloroethyl)ether	0.400	0.429	-7.2	111	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	110	0.00
8 S	Nitrobenzene-d5	0.400	0.392	2.0	107	0.00
9	Naphthalene	0.400	0.415	-3.7	111	-0.01
10	Hexachlorobutadiene	0.400	0.420	-5.0	112	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.410	-2.5	111	0.00
12	2-Methylnaphthalene	0.400	0.409	-2.2	110	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.400	0.316	21.0	91	0.00
15 S	2-Fluorobiphenyl	0.400	0.420	-5.0	112	0.00
16	Acenaphthylene	0.400	0.389	2.8	111	0.00
17	Acenaphthene	0.400	0.411	-2.7	112	0.01
18	Fluorene	0.400	0.402	-0.5	108	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.379	5.3	112	0.00
21	4-Bromophenyl-phenylether	0.400	0.377	5.8	99	0.00
22	Hexachlorobenzene	0.400	0.417	-4.2	109	0.00
23	Atrazine	0.400	0.367	8.3	103	0.00
24	Pentachlorophenol	0.400	0.351	12.3	109	0.00
25	Phenanthrene	0.400	0.407	-1.7	108	0.00
26	Anthracene	0.400	0.390	2.5	108	0.00
27 SURR	Fluoranthene-d10	0.400	0.401	-0.3	108	0.00
28	Fluoranthene	0.400	0.404	-1.0	109	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	112	0.00
30	Pyrene	0.400	0.404	-1.0	109	0.00
31 S	Terphenyl-d14	0.400	0.406	-1.5	108	0.00
32	Benzo(a)anthracene	0.400	0.411	-2.7	112	0.01
33	Chrysene	0.400	0.422	-5.5	114	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.437	-9.2	129	0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.429	-7.2	116	0.00
36 I	Perylene-d12	0.400	0.400	0.0	113	0.00
37	Benzo(b)fluoranthene	0.400	0.403	-0.8	112	0.00
38	Benzo(k)fluoranthene	0.400	0.412	-3.0	112	0.00
39 C	Benzo(a)pyrene	0.400	0.422	-5.5	118	0.00
40	Dibenzo(a,h)anthracene	0.400	0.418	-4.5	116	0.00
41	Benzo(g,h,i)perylene	0.400	0.418	-4.5	116	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN032021