

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032021\
 Data File : BN013998.D
 Acq On : 19 Mar 2021 13:57
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Mar 19 14:27:19 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 12:16:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	7508	0.40	ng	0.00
7) Naphthalene-d8	10.480	136	25540	0.40	ng	0.00
13) Acenaphthene-d10	14.334	164	14526	0.40	ng	0.00
19) Phenanthrene-d10	17.068	188	30452	0.40	ng	0.00
29) Chrysene-d12	21.250	240	33309	0.40	ng	# 0.00
36) Perylene-d12	23.458	264	27440	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	81422	4.04	ng	0.00
5) Phenol-d6	6.871	99	108041	4.03	ng	0.00
8) Nitrobenzene-d5	8.845	82	88011	5.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	198175	4.80	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	28521	4.80	ng	0.00
15) 2-Fluorobiphenyl	12.951	172	256394	4.99	ng	0.00
27) Fluoranthene-d10	19.099	212	442133	4.92	ng	0.00
31) Terphenyl-d14	19.700	244	339157	4.61	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.247	88	41768	4.60	ng	99
3) n-Nitrosodimethylamine	3.553	42	36609	5.23	ng	95
6) bis(2-Chloroethyl)ether	7.129	93	88396	4.65	ng	96
9) Naphthalene	10.530	128	305303	4.86	ng	99
10) Hexachlorobutadiene	10.822	225	56099	4.64	ng	# 100
12) 2-Methylnaphthalene	12.142	142	208467	4.72	ng	99
16) Acenaphthylene	14.042	152	312226	4.81	ng	99
17) Acenaphthene	14.397	154	204979	5.08	ng	98
18) Fluorene	15.374	166	259783	4.99	ng	100
21) 4-Bromophenyl-phenylether	16.264	248	76724	4.63	ng	91
22) Hexachlorobenzene	16.386	284	90015	5.20	ng	100
23) Atrazine	16.532	200	63180	4.08	ng	# 92
25) Phenanthrene	17.104	178	417085	5.26	ng	100
26) Anthracene	17.201	178	375223	4.84	ng	100
28) Fluoranthene	19.129	202	477340	5.02	ng	99
30) Pyrene	19.491	202	486676	4.75	ng	100
32) Benzo(a)anthracene	21.229	228	496594	4.84	ng	97
33) Chrysene	21.282	228	494291	4.94	ng	99
34) Bis(2-ethylhexyl)phtha...	21.165	149	202014	4.15	ng	99
35) Indeno(1,2,3-cd)pyrene	25.676	276	547287	4.51	ng	100
37) Benzo(b)fluoranthene	22.798	252	504351	5.75	ng	# 91
38) Benzo(k)fluoranthene	22.842	252	483418	5.71	ng	# 90
39) Benzo(a)pyrene	23.363	252	441066	5.59	ng	# 87
40) Dibenzo(a,h)anthracene	25.690	278	462688	5.53	ng	96
41) Benzo(g,h,i)perylene	26.354	276	475102	5.62	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032021\
Data File : BN013998.D
Acq On : 19 Mar 2021 13:57
Operator : CG/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Mar 19 14:27:19 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 19 12:16:37 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.700 min Scan# 611
 Delta R.T. 0.000 min
 Lab File: BN013998.D
 Acq: 19 Mar 2021 13:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:152 Resp: 7508
 Ion Ratio Lower Upper
 152 100
 150 158.5 121.8 182.8
 115 61.5 46.6 69.8



#2
 1,4-Dioxane
 Concen: 4.60 ng
 RT: 3.247 min Scan# 58
 Delta R.T. 0.000 min
 Lab File: BN013998.D
 Acq: 19 Mar 2021 13:57

Tgt Ion: 88 Resp: 41768
 Ion Ratio Lower Upper
 88 100
 58 79.7 64.6 96.8
 43 32.9 26.6 40.0









#3
n-Nitrosodimethylamine
Concen: 5.23 ng
RT: 3.553 min Scan# 96
Delta R.T. -0.024 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	42	Resp:	36609
Ion	Ratio	Lower	Upper
42	100		
74	113.4	95.0	142.6
44	5.9	5.8	8.8



#4
2-Fluorophenol
Concen: 4.04 ng
RT: 5.300 min Scan# 313
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:	112	Resp:	81422
Ion	Ratio	Lower	Upper
112	100		
64	59.5	44.3	66.5
63	31.7	23.0	34.6









#5
Phenol-d6
Concen: 4.03 ng
RT: 6.871 min Scan# 508
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 108041
Ion Ratio Lower Upper
99 100
42 22.0 15.2 22.8
71 33.2 26.1 39.1



#6
bis(2-Chloroethyl)ether
Concen: 4.65 ng
RT: 7.129 min Scan# 540
Delta R.T. -0.008 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion: 93 Resp: 88396
Ion Ratio Lower Upper
93 100
63 80.6 60.8 91.2
95 32.8 26.0 39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.480 min Scan# 858
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	25540
Ion Ratio	Lower	Upper	
136	100		
137	11.1	9.0	13.6
54	6.7	6.4	9.6
68	5.3	4.5	6.7



#8
Nitrobenzene-d5
Concen: 5.37 ng
RT: 8.845 min Scan# 729
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:	82	Resp:	88011
Ion Ratio	Lower	Upper	
82	100		
128	38.9	39.2	58.8#
54	52.1	40.1	60.1









#9
Naphthalene
Concen: 4.86 ng
RT: 10.530 min Scan# 862
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 305303
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.4
127 12.5 10.4 15.6



#10
Hexachlorobutadiene
Concen: 4.64 ng
RT: 10.822 min Scan# 885
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:225 Resp: 56099
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 4.80 ng
RT: 12.071 min Scan# 1066
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 198175
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2



#12
2-Methylnaphthalene
Concen: 4.72 ng
RT: 12.142 min Scan# 1082
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:142 Resp: 208467
Ion Ratio Lower Upper
142 100
141 88.9 70.5 105.7
115 35.2 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.334 min Scan# 1308
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 14526
Ion Ratio Lower Upper
164 100
162 99.1 79.6 119.4
160 43.6 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 4.80 ng
RT: 15.818 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:330 Resp: 28521
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 36.5 26.8 40.2









#15
2-Fluorobiphenyl
Concen: 4.99 ng
RT: 12.951 min Scan# 1199
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 256394
Ion Ratio Lower Upper
172 100
171 34.1 28.6 42.8
170 22.8 18.9 28.3



#16
Acenaphthylene
Concen: 4.81 ng
RT: 14.042 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:152 Resp: 312226
Ion Ratio Lower Upper
152 100
151 19.2 15.7 23.5
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 5.08 ng
RT: 14.397 min Scan# 1313
Delta R.T. 0.013 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 204979
Ion Ratio Lower Upper
154 100
153 108.1 88.4 132.6
152 52.0 43.0 64.4



#18
Fluorene
Concen: 4.99 ng
RT: 15.374 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:166 Resp: 259783
Ion Ratio Lower Upper
166 100
165 97.4 77.8 116.8
167 13.5 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.068 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188	Resp:	30452
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	9.7	13.6



#21
4-Bromophenyl-phenylether
Concen: 4.63 ng
RT: 16.264 min Scan# 1461
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:248	Resp:	76724
Ion Ratio	Lower	Upper
248	100	
250	90.2	75.5
141	79.3	53.8









#22
Hexachlorobenzene
Concen: 5.20 ng
RT: 16.386 min Scan# 1471
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:284 Resp: 90015
Ion Ratio Lower Upper
284 100
142 34.7 27.9 41.9
249 30.1 24.2 36.2



#23
Atrazine
Concen: 4.08 ng
RT: 16.532 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:200 Resp: 63180
Ion Ratio Lower Upper
200 100
173 29.5 18.6 28.0#
215 46.7 41.0 61.6









#25
 Phenanthrene
 Concen: 5.26 ng
 RT: 17.104 min Scan# 1530
 Delta R.T. 0.000 min
 Lab File: BN013998.D
 Acq: 19 Mar 2021 13:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:	178	Resp:	417085
Ion Ratio	Lower	Upper	
178	100		
176	18.8	15.0	22.4
179	15.3	12.3	18.5



#26
 Anthracene
 Concen: 4.84 ng
 RT: 17.201 min Scan# 1538
 Delta R.T. 0.000 min
 Lab File: BN013998.D
 Acq: 19 Mar 2021 13:57

Tgt Ion:	178	Resp:	375223
Ion Ratio	Lower	Upper	
178	100		
176	18.2	14.6	21.8
179	15.4	12.2	18.2









#27
Fluoranthene-d10
Concen: 4.92 ng
RT: 19.099 min Scan# 1823
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:212 Resp: 442133
Ion Ratio Lower Upper
212 100
106 12.6 9.7 14.5
104 7.0 5.5 8.3



#28
Fluoranthene
Concen: 5.02 ng
RT: 19.129 min Scan# 1829
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:202 Resp: 477340
Ion Ratio Lower Upper
202 100
101 10.9 8.1 12.1
203 17.4 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:240 Resp: 33309
Ion Ratio Lower Upper
240 100
120 7.1 10.2 15.4#
236 25.6 21.3 31.9



#30
Pyrene
Concen: 4.75 ng
RT: 19.491 min Scan# 1902
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:202 Resp: 486676
Ion Ratio Lower Upper
202 100
200 20.5 16.3 24.5
203 17.6 14.0 21.0









#31
 Terphenyl-d14
 Concen: 4.61 ng
 RT: 19.700 min Scan# 1944
 Delta R.T. 0.000 min
 Lab File: BN013998.D
 Acq: 19 Mar 2021 13:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:244 Resp: 339157
 Ion Ratio Lower Upper
 244 100
 212 7.4 6.1 9.1
 122 11.2 9.8 14.8



#32
 Benzo(a)anthracene
 Concen: 4.84 ng
 RT: 21.229 min Scan# 2121
 Delta R.T. 0.000 min
 Lab File: BN013998.D
 Acq: 19 Mar 2021 13:57

Tgt Ion:228 Resp: 496594
 Ion Ratio Lower Upper
 228 100
 226 27.2 20.5 30.7
 229 19.1 16.2 24.4









#33
Chrysene
Concen: 4.94 ng
RT: 21.282 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 494291
Ion Ratio Lower Upper
228 100
226 29.6 24.3 36.5
229 19.3 15.1 22.7



#34
Bis(2-ethylhexyl)phthalate
Concen: 4.15 ng
RT: 21.165 min Scan# 2115
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:149 Resp: 202014
Ion Ratio Lower Upper
149 100
167 27.1 22.1 33.1
279 3.8 3.5 5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 4.51 ng
RT: 25.676 min Scan# 3187
Delta R.T. 0.003 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:276 Resp: 547287
Ion Ratio Lower Upper
276 100
138 24.3 19.8 29.6
277 24.8 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.458 min Scan# 2539
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:264 Resp: 27440
Ion Ratio Lower Upper
264 100
260 24.4 19.3 28.9
265 31.6 66.2 99.4#









#37
Benzo(b)fluoranthene
Concen: 5.75 ng
RT: 22.798 min Scan# 2346
Delta R.T. 0.003 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 504351
Ion Ratio Lower Upper
252 100
253 21.9 20.4 30.6
125 9.6 11.6 17.4#



#38
Benzo(k)fluoranthene
Concen: 5.71 ng
RT: 22.842 min Scan# 2359
Delta R.T. 0.003 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:252 Resp: 483418
Ion Ratio Lower Upper
252 100
253 21.9 20.5 30.7
125 9.6 12.3 18.5#









#39
Benzo(a)pyrene
Concen: 5.59 ng
RT: 23.363 min Scan# 2511
Delta R.T. 0.000 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 441066
Ion Ratio Lower Upper
252 100
253 22.0 21.8 32.6
125 10.7 14.6 21.8#



#40
Dibenzo(a,h)anthracene
Concen: 5.53 ng
RT: 25.690 min Scan# 3191
Delta R.T. 0.003 min
Lab File: BN013998.D
Acq: 19 Mar 2021 13:57

Tgt Ion:278 Resp: 462688
Ion Ratio Lower Upper
278 100
139 15.8 13.1 19.7
279 23.6 21.4 32.0





#41
 Benzo(g,h,i)perylene
 Concen: 5.62 ng
 RT: 26.354 min Scan# 3385
 Delta R.T. 0.003 min
 Lab File: BN013998.D
 Acq: 19 Mar 2021 13:57

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:276 Resp: 475102

Ion	Ratio	Lower	Upper
276	100		
277	23.7	20.0	30.0
138	20.5	17.8	26.8

