

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032024\
 Data File : BN030432.D
 Acq On : 20 Mar 2024 16:29
 Operator : MA/JU
 Sample : SSTD00520
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD005246

Quant Time: Mar 21 00:01:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 21 00:01:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.728	152	396118	20.000	ng/u1	0.00
20) Naphthalene-d8	10.516	136	1735159	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.369	164	1089422	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.116	188	2341890	20.000	ng/u1	0.00
79) Chrysene-d12	21.363	240	2095851	20.000	ng/u1	0.00
88) Perylene-d12	24.333	264	2270380	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.223	96	20634	1.758	ng/uL	0.00
4) Pyridine-d5	3.622	84	133910	4.779	ng/u1	0.00
7) Phenol-d5	6.881	99	167544	4.767	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.063	67	109406	5.482	ng/u1	0.00
11) 2-Chlorophenol-d4	7.258	132	127617	4.533	ng/u1	0.00
15) 4-Methylphenol-d8	8.422	113	129693	4.682	ng/u1	0.00
21) Nitrobenzene-d5	8.881	128	54088	3.736	ng/u1	0.00
24) 2-Nitrophenol-d4	9.599	143	41627	2.742	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.128	165	111890	4.133	ng/u1	0.00
31) 4-Chloroaniline-d4	10.646	131	201803	5.022	ng/u1	0.00
46) Dimethylphthalate-d6	13.787	166	402343	4.580	ng/u1	0.00
49) Acenaphthylene-d8	14.063	160	461648	4.639	ng/u1	0.00
54) 4-Nitrophenol-d4	14.551	143	55782	3.582	ng/u1	0.00
60) Fluorene-d10	15.363	176	347919	4.828	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.475	200	23271	1.591	ng/u1	0.00
73) Anthracene-d10	17.216	188	536963	4.722	ng/u1	0.00
81) Pyrene-d10	19.516	212	581739	4.371	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.121	264	511159	4.298	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.258	88	21258	1.887	ng/uL	92
5) Pyridine	3.640	79	140714	4.936	ng/u1	96
6) Benzaldehyde	6.875	77	96322	5.880	ng/u1	99
8) Phenol	6.911	94	173940	4.974	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.158	93	150875	5.398	ng/u1	100
12) 2-Chlorophenol	7.287	128	132499	4.574	ng/u1	99
13) 2-Methylphenol	8.158	108	128951	4.918	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.252	45	198721	7.820	ng/u1	100
16) Acetophenone	8.546	105	228165	5.125	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.528	70	117341	5.508	ng/u1	98
18) 4-Methylphenol	8.481	108	139432	5.033	ng/u1	96
19) Hexachloroethane	8.799	117	55670	4.137	ng/u1	93
22) Nitrobenzene	8.922	77	144818	3.876	ng/u1	96
23) Isophorone	9.446	82	340520	5.272	ng/u1	98
25) 2-Nitrophenol	9.628	139	49672	3.038	ng/u1	95
26) 2,4-Dimethylphenol	9.687	107	144611	4.284	ng/u1	96
27) Bis(2-Chloroethoxy)met...	9.934	93	205600	5.427	ng/u1	99
29) 2,4-Dichlorophenol	10.157	162	117159	4.309	ng/u1	99
30) Naphthalene	10.563	128	472825	4.774	ng/u1	99
32) 4-Chloroaniline	10.669	127	201001	5.143	ng/u1	98
33) Hexachlorobutadiene	10.852	225	76366	4.026	ng/u1	97
34) Caprolactam	11.422	113	43256	5.342	ng/u1	95
35) 4-Chloro-3-methylphenol	11.793	107	138401	4.498	ng/u1	96
36) 2-Methylnaphthalene	12.181	142	322855	5.005	ng/u1	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032024\
 Data File : BN030432.D
 Acq On : 20 Mar 2024 16:29
 Operator : MA/JU
 Sample : SSTD00520
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD005246

Quant Time: Mar 21 00:01:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 21 00:01:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.404	142	318862	5.000	ng/ul	93
39) 1,2,4,5-Tetrachloroben...	12.546	216	153023	4.448	ng/ul	97
40) Hexachlorocyclopentadiene	12.528	237	69431	2.694	ng/ul	97
41) 2,4,6-Trichlorophenol	12.787	196	87504	4.009	ng/ul	99
42) 2,4,5-Trichlorophenol	12.851	196	92086	3.861	ng/ul	95
43) 1,1'-Biphenyl	13.204	154	419903	4.567	ng/ul	99
44) 2-Chloronaphthalene	13.240	162	330559	4.629	ng/ul	99
45) 2-Nitroaniline	13.440	65	66348	3.223	ng/ul	90
47) Dimethylphthalate	13.834	163	415844	4.658	ng/ul	100
48) 2,6-Dinitrotoluene	13.945	165	52320	3.035	ng/ul#	92
50) Acenaphthylene	14.087	152	548188	4.735	ng/ul	99
51) 3-Nitroaniline	14.269	138	60132	3.560	ng/ul#	91
52) Acenaphthene	14.434	153	362602	4.727	ng/ul	99
53) 2,4-Dinitrophenol	14.469	184	14541	1.501	ng/ul	93
55) 4-Nitrophenol	14.563	109	50590	2.865	ng/ul	95
56) Dibenzofuran	14.769	168	488935	4.779	ng/ul	99
57) 2,4-Dinitrotoluene	14.728	165	75471	3.067	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.992	232	75499	3.933	ng/ul	98
59) Diethylphthalate	15.204	149	413221	4.331	ng/ul	99
61) Fluorene	15.422	166	410009	4.854	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.422	204	193818	4.793	ng/ul	99
63) 4-Nitroaniline	15.428	138	61293	3.919	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.487	198	28374	1.798	ng/ul	93
67) N-Nitrosodiphenylamine	15.634	169	341405	4.621	ng/ul	96
68) 4-Bromophenyl-phenylether	16.316	248	108381	4.320	ng/ul	98
69) Hexachlorobenzene	16.416	284	133274	4.850	ng/ul#	94
70) Atrazine	16.587	200	107031	4.209	ng/ul	98
71) Pentachlorophenol	16.757	266	59418	3.196	ng/ul	93
72) Phenanthrene	17.157	178	647649	4.742	ng/ul	99
74) Anthracene	17.251	178	645948	4.649	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.157	216	148860	4.258	ng/uL	95
76) Pentachlorobenzene	14.681	250	150280	4.609	ng/uL	98
77) Carbazole	17.516	167	590510	4.786	ng/ul	99
78) Di-n-butylphthalate	18.104	149	687954	4.013	ng/ul	100
80) Fluoranthene	19.180	202	701614	4.302	ng/ul	98
82) Pyrene	19.545	202	745229	4.446	ng/ul	98
83) Butylbenzylphthalate	20.469	149	224600	2.797	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.275	252	171778	3.771	ng/ul	98
85) Benzo(a)anthracene	21.345	228	738990	4.699	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.304	149	365551	3.007	ng/ul	98
87) Chrysene	21.404	228	693576	4.755	ng/ul	98
89) Di-n-octyl phthalate	22.445	149	537439	2.529	ng/ul	100
90) Benzo(b)fluoranthene	23.398	252	652905	4.362	ng/ul	100
91) Benzo(k)fluoranthene	23.457	252	668499	4.324	ng/ul	100
93) Benzo(a)pyrene	24.198	252	613768	4.332	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.674	276	557858	3.311	ng/ul	96
95) Dibenzo(a,h)anthracene	27.751	278	557991	4.062	ng/ul	99
96) Benzo(g,h,i)perylene	28.715	276	526849	3.861	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 1.887 ng/uL
 RT: 3.258 min Scan# 29
 Delta R.T. 0.006 min
 Lab File: BN030432.D
 Acq: 20 Mar 2024 16:29




Instrument :
 BNA_N
 ClientSampleId :
 SST005246


Tgt Ion:	88	Resp:	21258
Ion	Ratio	Lower	Upper
88	100		
43	30.3	21.6	32.4
58	73.3	53.5	80.3



#5
 Pyridine
 Concen: 4.936 ng/uL
 RT: 3.640 min Scan# 94
 Delta R.T. -0.000 min
 Lab File: BN030432.D
 Acq: 20 Mar 2024 16:29

Tgt Ion:	79	Resp:	140714
Ion	Ratio	Lower	Upper
79	100		
52	70.9	54.2	81.4
51	35.4	26.0	39.0









#6
Benzaldehyde
Concen: 5.880 ng/ul
RT: 6.875 min Scan# 644
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion: 77 Resp: 96322
Ion Ratio Lower Upper
77 100
105 96.5 76.7 115.1
106 96.0 75.4 113.2



#8
Phenol
Concen: 4.974 ng/ul
RT: 6.911 min Scan# 650
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion: 94 Resp: 173940
Ion Ratio Lower Upper
94 100
65 25.0 21.3 31.9
66 34.1 27.8 41.8









#10
Bis(2-Chloroethyl)ether
Concen: 5.398 ng/ul
RT: 7.158 min Scan# 692
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion: 93 Resp: 150875
Ion Ratio Lower Upper
93 100
63 74.0 59.4 89.0
95 31.8 25.4 38.0



#12
2-Chlorophenol
Concen: 4.574 ng/ul
RT: 7.287 min Scan# 714
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:128 Resp: 132499
Ion Ratio Lower Upper
128 100
64 47.3 37.3 55.9
130 32.5 25.2 37.8









#13
2-Methylphenol
Concen: 4.918 ng/ul
RT: 8.158 min Scan# 862
Delta R.T. -0.006 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:108 Resp: 128951
Ion Ratio Lower Upper
108 100
107 85.9 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 7.820 ng/ul
RT: 8.252 min Scan# 878
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion: 45 Resp: 198721
Ion Ratio Lower Upper
45 100
77 15.0 12.0 18.0
79 11.2 9.0 13.4









#16
Acetophenone
Concen: 5.125 ng/ul
RT: 8.546 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:105 Resp: 228165
Ion Ratio Lower Upper
105 100
77 77.0 62.5 93.7
51 24.8 20.0 30.0



#17
N-Nitroso-di-n-propylamine
Concen: 5.508 ng/ul
RT: 8.528 min Scan# 925
Delta R.T. -0.012 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion: 70 Resp: 117341
Ion Ratio Lower Upper
70 100
42 54.4 42.4 63.6
101 10.0 8.0 12.0
130 20.2 17.0 25.4





#18

4-Methylphenol

Concen: 5.033 ng/ul

RT: 8.481 min Scan# 918

Delta R.T. -0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :





SSTD005246

Tgt Ion:108 Resp: 139432

Ion Ratio Lower Upper

108 100

107 108.8 90.3 135.5



#19

Hexachloroethane

Concen: 4.137 ng/ul

RT: 8.799 min Scan# 971

Delta R.T. -0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion:117 Resp: 55670


Ion Ratio Lower Upper

117 100

201 84.6 74.2 111.2

199 54.4 45.9 68.9









#22
Nitrobenzene
Concen: 3.876 ng/ul
RT: 8.922 min Scan# 992
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion: 77 Resp: 144818
Ion Ratio Lower Upper
77 100
123 41.3 35.9 53.9
65 12.8 10.2 15.4



#23
Isophorone
Concen: 5.272 ng/ul
RT: 9.446 min Scan# 1081
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion: 82 Resp: 340520
Ion Ratio Lower Upper
82 100
95 6.6 5.6 8.4
138 18.0 13.5 20.3









#25
2-Nitrophenol
Concen: 3.038 ng/ul
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:139 Resp: 49672
Ion Ratio Lower Upper
139 100
65 45.7 40.2 60.4
109 28.7 23.5 35.3



#26
2,4-Dimethylphenol
Concen: 4.284 ng/ul
RT: 9.687 min Scan# 1122
Delta R.T. -0.006 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:107 Resp: 144611
Ion Ratio Lower Upper
107 100
121 53.5 40.7 61.1
122 90.8 69.3 103.9





#27

Bis(2-Chloroethoxy)methane

Concen: 5.427 ng/ul

RT: 9.934 min Scan# 1165

Delta R.T. -0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion: 93 Resp: 205600

Ion Ratio Lower Upper

93 100

95 33.5 26.6 39.8

123 10.8 8.9 13.3



#29

2,4-Dichlorophenol

Concen: 4.309 ng/ul

RT: 10.157 min Scan# 1202

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion: 162 Resp: 117159


Ion Ratio Lower Upper

162 100

164 62.9 51.0 76.4

98 37.3 30.2 45.4









#30
Naphthalene
Concen: 4.774 ng/ul
RT: 10.563 min Scan# 1271
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument : BNA_N
ClientSampleId : SST005246


Tgt Ion:	128	Resp:	472825
Ion	Ratio	Lower	Upper
128	100		
129	10.3	8.6	12.8
127	12.9	10.4	15.6



#32
4-Chloroaniline
Concen: 5.143 ng/ul
RT: 10.669 min Scan# 1289
Delta R.T. -0.006 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:	127	Resp:	201001
Ion	Ratio	Lower	Upper
127	100		
129	31.6	26.0	39.0





#33

Hexachlorobutadiene

Concen: 4.026 ng/ul

RT: 10.852 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:225 Resp: 76366

Ion Ratio Lower Upper

225 100

223 58.3 48.0 72.0

227 66.1 50.6 76.0



#34

Caprolactam

Concen: 5.342 ng/ul

RT: 11.422 min Scan# 1417

Delta R.T. -0.012 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion:113 Resp: 43256


Ion Ratio Lower Upper

113 100

55 160.0 121.4 182.0

56 116.6 95.1 142.7









#35
4-Chloro-3-methylphenol
Concen: 4.498 ng/ul
RT: 11.793 min Scan# 1480
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:107 Resp: 138401
Ion Ratio Lower Upper
107 100
144 25.9 19.9 29.9
142 79.5 60.6 90.8



#36
2-Methylnaphthalene
Concen: 5.005 ng/ul
RT: 12.181 min Scan# 1546
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:142 Resp: 322855
Ion Ratio Lower Upper
142 100
141 85.6 70.6 105.8









#37
1-Methylnaphthalene
Concen: 5.000 ng/ul
RT: 12.404 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:142 Resp: 318862
Ion Ratio Lower Upper
142 100
141 87.1 75.5 113.3
116 4.0 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.448 ng/ul
RT: 12.546 min Scan# 1608
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:216 Resp: 153023
Ion Ratio Lower Upper
216 100
214 77.7 60.7 91.1
179 21.2 17.7 26.5
108 20.9 14.7 22.1









#40
Hexachlorocyclopentadiene
Concen: 2.694 ng/ul
RT: 12.528 min Scan# 1649
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:237 Resp: 69431
Ion Ratio Lower Upper
237 100
235 64.1 53.6 80.4
272 12.7 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 4.009 ng/ul
RT: 12.787 min Scan# 1649
Delta R.T. -0.006 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:196 Resp: 87504
Ion Ratio Lower Upper
196 100
198 95.1 77.0 115.6
200 30.4 24.3 36.5





#42

2,4,5-Trichlorophenol

Concen: 3.861 ng/ul

RT: 12.851 min Scan# 1660

Delta R.T. -0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:196 Resp: 92086

Ion Ratio Lower Upper

196 100

198 94.4 80.4 120.6

200 32.0 25.1 37.7



#43

1,1'-Biphenyl

Concen: 4.567 ng/ul

RT: 13.204 min Scan# 1720

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion:154 Resp: 419903


Ion Ratio Lower Upper

154 100

153 39.4 32.3 48.5

76 13.8 11.2 16.8









#44
 2-Chloronaphthalene
 Concen: 4.629 ng/ul
 RT: 13.240 min Scan# 1726
 Delta R.T. -0.000 min
 Lab File: BN030432.D
 Acq: 20 Mar 2024 16:29

Instrument :
 BNA_N
 ClientSampleId :
 SST005246

Tgt Ion: 162 Resp: 330559




Ion	Ratio	Lower	Upper
162	100		
164	32.4	26.0	39.0
127	41.2	32.3	48.5




#45
 2-Nitroaniline
 Concen: 3.223 ng/ul
 RT: 13.440 min Scan# 1760
 Delta R.T. -0.006 min
 Lab File: BN030432.D
 Acq: 20 Mar 2024 16:29

Tgt Ion: 65 Resp: 66348

Ion	Ratio	Lower	Upper
65	100		
92	70.1	58.9	88.3
138	96.8	89.5	134.3









#47
Dimethylphthalate
Concen: 4.658 ng/ul
RT: 13.834 min Scan# 1827
Delta R.T. -0.006 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:163 Resp: 415844
Ion Ratio Lower Upper
163 100
194 3.9 3.3 4.9
164 10.0 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 3.035 ng/ul
RT: 13.945 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:165 Resp: 52320
Ion Ratio Lower Upper
165 100
89 65.0 47.5 71.3
121 28.0 18.6 28.0#





#50

Acenaphthylene

Concen: 4.735 ng/ul

RT: 14.087 min Scan# 1871

Delta R.T. -0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:152 Resp: 548188

Ion Ratio Lower Upper

152 100

151 20.0 16.2 24.4

153 13.6 10.5 15.7



#51

3-Nitroaniline

Concen: 3.560 ng/ul

RT: 14.269 min Scan# 1901

Delta R.T. -0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion:138 Resp: 60132


Ion Ratio Lower Upper

138 100

108 25.4 12.4 18.6#

92 127.9 95.5 143.3









#52
Acenaphthene
Concen: 4.727 ng/ul
RT: 14.434 min Scan# 1929
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:153 Resp: 362602
Ion Ratio Lower Upper
153 100
152 46.3 38.1 57.1
154 85.3 68.0 102.0



#53
2,4-Dinitrophenol
Concen: 1.501 ng/ul
RT: 14.469 min Scan# 1935
Delta R.T. -0.006 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:184 Resp: 14541
Ion Ratio Lower Upper
184 100
63 78.7 56.1 84.1
154 64.7 49.8 74.6









#55
4-Nitrophenol
Concen: 2.865 ng/ul
RT: 14.563 min Scan# 1951
Delta R.T. -0.012 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:109 Resp: 50590
Ion Ratio Lower Upper
109 100
139 134.1 113.9 170.9
65 133.5 103.7 155.5



#56
Dibenzofuran
Concen: 4.779 ng/ul
RT: 14.769 min Scan# 1986
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:168 Resp: 488935
Ion Ratio Lower Upper
168 100
139 37.6 30.6 46.0





#57

2,4-Dinitrotoluene

Concen: 3.067 ng/ul

RT: 14.728 min Scan# 1980

Delta R.T. -0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:165 Resp: 75471

Ion Ratio Lower Upper

165 100

63 46.7 35.4 53.0

182 2.6 2.4 3.6



#58

2,3,4,6-Tetrachlorophenol

Concen: 3.933 ng/ul

RT: 14.992 min Scan# 2024

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Tgt Ion:232 Resp: 75499




Ion Ratio Lower Upper


232 100

131 47.9 37.6 56.4

130 2.2 1.8 2.8

166 31.2 25.8 38.6





#59

Diethylphthalate

Concen: 4.331 ng/ul

RT: 15.204 min Scan# 2060

Delta R.T. -0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:149 Resp: 413221

Ion Ratio Lower Upper

149 100

177 22.4 17.5 26.3

150 12.2 9.8 14.8



#61

Fluorene

Concen: 4.854 ng/ul

RT: 15.422 min Scan# 2097

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion:166 Resp: 410009


Ion Ratio Lower Upper

166 100

165 96.7 75.4 113.0

167 13.5 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 4.793 ng/ul
RT: 15.422 min Scan# 2097
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:204 Resp: 193818
Ion Ratio Lower Upper
204 100
206 33.8 26.5 39.7
141 63.0 50.9 76.3



#63
4-Nitroaniline
Concen: 3.919 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.012 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:138 Resp: 61293
Ion Ratio Lower Upper
138 100
92 56.4 42.6 64.0
108 84.5 63.4 95.0









#66
4,6-Dinitro-2-methylphenol
Concen: 1.798 ng/ul
RT: 15.487 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:198 Resp: 28374
Ion Ratio Lower Upper
198 100
182 5.3 3.6 5.4
77 30.6 21.2 31.8



#67
N-Nitrosodiphenylamine
Concen: 4.621 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:169 Resp: 341405
Ion Ratio Lower Upper
169 100
168 67.4 50.2 75.2
167 36.3 28.3 42.5









#68
4-Bromophenyl-phenylether
Concen: 4.320 ng/ul
RT: 16.316 min Scan# 2249
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:248 Resp: 108381
Ion Ratio Lower Upper
248 100
250 98.6 78.6 117.8
141 81.2 61.5 92.3



#69
Hexachlorobenzene
Concen: 4.850 ng/ul
RT: 16.416 min Scan# 2266
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:284 Resp: 133274
Ion Ratio Lower Upper
284 100
142 36.5 28.5 42.7
249 33.2 22.1 33.1#





#70

Atrazine

Concen: 4.209 ng/ul

RT: 16.587 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:200 Resp: 107031

Ion Ratio Lower Upper

200 100

173 28.8 21.9 32.9

215 48.5 37.7 56.5



#71

Pentachlorophenol

Concen: 3.196 ng/ul

RT: 16.757 min Scan# 2324

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion:266 Resp: 59418


Ion Ratio Lower Upper

266 100

264 70.2 50.7 76.1

268 68.9 52.2 78.2









#72
Phenanthrene
Concen: 4.742 ng/ul
RT: 17.157 min Scan# 2392
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:178 Resp: 647649
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.2
176 19.1 14.8 22.2



#74
Anthracene
Concen: 4.649 ng/ul
RT: 17.251 min Scan# 2408
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:178 Resp: 645948
Ion Ratio Lower Upper
178 100
179 15.0 11.6 17.4
176 18.8 14.1 21.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 4.258 ng/uL

RT: 13.157 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:216 Resp: 148860

Ion Ratio Lower Upper

216 100

214 82.3 61.1 91.7

218 49.6 39.0 58.4



#76

Pentachlorobenzene

Concen: 4.609 ng/uL

RT: 14.681 min Scan# 1971

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion:250 Resp: 150280


Ion Ratio Lower Upper

250 100

248 65.5 50.1 75.1

252 63.6 51.3 76.9









#77
 Carbazole
 Concen: 4.786 ng/ul
 RT: 17.516 min Scan# 2453
 Delta R.T. -0.006 min
 Lab File: BN030432.D
 Acq: 20 Mar 2024 16:29




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005246


Tgt Ion:167 Resp: 590510
 Ion Ratio Lower Upper
 167 100
 166 21.0 16.3 24.5
 139 13.1 10.2 15.4



#78
 Di-n-butylphthalate
 Concen: 4.013 ng/ul
 RT: 18.104 min Scan# 2553
 Delta R.T. -0.006 min
 Lab File: BN030432.D
 Acq: 20 Mar 2024 16:29

Tgt Ion:149 Resp: 687954
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.2 10.8
 104 5.3 4.1 6.1









#80
Fluoranthene
Concen: 4.302 ng/ul
RT: 19.180 min Scan# 2736
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29




Instrument :
BNA_N
ClientSampleId :
SSTD005246


Tgt Ion:202 Resp: 701614
Ion Ratio Lower Upper
202 100
101 13.5 10.2 15.2
100 9.9 7.7 11.5



#82
Pyrene
Concen: 4.446 ng/ul
RT: 19.545 min Scan# 2798
Delta R.T. -0.000 min
Lab File: BN030432.D
Acq: 20 Mar 2024 16:29

Tgt Ion:202 Resp: 745229
Ion Ratio Lower Upper
202 100
101 15.4 11.5 17.3
100 12.5 9.3 13.9





#83

Butylbenzylphthalate

Concen: 2.797 ng/ul

RT: 20.469 min Scan# 2955

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:149 Resp: 224600

Ion Ratio Lower Upper

149 100

91 76.6 57.4 86.0

206 18.9 16.3 24.5



#84

3,3'-Dichlorobenzidine

Concen: 3.771 ng/ul

RT: 21.275 min Scan# 3092

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion:252 Resp: 171778


Ion Ratio Lower Upper

252 100

254 65.1 53.9 80.9

126 12.8 10.1 15.1





#85

Benzo(a)anthracene

Concen: 4.699 ng/ul

RT: 21.345 min Scan# 3104

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:228 Resp: 738990

Ion Ratio Lower Upper

228 100

229 20.1 14.7 22.1

226 26.7 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 3.007 ng/ul

RT: 21.304 min Scan# 3097

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion:149 Resp: 365551


Ion Ratio Lower Upper

149 100

167 27.3 21.3 31.9

279 3.7 3.4 5.2





#87

Chrysene

Concen: 4.755 ng/ul

RT: 21.404 min Scan# 3114

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:228 Resp: 693576

Ion Ratio Lower Upper

228 100

226 29.9 22.3 33.5

229 19.0 15.2 22.8



#89

Di-n-octyl phthalate

Concen: 2.529 ng/ul




RT: 22.445 min Scan# 3291


Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Tgt Ion:149 Resp: 537439





#90

Benzo(b)fluoranthene

Concen: 4.362 ng/ul

RT: 23.398 min Scan# 3452

Delta R.T. 0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:252 Resp: 652905

Ion Ratio Lower Upper

252 100

253 21.8 17.2 25.8

125 11.2 9.0 13.4



#91

Benzo(k)fluoranthene

Concen: 4.324 ng/ul

RT: 23.457 min Scan# 3463

Delta R.T. -0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion:252 Resp: 668499


Ion Ratio Lower Upper

252 100

253 21.1 16.9 25.3

125 11.5 9.5 14.3





#93

Benzo(a)pyrene

Concen: 4.332 ng/ul

RT: 24.198 min Scan# 3

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:252 Resp: 613768

Ion Ratio Lower Upper

252 100

253 21.7 17.8 26.8

125 12.7 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 3.311 ng/ul

RT: 27.674 min Scan# 4180

Delta R.T. -0.012 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29




Tgt Ion:276 Resp: 557858


Ion Ratio Lower Upper

276 100

138 23.6 17.3 25.9

277 22.6 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 4.062 ng/ul

RT: 27.751 min Scan# 4193

Delta R.T. -0.000 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Instrument :

BNA_N

ClientSampleId :

SSTD005246





Tgt Ion:278 Resp: 557991

Ion Ratio Lower Upper

278 100

139 19.1 15.0 22.4

279 24.1 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 3.861 ng/ul

RT: 28.715 min Scan# 4357

Delta R.T. -0.006 min

Lab File: BN030432.D

Acq: 20 Mar 2024 16:29

Tgt Ion:276 Resp: 526849

Ion Ratio Lower Upper

276 100

138 23.8 19.4 29.2

277 23.1 17.7 26.5

