

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032024\
 Data File : BN030435.D
 Acq On : 20 Mar 2024 18:28
 Operator : MA/JU
 Sample : SSTD04023
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD040249

Quant Time: Mar 21 00:02:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 21 00:01:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.728	152	376768	20.000	ng/u1	0.00
20) Naphthalene-d8	10.516	136	1666579	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.369	164	1020479	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.122	188	2134000	20.000	ng/u1	0.00
79) Chrysene-d12	21.363	240	1873256	20.000	ng/u1	0.00
88) Perylene-d12	24.333	264	1986320	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	171603	15.371	ng/uL	0.00
4) Pyridine-d5	3.617	84	1248216	46.836	ng/u1	0.00
7) Phenol-d5	6.887	99	1539005	46.039	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.069	67	928315	48.907	ng/u1	0.00
11) 2-Chlorophenol-d4	7.258	132	1157375	43.221	ng/u1	0.00
15) 4-Methylphenol-d8	8.428	113	1212704	46.028	ng/u1	0.00
21) Nitrobenzene-d5	8.887	128	572986	41.204	ng/u1	0.00
24) 2-Nitrophenol-d4	9.599	143	539655	37.015	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.134	165	1127639	43.368	ng/u1	0.00
31) 4-Chloroaniline-d4	10.651	131	1757907	45.551	ng/u1	0.00
46) Dimethylphthalate-d6	13.793	166	3400391	41.327	ng/u1	0.00
49) Acenaphthylene-d8	14.063	160	3937061	42.238	ng/u1	0.00
54) 4-Nitrophenol-d4	14.569	143	658209	45.119	ng/u1	0.01
60) Fluorene-d10	15.369	176	2836818	42.029	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.487	200	410825	30.822	ng/u1	0.00
73) Anthracene-d10	17.222	188	4462696	43.067	ng/u1	0.00
81) Pyrene-d10	19.516	212	4781303	40.192	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.133	264	4622358	44.429	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.258	88	176457	16.466	ng/uL	96
5) Pyridine	3.640	79	1264544	46.638	ng/u1	99
6) Benzaldehyde	6.875	77	742677	47.668	ng/u1	99
8) Phenol	6.916	94	1564563	47.035	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.163	93	1280385	48.163	ng/u1	98
12) 2-Chlorophenol	7.293	128	1193623	43.321	ng/u1	98
13) 2-Methylphenol	8.163	108	1171917	46.991	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.258	45	1677829	69.413	ng/u1	98
16) Acetophenone	8.552	105	1910542	45.115	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.546	70	999883	49.348	ng/u1	99
18) 4-Methylphenol	8.493	108	1285283	48.773	ng/u1	100
19) Hexachloroethane	8.805	117	489744	38.260	ng/u1	97
22) Nitrobenzene	8.928	77	1408564	39.252	ng/u1	99
23) Isophorone	9.452	82	2912431	46.948	ng/u1	99
25) 2-Nitrophenol	9.634	139	621231	39.553	ng/u1	98
26) 2,4-Dimethylphenol	9.693	107	1339237	41.310	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.940	93	1745040	47.957	ng/u1	100
29) 2,4-Dichlorophenol	10.163	162	1102583	42.220	ng/u1	99
30) Naphthalene	10.563	128	3965792	41.687	ng/u1	100
32) 4-Chloroaniline	10.675	127	1722038	45.877	ng/u1	98
33) Hexachlorobutadiene	10.851	225	648652	35.605	ng/u1	97
34) Caprolactam	11.451	113	232762	29.931	ng/u1	95
35) 4-Chloro-3-methylphenol	11.798	107	1300153	43.995	ng/u1	97
36) 2-Methylnaphthalene	12.187	142	2677401	43.214	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032024\
 Data File : BN030435.D
 Acq On : 20 Mar 2024 18:28
 Operator : MA/JU
 Sample : SSTD04023
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD040249

Quant Time: Mar 21 00:02:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 21 00:01:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.404	142	2648923	43.246	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.551	216	1294889	40.183	ng/ul	98
40) Hexachlorocyclopentadiene	12.534	237	741307	30.712	ng/ul	94
41) 2,4,6-Trichlorophenol	12.793	196	851109	41.629	ng/ul	96
42) 2,4,5-Trichlorophenol	12.857	196	947370	42.403	ng/ul	91
43) 1,1'-Biphenyl	13.204	154	3429510	39.819	ng/ul	99
44) 2-Chloronaphthalene	13.245	162	2689518	40.208	ng/ul	100
45) 2-Nitroaniline	13.445	65	791865	41.061	ng/ul	95
47) Dimethylphthalate	13.840	163	3406048	40.732	ng/ul	100
48) 2,6-Dinitrotoluene	13.951	165	668227	41.379	ng/ul	97
50) Acenaphthylene	14.092	152	4466640	41.187	ng/ul	99
51) 3-Nitroaniline	14.275	138	696783	44.033	ng/ul	98
52) Acenaphthene	14.434	153	2949333	41.048	ng/ul	96
53) 2,4-Dinitrophenol	14.481	184	244658	26.963	ng/ul	96
55) 4-Nitrophenol	14.581	109	499392	30.194	ng/ul	98
56) Dibenzofuran	14.775	168	3954382	41.262	ng/ul	99
57) 2,4-Dinitrotoluene	14.740	165	974098	42.265	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.992	232	784803	43.647	ng/ul	98
59) Diethylphthalate	15.216	149	3504589	39.211	ng/ul	99
61) Fluorene	15.422	166	3132170	39.589	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.428	204	1509447	39.852	ng/ul	97
63) 4-Nitroaniline	15.445	138	681491	46.516	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.498	198	478199	33.249	ng/ul	99
67) N-Nitrosodiphenylamine	15.639	169	2835601	42.116	ng/ul	97
68) 4-Bromophenyl-phenylether	16.322	248	932646	40.792	ng/ul	96
69) Hexachlorobenzene	16.416	284	1106164	44.178	ng/ul	95
70) Atrazine	16.598	200	949758	40.987	ng/ul	98
71) Pentachlorophenol	16.763	266	694816	41.008	ng/ul	94
72) Phenanthrene	17.163	178	5116803	41.113	ng/ul	99
74) Anthracene	17.257	178	5193097	41.019	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.157	216	1295816	40.676	ng/ul	99
76) Pentachlorobenzene	14.687	250	1243858	41.869	ng/ul	94
77) Carbazole	17.522	167	4839971	43.049	ng/ul	96
78) Di-n-butylphthalate	18.110	149	5986734	38.320	ng/ul	99
80) Fluoranthene	19.180	202	5916542	40.588	ng/ul	99
82) Pyrene	19.545	202	6017324	40.162	ng/ul	98
83) Butylbenzylphthalate	20.469	149	2595618	36.163	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.274	252	1883404	46.265	ng/ul	94
85) Benzo(a)anthracene	21.351	228	5927686	42.168	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.304	149	3872290	35.643	ng/ul	97
87) Chrysene	21.410	228	5511666	42.276	ng/ul	96
89) Di-n-octyl phthalate	22.445	149	6389822	34.366	ng/ul	100
90) Benzo(b)fluoranthene	23.404	252	5676515	43.344	ng/ul	98
91) Benzo(k)fluoranthene	23.468	252	5637291	41.676	ng/ul	98
93) Benzo(a)pyrene	24.204	252	5393490	43.513	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	27.692	276	5254735	35.647	ng/ul	99
95) Dibenzo(a,h)anthracene	27.768	278	4828446	40.181	ng/ul	100
96) Benzo(g,h,i)perylene	28.733	276	4577763	38.342	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 16.466 ng/uL
RT: 3.258 min Scan# 29
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion: 88 Resp: 176457
Ion Ratio Lower Upper
88 100
43 24.7 21.6 32.4
58 70.4 53.5 80.3



#5
Pyridine
Concen: 46.638 ng/uL
RT: 3.640 min Scan# 94
Delta R.T. -0.000 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion: 79 Resp: 1264544
Ion Ratio Lower Upper
79 100
52 67.6 54.2 81.4
51 31.8 26.0 39.0









#6
Benzaldehyde
Concen: 47.668 ng/ul
RT: 6.875 min Scan# 644
Delta R.T. -0.000 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion: 77 Resp: 742677
Ion Ratio Lower Upper
77 100
105 97.1 76.7 115.1
106 94.8 75.4 113.2



#8
Phenol
Concen: 47.035 ng/ul
RT: 6.916 min Scan# 651
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion: 94 Resp: 1564563
Ion Ratio Lower Upper
94 100
65 26.3 21.3 31.9
66 35.0 27.8 41.8









#10
Bis(2-Chloroethyl)ether
Concen: 48.163 ng/ul
RT: 7.163 min Scan# 692
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion: 93 Resp: 1280385
Ion Ratio Lower Upper
93 100
63 71.7 59.4 89.0
95 32.1 25.4 38.0



#12
2-Chlorophenol
Concen: 43.321 ng/ul
RT: 7.293 min Scan# 715
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion:128 Resp: 1193623
Ion Ratio Lower Upper
128 100
64 47.5 37.3 55.9
130 33.8 25.2 37.8









#13
2-Methylphenol
Concen: 46.991 ng/ul
RT: 8.163 min Scan# 80
Delta R.T. -0.000 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:108 Resp: 1171917
Ion Ratio Lower Upper
108 100
107 90.7 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 69.413 ng/ul
RT: 8.258 min Scan# 879
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion: 45 Resp: 1677829
Ion Ratio Lower Upper
45 100
77 15.7 12.0 18.0
79 11.7 9.0 13.4









#16
Acetophenone
Concen: 45.115 ng/ul
RT: 8.552 min Scan# 91
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:105 Resp: 1910542
Ion Ratio Lower Upper
105 100
77 78.0 62.5 93.7
51 24.8 20.0 30.0



#17
N-Nitroso-di-n-propylamine
Concen: 49.348 ng/ul
RT: 8.546 min Scan# 928
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion: 70 Resp: 999883
Ion Ratio Lower Upper
70 100
42 51.9 42.4 63.6
101 9.8 8.0 12.0
130 21.8 17.0 25.4





#18

4-Methylphenol

Concen: 48.773 ng/ul

RT: 8.493 min Scan# 918

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :





SSTD040249

Tgt Ion:108 Resp: 1285283

Ion Ratio Lower Upper

108 100

107 113.4 90.3 135.5



#19

Hexachloroethane

Concen: 38.260 ng/ul

RT: 8.805 min Scan# 972

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:117 Resp: 489744


Ion Ratio Lower Upper

117 100

201 91.2 74.2 111.2

199 60.4 45.9 68.9









#22
Nitrobenzene
Concen: 39.252 ng/ul
RT: 8.928 min Scan# 992
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion: 77 Resp: 1408564
Ion Ratio Lower Upper
77 100
123 45.3 35.9 53.9
65 12.9 10.2 15.4



#23
Isophorone
Concen: 46.948 ng/ul
RT: 9.452 min Scan# 1082
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion: 82 Resp: 2912431
Ion Ratio Lower Upper
82 100
95 6.8 5.6 8.4
138 17.2 13.5 20.3





#25

2-Nitrophenol

Concen: 39.553 ng/ul

RT: 9.634 min Scan# 1113

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:139 Resp: 621231

Ion Ratio Lower Upper

139 100

65 48.4 40.2 60.4

109 29.1 23.5 35.3



#26

2,4-Dimethylphenol

Concen: 41.310 ng/ul

RT: 9.693 min Scan# 1123

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:107 Resp: 1339237


Ion Ratio Lower Upper

107 100

121 50.7 40.7 61.1

122 88.2 69.3 103.9





#27

Bis(2-Chloroethoxy)methane

Concen: 47.957 ng/ul

RT: 9.940 min Scan# 1165

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :





BNA_N

ClientSampleId :

SSTD040249

Tgt Ion: 93 Resp: 1745040

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.6	39.8
123	11.1	8.9	13.3



#29

2,4-Dichlorophenol

Concen: 42.220 ng/ul

RT: 10.163 min Scan# 1203




Delta R.T. 0.006 min


Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Tgt Ion:162 Resp: 1102583

Ion	Ratio	Lower	Upper
162	100		
164	64.4	51.0	76.4
98	37.9	30.2	45.4









#30
Naphthalene
Concen: 41.687 ng/ul
RT: 10.563 min Scan# 1271
Delta R.T. -0.000 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:128 Resp: 3965792
Ion Ratio Lower Upper
128 100
129 10.8 8.6 12.8
127 12.8 10.4 15.6



#32
4-Chloroaniline
Concen: 45.877 ng/ul
RT: 10.675 min Scan# 1290
Delta R.T. -0.000 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion:127 Resp: 1722038
Ion Ratio Lower Upper
127 100
129 31.5 26.0 39.0









#33
Hexachlorobutadiene
Concen: 35.605 ng/ul
RT: 10.851 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:225 Resp: 648652
Ion Ratio Lower Upper
225 100
223 62.1 48.0 72.0
227 65.0 50.6 76.0



#34
Caprolactam
Concen: 29.931 ng/ul
RT: 11.451 min Scan# 1422
Delta R.T. 0.018 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion:113 Resp: 232762
Ion Ratio Lower Upper
113 100
55 147.3 121.4 182.0
56 111.5 95.1 142.7









#35
4-Chloro-3-methylphenol
Concen: 43.995 ng/ul
RT: 11.798 min Scan# 1480
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:107 Resp: 1300153
Ion Ratio Lower Upper
107 100
144 25.6 19.9 29.9
142 78.8 60.6 90.8



#36
2-Methylnaphthalene
Concen: 43.214 ng/ul
RT: 12.187 min Scan# 1547
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion:142 Resp: 2677401
Ion Ratio Lower Upper
142 100
141 89.2 70.6 105.8





#37

1-Methylnaphthalene

Concen: 43.246 ng/ul

RT: 12.404 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:142 Resp: 2648923

Ion Ratio Lower Upper

142 100

141 95.2 75.5 113.3

116 3.7 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 40.183 ng/ul

RT: 12.551 min Scan# 1609

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Tgt Ion:216 Resp: 1294889




Ion Ratio Lower Upper


216 100

214 77.3 60.7 91.1

179 20.8 17.7 26.5

108 17.6 14.7 22.1





#40

Hexachlorocyclopentadiene

Concen: 30.712 ng/ul

RT: 12.534 min Scan# 1606

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:237 Resp: 741307

Ion Ratio Lower Upper

237 100

235 61.4 53.6 80.4

272 12.3 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 41.629 ng/ul

RT: 12.793 min Scan# 1650

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:196 Resp: 851109


Ion Ratio Lower Upper

196 100

198 91.6 77.0 115.6

200 30.9 24.3 36.5





#42

2,4,5-Trichlorophenol

Concen: 42.403 ng/ul

RT: 12.857 min Scan# 1661

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:196 Resp: 947370

Ion Ratio Lower Upper

196 100

198 90.1 80.4 120.6

200 29.1 25.1 37.7



#43

1,1'-Biphenyl

Concen: 39.819 ng/ul

RT: 13.204 min Scan# 1720

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:154 Resp: 3429510


Ion Ratio Lower Upper

154 100

153 41.1 32.3 48.5

76 14.2 11.2 16.8









#44
2-Chloronaphthalene
Concen: 40.208 ng/ul
RT: 13.245 min Scan# 1726
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:162 Resp: 2689518
Ion Ratio Lower Upper
162 100
164 32.7 26.0 39.0
127 40.1 32.3 48.5



#45
2-Nitroaniline
Concen: 41.061 ng/ul
RT: 13.445 min Scan# 1761
Delta R.T. -0.000 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion: 65 Resp: 791865
Ion Ratio Lower Upper
65 100
92 70.6 58.9 88.3
138 105.1 89.5 134.3









#47
Dimethylphthalate
Concen: 40.732 ng/ul
RT: 13.840 min Scan# 1828
Delta R.T. -0.000 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:163 Resp: 3406048
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.0 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 41.379 ng/ul
RT: 13.951 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion:165 Resp: 668227
Ion Ratio Lower Upper
165 100
89 56.8 47.5 71.3
121 22.1 18.6 28.0





#50

Acenaphthylene

Concen: 41.187 ng/ul

RT: 14.092 min Scan# 1871

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:152 Resp: 4466640

Ion Ratio Lower Upper

152 100

151 20.8 16.2 24.4

153 13.8 10.5 15.7



#51

3-Nitroaniline

Concen: 44.033 ng/ul

RT: 14.275 min Scan# 1902

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:138 Resp: 696783


Ion Ratio Lower Upper

138 100

108 12.9 12.4 18.6

92 118.3 95.5 143.3









#52
 Acenaphthene
 Concen: 41.048 ng/ul
 RT: 14.434 min Scan# 1929
 Delta R.T. -0.000 min
 Lab File: BN030435.D
 Acq: 20 Mar 2024 18:28

Instrument : BNA_N
 ClientSampleId : SSTD040249

Tgt Ion:153 Resp: 2949333




Ion	Ratio	Lower	Upper
153	100		
152	47.2	38.1	57.1
154	91.0	68.0	102.0




#53
 2,4-Dinitrophenol
 Concen: 26.963 ng/ul
 RT: 14.481 min Scan# 1937
 Delta R.T. 0.006 min
 Lab File: BN030435.D
 Acq: 20 Mar 2024 18:28

Tgt Ion:184 Resp: 244658

Ion	Ratio	Lower	Upper
184	100		
63	66.7	56.1	84.1
154	59.7	49.8	74.6









#55
4-Nitrophenol
Concen: 30.194 ng/ul
RT: 14.581 min Scan# 1954
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:109 Resp: 499392
Ion Ratio Lower Upper
109 100
139 144.4 113.9 170.9
65 132.2 103.7 155.5



#56
Dibenzofuran
Concen: 41.262 ng/ul
RT: 14.775 min Scan# 1987
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion:168 Resp: 3954382
Ion Ratio Lower Upper
168 100
139 38.7 30.6 46.0





#57

2,4-Dinitrotoluene

Concen: 42.265 ng/ul

RT: 14.740 min Scan# 1980

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:165 Resp: 974098

Ion Ratio Lower Upper

165 100

63 41.6 35.4 53.0

182 3.0 2.4 3.6



#58

2,3,4,6-Tetrachlorophenol

Concen: 43.647 ng/ul

RT: 14.992 min Scan# 2024

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Tgt Ion:232 Resp: 784803




Ion Ratio Lower Upper


232 100

131 47.7 37.6 56.4

130 2.4 1.8 2.8

166 30.9 25.8 38.6





#59

Diethylphthalate

Concen: 39.211 ng/ul

RT: 15.216 min Scan# 2062

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:149 Resp: 3504589

Ion Ratio Lower Upper

149 100

177 22.2 17.5 26.3

150 12.9 9.8 14.8



#61

Fluorene

Concen: 39.589 ng/ul

RT: 15.422 min Scan# 2097

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:166 Resp: 3132170


Ion Ratio Lower Upper

166 100

165 99.5 75.4 113.0

167 14.1 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 39.852 ng/ul
RT: 15.428 min Scan# 2097
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:204 Resp: 1509447
Ion Ratio Lower Upper
204 100
206 33.8 26.5 39.7
141 60.8 50.9 76.3



#63
4-Nitroaniline
Concen: 46.516 ng/ul
RT: 15.445 min Scan# 2101
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion:138 Resp: 681491
Ion Ratio Lower Upper
138 100
92 51.7 42.6 64.0
108 77.8 63.4 95.0









#66
4,6-Dinitro-2-methylphenol
Concen: 33.249 ng/ul
RT: 15.498 min Scan# 2109
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:198 Resp: 478199
Ion Ratio Lower Upper
198 100
182 4.0 3.6 5.4
77 26.1 21.2 31.8



#67
N-Nitrosodiphenylamine
Concen: 42.116 ng/ul
RT: 15.639 min Scan# 2134
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion:169 Resp: 2835601
Ion Ratio Lower Upper
169 100
168 65.8 50.2 75.2
167 36.3 28.3 42.5









#68
4-Bromophenyl-phenylether
Concen: 40.792 ng/ul
RT: 16.322 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:248 Resp: 932646
Ion Ratio Lower Upper
248 100
250 103.4 78.6 117.8
141 75.8 61.5 92.3



#69
Hexachlorobenzene
Concen: 44.178 ng/ul
RT: 16.416 min Scan# 2266
Delta R.T. -0.000 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion:284 Resp: 1106164
Ion Ratio Lower Upper
284 100
142 38.3 28.5 42.7
249 30.9 22.1 33.1





#70

Atrazine

Concen: 40.987 ng/ul

RT: 16.598 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:200 Resp: 949758

Ion Ratio Lower Upper

200 100

173 26.1 21.9 32.9

215 47.8 37.7 56.5



#71

Pentachlorophenol

Concen: 41.008 ng/ul

RT: 16.763 min Scan# 2325

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:266 Resp: 694816


Ion Ratio Lower Upper

266 100

264 66.9 50.7 76.1

268 70.8 52.2 78.2









#72
Phenanthrene
Concen: 41.113 ng/ul
RT: 17.163 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28




Instrument :
BNA_N
ClientSampleId :
SSTD040249


Tgt Ion:178 Resp: 5116803
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.2
176 19.2 14.8 22.2



#74
Anthracene
Concen: 41.019 ng/ul
RT: 17.257 min Scan# 2409
Delta R.T. 0.006 min
Lab File: BN030435.D
Acq: 20 Mar 2024 18:28

Tgt Ion:178 Resp: 5193097
Ion Ratio Lower Upper
178 100
179 15.2 11.6 17.4
176 18.5 14.1 21.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 40.676 ng/uL

RT: 13.157 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:216 Resp: 1295816

Ion Ratio Lower Upper

216 100

214 76.4 61.1 91.7

218 46.6 39.0 58.4



#76

Pentachlorobenzene

Concen: 41.869 ng/uL

RT: 14.687 min Scan# 1972

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:250 Resp: 1243858


Ion Ratio Lower Upper

250 100

248 69.1 50.1 75.1

252 66.7 51.3 76.9









#77
 Carbazole
 Concen: 43.049 ng/ul
 RT: 17.522 min Scan# 2454
 Delta R.T. -0.000 min
 Lab File: BN030435.D
 Acq: 20 Mar 2024 18:28

Instrument :
 BNA_N
 ClientSampleId :
 SSTD040249

Tgt Ion:167 Resp: 4839971




Ion	Ratio	Lower	Upper
167	100		
166	22.2	16.3	24.5
139	14.0	10.2	15.4




#78
 Di-n-butylphthalate
 Concen: 38.320 ng/ul
 RT: 18.110 min Scan# 2554
 Delta R.T. -0.000 min
 Lab File: BN030435.D
 Acq: 20 Mar 2024 18:28

Tgt Ion:149 Resp: 5986734

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.2	10.8
104	5.2	4.1	6.1









#80
 Fluoranthene
 Concen: 40.588 ng/u1
 RT: 19.180 min Scan# 2736
 Delta R.T. -0.000 min
 Lab File: BN030435.D
 Acq: 20 Mar 2024 18:28

Instrument :
 BNA_N
 ClientSampleId :
 SST040249

Tgt Ion:202 Resp: 5916542




Ion	Ratio	Lower	Upper
202	100		
101	13.1	10.2	15.2
100	10.1	7.7	11.5




#82
 Pyrene
 Concen: 40.162 ng/u1
 RT: 19.545 min Scan# 2798
 Delta R.T. -0.000 min
 Lab File: BN030435.D
 Acq: 20 Mar 2024 18:28

Tgt Ion:202 Resp: 6017324

Ion	Ratio	Lower	Upper
202	100		
101	15.0	11.5	17.3
100	12.3	9.3	13.9





#83

Butylbenzylphthalate

Concen: 36.163 ng/ul

RT: 20.469 min Scan# 2955

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:149 Resp: 2595618

Ion Ratio Lower Upper

149 100

91 74.3 57.4 86.0

206 19.8 16.3 24.5



#84

3,3'-Dichlorobenzidine

Concen: 46.265 ng/ul

RT: 21.274 min Scan# 3092

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:252 Resp: 1883404


Ion Ratio Lower Upper

252 100

254 61.8 53.9 80.9

126 12.2 10.1 15.1





#85

Benzo(a)anthracene

Concen: 42.168 ng/ul

RT: 21.351 min Scan# 3105

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:228 Resp: 5927686

Ion Ratio Lower Upper

228 100

229 18.2 14.7 22.1

226 27.3 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 35.643 ng/ul

RT: 21.304 min Scan# 3097

Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:149 Resp: 3872290


Ion Ratio Lower Upper

149 100

167 28.3 21.3 31.9

279 4.4 3.4 5.2





#87

Chrysene

Concen: 42.276 ng/ul

RT: 21.410 min Scan# 3114

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:228 Resp: 5511666

Ion Ratio Lower Upper

228 100

226 30.3 22.3 33.5

229 20.4 15.2 22.8



#89

Di-n-octyl phthalate

Concen: 34.366 ng/ul




RT: 22.445 min Scan# 3291


Delta R.T. -0.000 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Tgt Ion:149 Resp: 6389822





#90

Benzo(b)fluoranthene

Concen: 43.344 ng/ul

RT: 23.404 min Scan# 3454

Delta R.T. 0.012 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:252 Resp: 5676515

Ion Ratio Lower Upper

252 100

253 22.6 17.2 25.8

125 12.0 9.0 13.4



#91

Benzo(k)fluoranthene

Concen: 41.676 ng/ul

RT: 23.468 min Scan# 3465

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:252 Resp: 5637291


Ion Ratio Lower Upper

252 100

253 22.1 16.9 25.3

125 11.5 9.5 14.3





#93

Benzo(a)pyrene

Concen: 43.513 ng/uI

RT: 24.204 min Scan# 3590

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:252 Resp: 5393490

Ion Ratio Lower Upper

252 100

253 22.3 17.8 26.8

125 12.2 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 35.647 ng/uI

RT: 27.692 min Scan# 4183

Delta R.T. 0.006 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28




Tgt Ion:276 Resp: 5254735


Ion Ratio Lower Upper

276 100

138 22.2 17.3 25.9

277 23.4 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 40.181 ng/ul

RT: 27.768 min Scan# 4193

Delta R.T. 0.018 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Instrument :

BNA_N

ClientSampleId :

SSTD040249





Tgt Ion:278 Resp: 4828446

Ion Ratio Lower Upper

278 100

139 18.8 15.0 22.4

279 23.8 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 38.342 ng/ul

RT: 28.733 min Scan# 4360

Delta R.T. 0.012 min

Lab File: BN030435.D

Acq: 20 Mar 2024 18:28

Tgt Ion:276 Resp: 4577763

Ion Ratio Lower Upper

276 100

138 24.7 19.4 29.2

277 25.3 17.7 26.5

