


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032025\
Data File : BN036688.D
Acq On : 20 Mar 2025 08:40
Operator : RC/JU
Sample : DFTPP371
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP371

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN022825.MA.M
Title : SVOA CALIBRATION
Last Update : Thu Mar 20 13:16:23 2025



AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2451

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	67.1	63941	PASS
68	69	0.00	2	1.4	712	PASS
69	198	0.00	100	54.9	52336	PASS
70	69	0.00	2	0.5	259	PASS
127	198	10	80	52.3	49856	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	95251	PASS
199	198	5	9	7.1	6739	PASS
275	198	10	60	25.4	24224	PASS
365	198	1	100	4.2	4001	PASS
441	198	0.01	100	8.8	8352	PASS
442	198	50	100	54.2	51589	PASS
443	442	15	24	19.3	9956	PASS

m/z	Abundance
39.90	999.0
44.00	1033.0
73.00	308.0
207.00	591.0

Instrument :

BNA_N

ClientSampleId :

DFTPP371

m/z	Abundance
36.00	156.0
37.00	546.0
38.00	1108.0
39.00	5410.0
40.00	1017.0
44.00	1293.0
49.10	614.0
50.00	20168.0
51.00	65080.0
52.00	3565.0
52.90	166.0
54.90	388.0
56.00	1791.0
57.00	3807.0
60.90	760.0
61.90	838.0
63.00	2590.0
64.10	237.0
65.00	1016.0
68.00	1292.0
69.00	51760.0
69.90	331.0
73.00	801.0
74.00	6760.0
75.00	9136.0
76.10	3430.0
77.00	57744.0
78.00	4221.0
78.90	5197.0
80.00	3365.0
81.00	4525.0
81.90	1268.0
83.00	845.0
84.80	801.0
86.00	1618.0
87.00	586.0
87.90	251.0
91.00	787.0
92.00	1030.0
93.00	8914.0
93.90	631.0
95.90	285.0
97.90	7089.0
99.00	5593.0
100.10	337.0
101.00	2445.0
103.00	755.0
103.90	1693.0
104.90	1594.0
107.00	18232.0
108.00	2218.0
110.00	27608.0
111.00	4800.0
111.90	510.0
112.90	193.0
116.10	999.0
116.90	21936.0
117.90	1589.0
118.90	218.0
122.00	1294.0
123.00	2050.0
124.00	939.0
124.80	769.0
127.00	47520.0
128.00	3836.0
129.00	21984.0
129.90	1987.0
130.90	414.0
132.00	162.0
133.90	696.0
134.90	1501.0
135.90	869.0
137.00	903.0
138.80	163.0
139.90	363.0
140.90	3190.0
141.90	735.0
142.90	659.0
144.00	198.0
145.00	203.0
145.90	632.0
146.90	1782.0
148.00	4807.0
149.00	1100.0
150.00	163.0
150.90	401.0
153.00	889.0
154.00	653.0
154.90	1336.0
156.00	2265.0
157.10	425.0
158.00	544.0

Instrument :
BNA_N
ClientSampleId :
DFTPP371

159.00	534.0
160.00	849.0
161.00	1187.0
162.00	417.0
165.00	1156.0
165.90	920.0
167.00	6142.0
168.00	4934.0
169.00	584.0
169.90	337.0
170.90	318.0
171.90	607.0
172.80	599.0
173.90	1121.0
175.00	2148.0
176.00	378.0
177.00	1050.0
178.90	4195.0
180.00	2788.0
181.00	1295.0
181.90	154.0
183.90	446.0
185.00	2175.0
186.00	14382.0
187.00	4128.0
187.90	498.0
188.90	990.0
190.10	224.0
190.90	545.0
191.90	1493.0
193.00	1609.0
193.90	290.0
194.90	215.0
196.00	2382.0
197.90	81712.0
198.90	6408.0
200.00	379.0
201.50	416.0
202.90	892.0
203.90	3599.0
205.00	6321.0
206.00	25112.0
207.00	3696.0
207.90	1142.0
209.00	529.0
209.90	399.0
210.90	968.0
211.60	244.0
215.00	503.0
216.00	804.0
216.90	7294.0
217.90	888.0
219.90	268.0
220.90	2005.0
221.90	894.0
222.90	1683.0
224.00	14634.0
225.00	3211.0
226.00	514.0
226.90	7797.0
227.90	1258.0
228.90	1534.0
229.90	207.0
231.00	480.0
233.80	402.0
235.00	613.0
235.90	403.0
237.00	848.0
239.00	329.0
239.90	246.0
240.90	419.0
241.90	731.0
243.00	955.0
244.00	10338.0
244.90	1400.0
245.90	2575.0
246.80	503.0
247.90	190.0
248.90	326.0
252.80	251.0
255.00	49136.0
256.00	7694.0
257.00	663.0
257.90	3540.0
258.90	587.0
264.90	1366.0
265.90	244.0
271.90	185.0
272.90	1492.0
274.00	3496.0
275.00	18480.0
276.00	2842.0
277.00	2347.0
278.00	371.0
281.00	219.0
282.80	206.0

Instrument :
BNA_N
ClientSampleId :
DFTPP371

284.90	396.0
293.00	351.0
296.00	7091.0
296.90	1128.0
303.00	719.0
303.90	294.0
314.00	290.0
314.80	726.0
315.90	440.0
321.00	154.0
323.00	1375.0
323.90	284.0
326.90	362.0
334.00	1034.0
335.00	369.0
340.90	220.0
345.90	416.0
351.90	536.0
353.00	400.0
354.00	332.0
365.00	2925.0
365.90	501.0
370.90	168.0
371.90	873.0
382.80	264.0
401.90	258.0
403.00	254.0
403.90	181.0
421.00	353.0
422.00	433.0
423.00	1972.0
424.00	518.0
441.00	5446.0
441.90	31552.0
443.00	7126.0
444.00	620.0

Instrument :
BNA_N
ClientSampleId :
DFTPP371

m/z	Abundance
37.00	497.0
38.00	1287.0
39.00	6628.0
40.00	1053.0
41.00	164.0
44.00	1275.0
49.00	747.0
50.00	22872.0
51.00	78800.0
52.00	3530.0
52.90	170.0
55.00	518.0
56.00	2021.0
57.00	4549.0
57.90	241.0
61.00	946.0
62.00	1058.0
63.00	2861.0
64.00	496.0
65.10	1085.0
69.00	62968.0
70.00	445.0
73.10	750.0
74.00	7519.0
75.00	10768.0
76.00	3936.0
77.00	70032.0
78.00	5109.0
79.00	5880.0
80.00	4477.0
81.00	5329.0
82.00	1375.0
82.90	1051.0
83.90	234.0
84.90	969.0
86.00	1931.0
87.00	729.0
87.90	234.0
90.90	1342.0
92.00	1323.0
93.00	11007.0
94.00	622.0
96.00	270.0
98.00	8258.0
99.00	6347.0
99.90	628.0
101.00	2683.0
103.00	1005.0
103.90	2029.0
104.90	2006.0
106.00	664.0
107.00	21720.0
108.00	3091.0
110.00	33456.0
111.00	5565.0
111.90	677.0
112.80	296.0
115.10	150.0
116.00	1451.0
116.90	27472.0
118.00	1632.0
118.90	200.0
119.90	290.0
121.90	1515.0
122.90	2052.0
124.00	1159.0
124.90	870.0
127.00	61288.0
128.00	4822.0
129.00	26848.0
130.00	2130.0
130.90	490.0
133.00	162.0
133.90	854.0
135.00	2408.0
135.90	875.0
136.90	1554.0
138.10	297.0
139.00	164.0
139.90	468.0
141.00	4054.0
141.90	1290.0
143.00	940.0
143.90	312.0
144.90	160.0
146.00	853.0
147.00	2019.0
147.90	5288.0
149.00	1142.0
150.00	255.0
151.00	549.0
151.80	269.0

Instrument :
BNA_N
ClientSampleId :
DFTPP371

153.00	1190.0
153.90	919.0
155.00	2135.0
156.00	3109.0
157.00	526.0
157.90	727.0
158.90	548.0
159.90	1314.0
161.00	1484.0
162.00	406.0
162.90	156.0
164.00	316.0
164.90	1556.0
165.90	1302.0
167.00	9095.0
168.00	4926.0
169.00	701.0
169.90	368.0
170.80	402.0
171.90	854.0
172.90	1018.0
174.00	1675.0
175.00	3028.0
176.10	629.0
177.00	1336.0
178.00	725.0
178.90	5726.0
180.00	4291.0
181.00	1834.0
182.00	315.0
183.90	501.0
185.00	2475.0
186.00	19872.0
187.00	5759.0
188.00	684.0
188.90	1559.0
189.90	392.0
191.00	764.0
192.00	2151.0
193.00	2095.0
194.00	398.0
195.00	324.0
196.00	3322.0
197.90	115824.0
198.90	8213.0
199.90	739.0
201.50	441.0
202.90	1237.0
204.00	5764.0
205.00	10074.0
206.00	35808.0
207.00	5100.0
208.00	1721.0
208.90	721.0
210.00	796.0
211.00	1593.0
214.90	524.0
216.00	980.0
216.90	10448.0
217.90	1357.0
219.00	197.0
220.90	3468.0
221.90	1354.0
222.90	2323.0
224.00	21704.0
225.00	5173.0
226.00	480.0
226.90	9763.0
227.90	1625.0
229.00	1977.0
229.90	253.0
230.90	551.0
231.90	185.0
233.90	841.0
234.90	895.0
235.80	431.0
237.00	919.0
238.80	286.0
239.90	386.0
241.00	582.0
241.90	1143.0
243.00	1309.0
244.00	13871.0
245.00	1701.0
245.90	3944.0
246.90	708.0
248.00	172.0
248.90	495.0
250.80	158.0
251.90	157.0
252.90	485.0
255.00	71816.0
256.00	10048.0
257.00	987.0
257.90	5565.0
258.90	922.0

Instrument :
BNA_N
ClientSampleId :
DFTPP371

259.80	170.0
264.90	2335.0
265.90	410.0
271.90	240.0
273.00	2042.0
274.00	6505.0
275.00	30456.0
276.00	3347.0
276.90	3179.0
277.90	589.0
281.00	156.0
282.90	524.0
283.90	221.0
285.00	581.0
289.90	159.0
292.10	158.0
292.90	533.0
293.90	235.0
295.10	313.0
296.00	11917.0
297.00	1794.0
302.10	202.0
303.00	1317.0
303.90	298.0
308.00	156.0
313.00	156.0
314.00	356.0
314.90	1255.0
315.90	492.0
321.10	333.0
323.00	2883.0
324.00	433.0
326.90	771.0
328.00	190.0
331.90	207.0
333.00	293.0
333.90	2031.0
335.00	392.0
340.90	237.0
345.90	662.0
352.00	1087.0
353.00	632.0
354.10	643.0
354.90	194.0
364.90	4703.0
365.90	632.0
370.90	268.0
372.00	1203.0
372.80	364.0
382.80	313.0
390.10	158.0
401.90	619.0
403.00	932.0
403.90	196.0
420.90	467.0
422.00	537.0
423.00	3939.0
423.90	1206.0
441.00	9577.0
442.00	65512.0
443.00	10697.0
443.90	1118.0

Instrument :
BNA_N
ClientSampleId :
DFTPP371

m/z	Abundance
37.00	320.0
38.00	776.0
39.00	4013.0
40.00	903.0
44.00	927.0
50.00	14220.0
51.00	47944.0
52.00	2276.0
54.90	284.0
55.90	1248.0
57.00	2947.0
60.90	645.0
62.00	825.0
63.00	1742.0
64.00	257.0
65.00	808.0
68.00	843.0
69.00	42280.0
73.00	744.0
74.00	5080.0
75.00	7760.0
76.10	2909.0
77.00	44832.0
78.00	3435.0
79.00	3925.0
80.00	2599.0
81.00	3605.0
82.00	743.0
82.90	583.0
84.90	619.0
85.90	963.0
87.00	461.0
87.90	180.0
91.00	755.0
91.90	872.0
93.00	7517.0
93.90	486.0
95.90	169.0
98.00	5613.0
99.00	3765.0
99.90	399.0
101.00	1823.0
103.00	566.0
104.00	1192.0
105.00	1522.0
105.90	523.0
107.00	14443.0
108.00	2312.0
109.00	624.0
110.00	21640.0
110.90	3454.0
112.00	438.0
112.90	232.0
115.90	966.0
116.90	18424.0
117.90	1236.0
119.80	172.0
121.90	1095.0
123.00	1524.0
123.90	722.0
124.90	606.0
127.00	40760.0
128.00	3254.0
129.00	18296.0
129.90	1424.0
130.90	313.0
131.80	245.0
132.90	216.0
133.90	626.0
135.00	1807.0
135.80	571.0
137.00	1047.0
138.00	221.0
139.00	175.0
140.00	334.0
141.00	2609.0
141.90	787.0
142.90	446.0
143.80	151.0
146.00	709.0
147.00	1613.0
148.00	3948.0
148.90	751.0
149.90	287.0
151.00	297.0
151.90	238.0
152.90	1083.0
154.00	711.0
155.00	1570.0
156.00	2183.0
156.90	378.0
157.80	577.0

Instrument :
BNA_N
ClientSampleId :
DFTPP371

158.90	314.0
159.90	1008.0
161.00	1254.0
162.00	368.0
164.90	985.0
165.90	736.0
167.00	6627.0
167.90	3795.0
169.00	376.0
170.00	237.0
171.10	259.0
171.90	527.0
173.00	781.0
174.00	1122.0
175.00	2157.0
175.90	481.0
176.90	1091.0
177.90	457.0
178.90	4477.0
180.00	3094.0
181.00	1231.0
181.90	280.0
183.90	406.0
185.00	1673.0
186.00	14894.0
187.00	4148.0
187.90	550.0
188.90	1240.0
190.90	428.0
192.00	1389.0
193.00	1855.0
193.90	373.0
194.90	224.0
196.00	2491.0
197.90	88216.0
198.90	5595.0
200.00	627.0
201.10	266.0
201.50	297.0
202.90	1047.0
204.00	4138.0
205.00	6363.0
206.00	27880.0
207.00	4034.0
208.00	1337.0
208.90	418.0
209.80	425.0
210.90	1362.0
211.70	259.0
215.00	396.0
215.90	894.0
216.90	8423.0
218.00	1068.0
219.90	231.0
220.90	2834.0
221.90	953.0
223.00	1879.0
224.00	15616.0
225.00	3851.0
226.00	593.0
227.00	8216.0
227.90	1122.0
229.00	1861.0
231.00	536.0
233.90	417.0
235.00	514.0
235.80	398.0
236.90	572.0
238.80	316.0
239.90	296.0
241.00	458.0
241.90	951.0
243.00	979.0
244.00	10473.0
245.00	1632.0
245.90	3055.0
247.00	681.0
248.80	435.0
251.90	153.0
252.80	300.0
255.00	56176.0
256.00	9145.0
256.90	742.0
258.00	4436.0
258.90	678.0
259.80	173.0
264.90	1763.0
266.00	399.0
272.00	210.0
273.00	1869.0
274.00	4905.0
275.00	23736.0
276.00	3120.0
276.90	2814.0
277.90	423.0
281.00	168.0

Instrument :
BNA_N
ClientSampleId :
DFTPP371

283.00	392.0
283.90	218.0
285.00	429.0
292.90	661.0
296.00	9009.0
296.90	1335.0
302.00	167.0
303.00	957.0
304.00	264.0
307.90	208.0
314.10	363.0
314.90	1007.0
316.00	561.0
320.90	425.0
321.70	157.0
322.90	2001.0
324.00	283.0
326.90	474.0
327.70	211.0
332.00	227.0
333.00	198.0
334.00	1796.0
335.00	442.0
341.10	424.0
346.00	783.0
351.90	753.0
352.80	480.0
353.90	823.0
354.80	189.0
365.00	4376.0
365.90	619.0
372.00	1097.0
373.00	210.0
382.90	304.0
401.90	437.0
402.90	769.0
403.90	281.0
420.90	462.0
421.90	551.0
423.00	3401.0
423.90	853.0
425.10	185.0
441.00	10034.0
441.90	57704.0
443.00	12046.0
444.00	949.0

Instrument :
BNA_N
ClientSampleId :
DFTPP371