


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_N\Data\BN032118\
 Data File : BN000512.D
 Acq On : 21 Mar 2018 10:09
 Operator : JU/SJ
 Sample : DFTPP83
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP83

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
 Title : SVOA CALIBRATION
 Last Update : Thu Mar 22 01:34:06 2018



AutoFind: Scans 2500, 2501, 2502; Background Corrected with Scan 2494

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.3	97357	PASS
68	69	0.00	2	1.9	2224	PASS
69	198	0.00	100	40.5	118177	PASS
70	69	0.00	2	0.5	558	PASS
127	198	10	80	50.6	147720	PASS
197	198	0.00	2	0.7	1964	PASS
198	198	100	100	100.0	292010	PASS
199	198	5	9	6.8	19763	PASS
275	198	10	60	23.6	68853	PASS
365	198	1	100	2.9	8596	PASS
441	443	0.01	100	78.5	32208	PASS
442	198	50	100	70.6	206280	PASS
443	442	15	24	19.9	41032	PASS

m/z	Abundance
37.00	480.0
38.00	1095.0
39.10	6575.0
40.00	786.0
41.00	997.0
42.00	335.0
43.10	1011.0
44.00	1156.0
45.00	406.0
49.00	819.0
50.10	22736.0
51.10	88368.0
52.00	4376.0
53.00	404.0
55.10	1250.0
56.00	2980.0
57.00	6368.0
58.00	444.0
60.10	443.0
61.00	1519.0
62.00	1482.0
63.00	3971.0
64.00	650.0
65.10	2060.0
67.00	540.0
68.10	1915.0
69.00	105464.0
70.00	791.0
71.10	362.0
73.10	1900.0
74.00	10863.0
75.00	17472.0
76.00	6062.0
77.10	126632.0
78.10	9221.0
79.00	8417.0
80.00	6182.0
81.00	8686.0
82.00	2099.0
83.00	2309.0
84.00	487.0
85.00	2335.0
86.00	2710.0
87.00	1288.0
88.00	481.0
91.00	2359.0
92.00	2039.0
93.00	13967.0
94.00	1157.0
95.00	550.0
95.90	770.0
97.10	601.0
98.00	11115.0
99.00	8364.0
100.00	808.0
101.00	4949.0
102.00	346.0
103.00	1596.0
104.00	2773.0
105.00	2883.0
106.00	978.0
107.00	38120.0
108.00	5986.0
109.10	1323.0
110.00	68840.0
111.00	10821.0
112.00	1425.0
113.00	461.0
115.00	395.0
116.00	2158.0
117.00	29960.0
118.00	2444.0
119.00	525.0
120.00	496.0
120.90	327.0
122.00	2211.0
123.00	3825.0
124.00	1727.0
124.90	1497.0
126.10	776.0
127.00	123288.0
128.00	9220.0
129.00	46952.0
130.00	4392.0
131.00	956.0
132.00	514.0
133.00	399.0
134.00	1479.0
135.00	4003.0
136.00	1497.0
137.10	2403.0

Instrument :
BNA_N
ClientSampleId :
DFTPP83

138.10	681.0
139.00	317.0
139.90	594.0
141.00	5993.0
142.00	2172.0
143.00	1422.0
144.00	455.0
145.10	391.0
146.00	1273.0
147.00	3258.0
148.00	7020.0
149.00	1582.0
150.00	567.0
151.20	850.0
152.00	558.0
153.00	1905.0
154.00	1493.0
155.10	3072.0
156.10	4616.0
157.00	1014.0
157.90	1238.0
159.00	952.0
160.00	2020.0
161.00	2861.0
162.00	1098.0
163.00	371.0
164.10	400.0
165.00	2325.0
166.00	1652.0
167.00	12661.0
168.00	7066.0
169.00	1160.0
170.10	385.0
171.00	575.0
172.00	1219.0
173.10	1314.0
174.00	2383.0
175.00	4489.0
176.00	1378.0
177.00	2238.0
178.00	804.0
179.00	8125.0
180.00	6023.0
181.00	2824.0
182.00	629.0
183.90	611.0
185.00	4267.0
186.00	31544.0
187.00	8618.0
188.00	1043.0
189.00	1997.0
189.90	367.0
191.00	887.0
192.00	2843.0
193.00	3099.0
194.00	769.0
195.10	368.0
196.00	6004.0
197.10	1493.0
198.00	226816.0
199.00	15913.0
200.00	1285.0
201.50	960.0
203.00	1797.0
204.10	8119.0
205.10	13491.0
206.10	56872.0
207.10	8510.0
208.00	2201.0
209.00	851.0
211.10	2300.0
211.80	468.0
215.00	711.0
216.00	1306.0
217.00	14375.0
218.00	1936.0
221.10	8418.0
223.00	3296.0
224.10	30912.0
225.10	7604.0
226.00	975.0
227.00	12886.0
228.00	1866.0
229.00	2374.0
229.90	319.0
231.10	1042.0
234.00	764.0
235.00	941.0
236.00	615.0
237.00	1079.0
239.00	444.0
240.00	405.0
241.00	751.0
242.00	1537.0
243.10	1669.0

Instrument :
BNA_N
ClientSampleId :
DFTPP83

244.10	22408.0
245.10	2987.0
246.00	4541.0
247.00	999.0
249.00	920.0
253.00	483.0
254.20	1234.0
255.00	114496.0
256.10	17200.0
257.10	1333.0
258.00	6401.0
259.00	1112.0
265.00	2864.0
266.10	498.0
273.00	3265.0
274.00	8950.0
275.10	49408.0
276.10	7115.0
277.00	4428.0
278.00	620.0
283.10	545.0
284.10	360.0
285.10	681.0
293.00	895.0
295.10	323.0
296.00	13242.0
297.00	2059.0
303.10	1572.0
304.00	482.0
314.00	600.0
315.00	1214.0
316.00	728.0
320.90	374.0
323.10	3826.0
324.10	921.0
327.00	782.0
328.00	447.0
332.00	343.0
333.10	400.0
334.10	2406.0
335.10	596.0
341.00	565.0
346.00	945.0
352.10	1379.0
353.00	1042.0
354.10	1097.0
355.10	327.0
365.00	5136.0
366.00	979.0
371.10	396.0
372.10	2221.0
373.00	563.0
383.00	469.0
390.00	310.0
402.00	829.0
403.00	1171.0
404.10	451.0
421.10	1141.0
422.10	1046.0
423.10	6689.0
424.10	1761.0
441.10	18184.0
442.10	117272.0
443.10	24352.0
444.00	1955.0

Instrument :

BNA_N

ClientSampleId :

DFTPP83

m/z	Abundance
37.00	543.0
38.00	1492.0
39.10	7933.0
40.00	889.0
41.10	936.0
42.00	371.0
43.00	983.0
44.00	1185.0
45.00	502.0
49.00	951.0
50.10	28544.0
51.10	113488.0
52.10	6123.0
53.00	367.0
55.00	1260.0
56.00	3775.0
57.00	8535.0
58.00	605.0
60.00	416.0
61.00	1835.0
62.00	2000.0
63.00	4909.0
64.00	816.0
65.10	2705.0
66.10	310.0
67.00	485.0
68.10	2699.0
69.00	138624.0
70.00	966.0
71.10	390.0
73.10	2220.0
74.00	14654.0
75.00	22856.0
76.10	7926.0
77.10	171136.0
78.10	12258.0
79.00	11478.0
80.00	8331.0
81.00	12108.0
82.00	3171.0
83.00	2937.0
84.00	698.0
85.00	3407.0
86.00	3607.0
87.00	1793.0
88.00	732.0
88.90	464.0
91.00	3053.0
92.00	3164.0
93.00	19000.0
94.00	1270.0
95.00	630.0
96.00	1089.0
97.00	546.0
98.00	15143.0
99.00	12445.0
100.00	1134.0
101.00	6631.0
101.90	511.0
103.00	2471.0
104.00	4058.0
105.00	4030.0
106.10	1351.0
107.00	50920.0
108.00	7934.0
109.10	1656.0
110.00	95272.0
111.00	14141.0
112.00	1761.0
113.00	511.0
115.00	391.0
116.00	2933.0
117.00	43264.0
118.00	3238.0
119.00	549.0
120.00	871.0
121.00	339.0
122.00	3359.0
123.00	5385.0
124.00	2431.0
125.00	2309.0
126.20	1043.0
127.00	175040.0
128.00	13080.0
129.00	65768.0
130.00	5471.0
131.00	1264.0
132.10	727.0
132.90	400.0
134.00	1943.0
135.00	5409.0

Instrument :
BNA_N
ClientSampleId :
DFTPP83

136.00	2108.0
137.00	3077.0
137.90	565.0
139.00	576.0
140.00	816.0
141.00	8626.0
142.00	2651.0
143.00	1889.0
143.90	600.0
145.00	569.0
146.00	1672.0
147.00	4714.0
148.00	10684.0
149.00	2284.0
149.90	709.0
151.10	1303.0
152.10	531.0
153.00	2898.0
154.00	2048.0
155.10	5319.0
156.10	7357.0
157.10	1394.0
158.00	1594.0
159.00	1245.0
160.00	2555.0
161.00	4113.0
162.00	1374.0
163.00	415.0
164.00	502.0
165.00	3572.0
166.00	2783.0
167.00	17696.0
168.00	9697.0
169.00	1743.0
170.00	609.0
171.00	727.0
172.00	1624.0
173.00	2017.0
174.00	3573.0
175.10	6875.0
176.10	1833.0
177.00	2915.0
178.00	1149.0
179.00	12850.0
180.00	8869.0
181.10	4243.0
182.00	807.0
183.10	383.0
184.00	1145.0
185.00	6206.0
186.00	47648.0
187.00	13468.0
188.00	1384.0
189.00	2806.0
189.90	494.0
191.10	1652.0
192.00	3741.0
193.00	4738.0
194.00	1208.0
195.10	475.0
196.00	9047.0
197.20	2276.0
198.00	345344.0
199.00	23864.0
200.00	1770.0
201.50	1510.0
203.00	2335.0
204.00	12022.0
205.10	20256.0
206.10	87976.0
207.10	12478.0
208.10	3178.0
209.00	1063.0
210.10	1557.0
211.00	3164.0
211.80	693.0
215.00	963.0
216.00	1912.0
217.00	22696.0
218.00	3008.0
219.00	348.0
221.10	12654.0
221.80	3179.0
223.00	4946.0
224.10	47424.0
225.10	12741.0
226.10	1227.0
227.00	19800.0
228.00	2639.0
229.00	3750.0
229.90	520.0
231.10	1453.0
231.90	336.0
233.00	356.0
233.90	1102.0

Instrument :
BNA_N
ClientSampleId :
DFTPP83

235.00	1398.0
236.00	972.0
237.00	1800.0
239.00	727.0
240.00	471.0
241.00	1044.0
242.00	3023.0
243.00	2653.0
244.10	36096.0
245.10	4550.0
246.00	6787.0
247.00	1443.0
248.00	388.0
249.00	1383.0
251.90	473.0
253.10	832.0
255.00	177728.0
256.10	27152.0
257.00	2009.0
258.00	10028.0
259.00	1887.0
259.90	329.0
261.00	317.0
264.00	319.0
265.00	4141.0
265.90	625.0
271.10	320.0
272.10	531.0
273.00	5312.0
274.10	15133.0
275.10	80256.0
276.10	11699.0
277.00	6891.0
278.10	1057.0
281.10	328.0
283.10	798.0
284.10	502.0
285.10	1116.0
292.00	419.0
293.10	1557.0
294.00	428.0
295.00	420.0
296.00	21736.0
297.00	3086.0
302.00	389.0
303.10	2721.0
304.10	644.0
308.10	326.0
314.10	1071.0
315.10	2432.0
316.10	1288.0
321.00	653.0
322.00	391.0
323.10	6866.0
324.10	1365.0
327.00	1352.0
328.00	673.0
332.10	540.0
333.00	609.0
334.10	4357.0
335.10	1335.0
341.10	706.0
346.00	1681.0
347.00	361.0
352.10	2280.0
353.00	1457.0
354.00	2242.0
355.10	503.0
365.00	9713.0
366.00	1763.0
371.10	618.0
372.10	3711.0
373.00	874.0
383.00	965.0
390.10	480.0
391.10	390.0
402.10	1464.0
403.10	2034.0
404.00	779.0
421.00	2066.0
422.10	1980.0
423.10	12795.0
424.00	2802.0
441.10	37192.0
442.10	235072.0
443.10	46480.0
444.00	4207.0

Instrument :
BNA_N
ClientSampleId :
DFTPP83

m/z	Abundance
37.10	420.0
38.00	1049.0
39.10	6372.0
40.00	679.0
41.10	965.0
42.00	347.0
43.00	1112.0
44.00	1248.0
45.00	410.0
49.00	840.0
50.00	23856.0
51.10	90216.0
52.00	4716.0
53.10	395.0
55.00	1273.0
56.00	3035.0
57.00	7017.0
57.90	418.0
60.00	397.0
61.00	1389.0
62.00	1509.0
63.00	4014.0
64.10	586.0
65.00	2280.0
67.00	461.0
68.10	2060.0
69.00	112088.0
70.10	845.0
71.00	347.0
73.00	1871.0
74.00	12532.0
75.00	18272.0
76.10	6494.0
77.10	137472.0
78.10	9802.0
79.00	8852.0
80.00	6589.0
81.00	9499.0
82.00	2480.0
83.00	2351.0
83.90	529.0
85.00	2900.0
86.00	2829.0
87.00	1373.0
87.90	632.0
89.00	352.0
91.00	2582.0
92.00	2590.0
93.00	15168.0
94.10	1297.0
95.00	562.0
96.00	993.0
97.10	686.0
98.00	12706.0
99.00	9371.0
100.00	869.0
101.00	5546.0
102.10	335.0
103.00	1969.0
104.00	3753.0
105.00	3507.0
106.00	1094.0
107.00	41120.0
108.00	6611.0
109.10	1523.0
110.00	78856.0
111.00	11655.0
112.00	1488.0
113.00	527.0
114.90	314.0
116.00	2537.0
117.00	35040.0
118.00	2632.0
119.10	360.0
120.00	578.0
121.00	315.0
122.00	2766.0
123.00	4407.0
124.00	2111.0
125.00	1956.0
126.20	1010.0
127.00	144832.0
128.00	10776.0
129.00	54984.0
130.00	4721.0
131.00	1028.0
132.00	587.0
132.90	434.0
134.00	1626.0
135.00	4662.0
136.00	1993.0

Instrument :
BNA_N
ClientSampleId :
DFTPP83

137.00	2846.0
137.90	518.0
139.00	400.0
140.00	800.0
141.00	7289.0
142.00	2367.0
143.00	1706.0
144.00	529.0
145.10	499.0
146.00	1209.0
147.00	4030.0
148.00	8916.0
149.00	1991.0
150.00	670.0
151.10	1104.0
151.90	640.0
153.00	2452.0
154.00	1885.0
155.00	4250.0
156.00	6365.0
157.10	1409.0
158.00	1411.0
159.00	1117.0
160.00	2413.0
161.00	3604.0
162.00	1176.0
162.90	384.0
163.90	536.0
165.00	2940.0
166.10	2344.0
167.00	15928.0
168.00	9160.0
169.00	1524.0
170.00	562.0
171.00	671.0
172.00	1525.0
173.00	1779.0
174.00	3109.0
175.00	6100.0
176.00	1722.0
177.00	2780.0
178.00	887.0
179.00	11662.0
180.00	7332.0
181.00	3580.0
182.10	806.0
182.90	394.0
184.00	962.0
185.10	5332.0
186.00	40376.0
187.00	11778.0
188.10	1160.0
189.00	2687.0
190.00	327.0
191.00	1328.0
192.00	3522.0
193.00	4129.0
194.00	992.0
195.00	565.0
196.00	8000.0
197.10	2123.0
198.00	303872.0
199.00	19512.0
200.00	1655.0
201.60	1252.0
203.00	2197.0
204.10	10958.0
205.10	18224.0
206.10	78904.0
207.10	10764.0
208.00	2957.0
209.10	828.0
210.10	1290.0
211.00	2735.0
215.00	961.0
216.00	1727.0
217.00	20096.0
218.00	2889.0
221.10	12430.0
221.80	2904.0
223.00	4691.0
224.10	42520.0
225.10	10792.0
226.10	1141.0
227.00	17616.0
228.00	2502.0
229.00	3731.0
230.10	501.0
231.00	1512.0
232.00	305.0
234.00	1315.0
235.00	1375.0
236.00	858.0
237.00	1339.0
239.00	718.0

Instrument :
BNA_N
ClientSampleId :
DFTPP83

240.00	450.0
241.00	933.0
242.00	2466.0
243.10	2423.0
244.10	33640.0
245.10	4703.0
246.00	6364.0
247.00	1460.0
249.00	1281.0
250.90	349.0
252.10	321.0
253.00	867.0
255.00	171008.0
256.10	25448.0
257.00	1681.0
258.00	9629.0
259.00	1603.0
260.00	313.0
264.00	339.0
265.00	3978.0
266.00	684.0
272.00	359.0
273.10	5429.0
274.10	14198.0
275.10	76896.0
276.10	9607.0
277.00	6515.0
278.10	997.0
279.00	347.0
281.10	303.0
283.00	724.0
284.00	484.0
285.10	1120.0
293.00	1502.0
294.00	359.0
295.10	514.0
296.10	22752.0
297.00	2907.0
301.00	328.0
302.00	503.0
303.00	2691.0
304.10	615.0
308.10	367.0
314.00	1083.0
315.00	2277.0
316.10	1428.0
321.00	705.0
322.00	558.0
323.10	7274.0
324.10	1215.0
327.00	1576.0
328.00	750.0
332.10	529.0
333.10	686.0
334.10	4634.0
335.10	1212.0
341.00	867.0
346.00	1824.0
352.10	2062.0
353.00	1580.0
354.00	2636.0
355.10	529.0
365.00	10939.0
366.00	1662.0
371.00	576.0
372.10	4231.0
373.10	940.0
383.00	1035.0
390.00	474.0
391.10	360.0
402.00	1943.0
403.00	2158.0
404.00	792.0
421.10	1877.0
422.10	1872.0
423.10	14453.0
424.10	2912.0
441.10	41248.0
442.10	266496.0
443.10	52264.0
444.10	4636.0

Instrument :
BNA_N
ClientSampleId :
DFTPP83

m/z	Abundance
39.10	416.0
40.00	513.0
41.10	994.0
42.00	367.0
43.00	1440.0
44.00	1181.0
45.00	423.0
55.00	975.0
56.00	361.0
57.10	661.0
60.00	429.0
67.00	368.0
69.00	548.0
70.10	309.0
73.00	853.0
77.00	363.0
78.90	414.0
81.00	383.0
83.00	464.0
84.00	390.0
84.90	344.0
91.00	452.0
97.00	313.0
98.00	315.0
146.90	306.0
207.10	1104.0
208.00	328.0
281.00	332.0

Instrument :
BNA_N
ClientSampleId :
DFTPP83