

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032124\
 Data File : BN030442.D
 Acq On : 21 Mar 2024 11:40
 Operator : MA/JU
 Sample : PB159672TB 10X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLEB672

Quant Time: Mar 21 12:16:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 21 00:37:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	335790	20.000	ng/u1	0.00
20) Naphthalene-d8	10.510	136	1468666	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.369	164	977019	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.116	188	2024285	20.000	ng/u1	0.00
79) Chrysene-d12	21.357	240	1724943	20.000	ng/u1	0.00
88) Perylene-d12	24.321	264	1875161	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.222	96	69502	7.583	ng/uL	0.00
4) Pyridine-d5	3.617	84	997729	37.554	ng/u1	0.00
7) Phenol-d5	6.881	99	1160443	35.360	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.063	67	749280	37.707	ng/u1	0.00
11) 2-Chlorophenol-d4	7.258	132	888301	37.084	ng/u1	0.00
15) 4-Methylphenol-d8	8.422	113	920321	35.836	ng/u1	0.00
21) Nitrobenzene-d5	8.881	128	457604	40.808	ng/u1	0.00
24) 2-Nitrophenol-d4	9.599	143	415991	41.870	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.128	165	827020	36.823	ng/u1	0.00
31) 4-Chloroaniline-d4	10.646	131	1406182	38.420	ng/u1	0.00
46) Dimethylphthalate-d6	13.787	166	2954463	38.591	ng/u1	0.00
49) Acenaphthylene-d8	14.057	160	3370803	37.787	ng/u1	0.00
54) 4-Nitrophenol-d4	14.551	143	472306	32.624	ng/u1	0.00
60) Fluorene-d10	15.363	176	2476228	37.962	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.475	200	230068	25.501	ng/u1	0.00
73) Anthracene-d10	17.216	188	3894347	39.607	ng/u1	0.00
81) Pyrene-d10	19.510	212	4260322	40.303	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.121	264	3989571	40.093	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032124\
Data File : BN030442.D
Acq On : 21 Mar 2024 11:40
Operator : MA/JU
Sample : PB159672TB 10X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLEB672

Quant Time: Mar 21 12:16:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032024.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 21 00:37:15 2024
Response via : Initial Calibration

