

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032124\
 Data File : BN030445.D
 Acq On : 21 Mar 2024 13:38
 Operator : MA/JU
 Sample : P1743-07MS 10X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E0737MS

Quant Time: Mar 21 14:26:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 21 00:37:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.728	152	331777	20.000	ng/u1	0.00
20) Naphthalene-d8	10.510	136	1471155	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.369	164	937885	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.116	188	1909495	20.000	ng/u1	0.00
79) Chrysene-d12	21.363	240	1758389	20.000	ng/u1	0.00
88) Perylene-d12	24.327	264	1781636	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.223	96	25638	2.831	ng/uL	0.00
4) Pyridine-d5	3.623	84	345786	13.173	ng/u1	0.00
7) Phenol-d5	6.887	99	258312	7.966	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.064	67	692869	35.290	ng/u1	0.00
11) 2-Chlorophenol-d4	7.258	132	667302	28.195	ng/u1	0.00
15) 4-Methylphenol-d8	8.428	113	459376	18.104	ng/u1	0.00
21) Nitrobenzene-d5	8.881	128	439418	39.120	ng/u1	0.00
24) 2-Nitrophenol-d4	9.599	143	392714	39.460	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.128	165	762234	33.881	ng/u1	0.00
31) 4-Chloroaniline-d4	10.652	131	1114625	30.403	ng/u1	0.00
46) Dimethylphthalate-d6	13.787	166	2742724	37.320	ng/u1	0.00
49) Acenaphthylene-d8	14.057	160	3107880	36.293	ng/u1	0.00
54) 4-Nitrophenol-d4	14.551	143	120367	8.661	ng/u1	0.00
60) Fluorene-d10	15.363	176	2321912	37.081	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.481	200	327799	38.518	ng/u1	0.00
73) Anthracene-d10	17.216	188	3669998	39.570	ng/u1	0.00
81) Pyrene-d10	19.516	212	4193862	38.920	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.127	264	4029005	42.615	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.258	88	58500	6.286	ng/uL#	1
5) Pyridine	3.640	79	382990	14.441	ng/u1	99
6) Benzaldehyde	6.869	77	574919	38.029	ng/u1	99
8) Phenol	6.911	94	296061	9.004	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.158	93	1006121	37.510	ng/u1	99
12) 2-Chlorophenol	7.287	128	735688	29.935	ng/u1	98
13) 2-Methylphenol	8.163	108	565130	22.738	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.252	45	1294986	36.659	ng/u1	98
16) Acetophenone	8.546	105	1549908	38.991	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.534	70	779013	37.492	ng/u1	98
18) 4-Methylphenol	8.487	108	533894	19.797	ng/u1	99
19) Hexachloroethane	8.799	117	355718	35.258	ng/u1	99
22) Nitrobenzene	8.922	77	1115046	39.211	ng/u1	99
23) Isophorone	9.452	82	2250409	36.846	ng/u1	98
25) 2-Nitrophenol	9.628	139	473832	40.377	ng/u1	98
26) 2,4-Dimethylphenol	9.687	107	744548	27.377	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.940	93	1372714	37.405	ng/u1	98
29) 2,4-Dichlorophenol	10.157	162	778844	34.661	ng/u1	98
30) Naphthalene	10.563	128	3138155	37.486	ng/u1	99
32) 4-Chloroaniline	10.675	127	831327	23.544	ng/u1	97
33) Hexachlorobutadiene	10.846	225	493503	35.936	ng/u1	97
34) Caprolactam	11.440	113	52566	5.966	ng/u1	98
35) 4-Chloro-3-methylphenol	11.793	107	855506	32.328	ng/u1	95
36) 2-Methylnaphthalene	12.181	142	2153959	37.887	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032124\
 Data File : BN030445.D
 Acq On : 21 Mar 2024 13:38
 Operator : MA/JU
 Sample : P1743-07MS 10X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E0737MS

Quant Time: Mar 21 14:26:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 21 00:37:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.399	142	2096120	37.476	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.546	216	1036435	36.688	ng/ul	96
40) Hexachlorocyclopentadiene	12.528	237	427411	26.834	ng/ul	97
41) 2,4,6-Trichlorophenol	12.787	196	642710	36.153	ng/ul	96
42) 2,4,5-Trichlorophenol	12.857	196	714686	37.058	ng/ul	98
43) 1,1'-Biphenyl	13.204	154	2747762	36.662	ng/ul	100
44) 2-Chloronaphthalene	13.240	162	2173042	36.725	ng/ul	99
45) 2-Nitroaniline	13.446	65	670976	43.364	ng/ul	99
47) Dimethylphthalate	13.840	163	2896048	38.811	ng/ul	99
48) 2,6-Dinitrotoluene	13.951	165	576333	44.944	ng/ul	99
50) Acenaphthylene	14.093	152	3677004	37.215	ng/ul	99
51) 3-Nitroaniline	14.275	138	585934	39.618	ng/ul	96
52) Acenaphthene	14.434	153	2392680	37.092	ng/ul	97
53) 2,4-Dinitrophenol	14.475	184	190826	33.127	ng/ul	100
55) 4-Nitrophenol	14.569	109	100262	9.341	ng/ul	99
56) Dibenzofuran	14.769	168	3271679	37.552	ng/ul	100
57) 2,4-Dinitrotoluene	14.734	165	847635	45.497	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	14.993	232	593315	37.427	ng/ul	98
59) Diethylphthalate	15.210	149	2991604	39.495	ng/ul	100
61) Fluorene	15.422	166	2716167	38.750	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.422	204	1325074	39.379	ng/ul	98
63) 4-Nitroaniline	15.440	138	665018	46.674	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.493	198	391841	39.804	ng/ul	97
67) N-Nitrosodiphenylamine	15.634	169	2407233	40.390	ng/ul	95
68) 4-Bromophenyl-phenylether	16.316	248	855374	43.554	ng/ul	99
69) Hexachlorobenzene	16.416	284	992531	42.198	ng/ul	96
70) Atrazine	16.592	200	364972	19.043	ng/ul	99
71) Pentachlorophenol	16.757	266	537217	37.064	ng/ul	98
72) Phenanthrene	17.157	178	4535894	41.772	ng/ul	98
74) Anthracene	17.251	178	4497612	40.801	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.157	216	1053026	39.280	ng/ul	98
76) Pentachlorobenzene	14.681	250	1051442	39.848	ng/ul	98
77) Carbazole	17.522	167	4301203	43.964	ng/ul	97
78) Di-n-butylphthalate	18.104	149	5330301	43.325	ng/ul	100
80) Fluoranthene	19.181	202	5250380	40.458	ng/ul	99
82) Pyrene	19.545	202	5414445	40.378	ng/ul	99
83) Butylbenzylphthalate	20.463	149	2278073	44.145	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.269	252	1619449	40.900	ng/ul	96
85) Benzo(a)anthracene	21.345	228	5229795	39.967	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.298	149	3378306	42.728	ng/ul	97
87) Chrysene	21.404	228	4941285	40.865	ng/ul	96
89) Di-n-octyl phthalate	22.439	149	5565933	42.679	ng/ul	100
90) Benzo(b)fluoranthene	23.392	252	5228206	44.328	ng/ul	99
91) Benzo(k)fluoranthene	23.457	252	5162736	43.335	ng/ul	99
93) Benzo(a)pyrene	24.192	252	4737360	42.924	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.680	276	5191984	43.692	ng/ul	99
95) Dibenzo(a,h)anthracene	27.751	278	4356253	43.875	ng/ul	97
96) Benzo(g,h,i)perylene	28.721	276	4040370	43.087	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 6.286 ng/uL

RT: 3.258 min Scan# 29

Delta R.T. 0.006 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38

Instrument :

BNA_N

ClientSampleId :

E0737MS





Tgt Ion: 88 Resp: 58500

Ion Ratio Lower Upper

88 100

43 867.8 21.6 32.4#

58 66.9 53.5 80.3



#5

Pyridine

Concen: 14.441 ng/uL

RT: 3.640 min Scan# 94

Delta R.T. 0.000 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38




Tgt Ion: 79 Resp: 382990


Ion Ratio Lower Upper

79 100

52 68.9 54.2 81.4

51 32.8 26.0 39.0









#6
Benzaldehyde
Concen: 38.029 ng/ul
RT: 6.869 min Scan# 643
Delta R.T. -0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion: 77 Resp: 574919
Ion Ratio Lower Upper
77 100
105 96.8 76.7 115.1
106 95.3 75.4 113.2



#8
Phenol
Concen: 9.004 ng/ul
RT: 6.911 min Scan# 650
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion: 94 Resp: 296061
Ion Ratio Lower Upper
94 100
65 26.6 21.3 31.9
66 35.8 27.8 41.8









#10
Bis(2-Chloroethyl)ether
Concen: 37.510 ng/ul
RT: 7.158 min Scan# 692
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion: 93 Resp: 1006121
Ion Ratio Lower Upper
93 100
63 73.1 59.4 89.0
95 32.7 25.4 38.0



#12
2-Chlorophenol
Concen: 29.935 ng/ul
RT: 7.287 min Scan# 714
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:128 Resp: 735688
Ion Ratio Lower Upper
128 100
64 45.4 37.3 55.9
130 32.1 25.2 37.8









#13
2-Methylphenol
Concen: 22.738 ng/ul
RT: 8.163 min Scan# 863
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:108 Resp: 565130
Ion Ratio Lower Upper
108 100
107 89.7 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 36.659 ng/ul
RT: 8.252 min Scan# 878
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion: 45 Resp: 1294986
Ion Ratio Lower Upper
45 100
77 15.8 12.0 18.0
79 11.6 9.0 13.4









#16
Acetophenone
Concen: 38.991 ng/ul
RT: 8.546 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:105 Resp: 1549908
Ion Ratio Lower Upper
105 100
77 79.3 62.5 93.7
51 24.9 20.0 30.0



#17
N-Nitroso-di-n-propylamine
Concen: 37.492 ng/ul
RT: 8.534 min Scan# 926
Delta R.T. -0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion: 70 Resp: 779013
Ion Ratio Lower Upper
70 100
42 51.2 42.4 63.6
101 9.7 8.0 12.0
130 20.8 17.0 25.4









#18
4-Methylphenol
Concen: 19.797 ng/ul
RT: 8.487 min Scan# 918
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:108 Resp: 533894
Ion Ratio Lower Upper
108 100
107 111.6 90.3 135.5



#19
Hexachloroethane
Concen: 35.258 ng/ul
RT: 8.799 min Scan# 971
Delta R.T. -0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:117 Resp: 355718
Ion Ratio Lower Upper
117 100
201 92.4 74.2 111.2
199 58.3 45.9 68.9









#22
Nitrobenzene
Concen: 39.211 ng/ul
RT: 8.922 min Scan# 992
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion: 77 Resp: 1115046
Ion Ratio Lower Upper
77 100
123 45.7 35.9 53.9
65 13.0 10.2 15.4



#23
Isophorone
Concen: 36.846 ng/ul
RT: 9.452 min Scan# 1082
Delta R.T. 0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion: 82 Resp: 2250409
Ion Ratio Lower Upper
82 100
95 6.8 5.6 8.4
138 17.9 13.5 20.3





#25

2-Nitrophenol

Concen: 40.377 ng/ul

RT: 9.628 min Scan# 1112

Delta R.T. 0.000 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38

Instrument :

BNA_N

ClientSampleId :

E0737MS





Tgt Ion:139 Resp: 473832

Ion Ratio Lower Upper

139 100

65 48.7 40.2 60.4

109 28.3 23.5 35.3



#26

2,4-Dimethylphenol

Concen: 27.377 ng/ul

RT: 9.687 min Scan# 1122

Delta R.T. -0.006 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38




Tgt Ion:107 Resp: 744548


Ion Ratio Lower Upper

107 100

121 48.7 40.7 61.1

122 85.6 69.3 103.9





#27

Bis(2-Chloroethoxy)methane

Concen: 37.405 ng/ul

RT: 9.940 min Scan# 1164

Delta R.T. 0.000 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38

Instrument :





BNA_N

ClientSampleId :

E0737MS

Tgt Ion: 93 Resp: 1372714

Ion	Ratio	Lower	Upper
93	100		
95	32.0	26.6	39.8
123	11.2	8.9	13.3



#29

2,4-Dichlorophenol

Concen: 34.661 ng/ul

RT: 10.157 min Scan# 1202




Delta R.T. 0.000 min


Lab File: BN030445.D

Acq: 21 Mar 2024 13:38

Tgt Ion:162 Resp: 778844

Ion	Ratio	Lower	Upper
162	100		
164	65.4	51.0	76.4
98	39.1	30.2	45.4









#30
Naphthalene
Concen: 37.486 ng/ul
RT: 10.563 min Scan# 1271
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:128 Resp: 3138155
Ion Ratio Lower Upper
128 100
129 10.8 8.6 12.8
127 13.7 10.4 15.6



#32
4-Chloroaniline
Concen: 23.544 ng/ul
RT: 10.675 min Scan# 1290
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:127 Resp: 831327
Ion Ratio Lower Upper
127 100
129 30.9 26.0 39.0









#33
Hexachlorobutadiene
Concen: 35.936 ng/ul
RT: 10.846 min Scan# 1319
Delta R.T. -0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:225 Resp: 493503
Ion Ratio Lower Upper
225 100
223 63.7 48.0 72.0
227 62.6 50.6 76.0



#34
Caprolactam
Concen: 5.966 ng/ul
RT: 11.440 min Scan# 1420
Delta R.T. 0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:113 Resp: 52566
Ion Ratio Lower Upper
113 100
55 154.9 121.4 182.0
56 117.5 95.1 142.7









#35
4-Chloro-3-methylphenol
Concen: 32.328 ng/ul
RT: 11.793 min Scan# 1480
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:107 Resp: 855506
Ion Ratio Lower Upper
107 100
144 25.9 19.9 29.9
142 81.0 60.6 90.8



#36
2-Methylnaphthalene
Concen: 37.887 ng/ul
RT: 12.181 min Scan# 1546
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:142 Resp: 2153959
Ion Ratio Lower Upper
142 100
141 87.7 70.6 105.8









#37
1-Methylnaphthalene
Concen: 37.476 ng/ul
RT: 12.399 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:142 Resp: 2096120
Ion Ratio Lower Upper
142 100
141 94.4 75.5 113.3
116 3.8 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 36.688 ng/ul
RT: 12.546 min Scan# 1608
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:216 Resp: 1036435
Ion Ratio Lower Upper
216 100
214 80.4 60.7 91.1
179 21.6 17.7 26.5
108 18.9 14.7 22.1









#40
Hexachlorocyclopentadiene
Concen: 26.834 ng/ul
RT: 12.528 min Scan# 1605
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:237 Resp: 427411
Ion Ratio Lower Upper
237 100
235 63.9 53.6 80.4
272 13.1 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 36.153 ng/ul
RT: 12.787 min Scan# 1649
Delta R.T. -0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:196 Resp: 642710
Ion Ratio Lower Upper
196 100
198 91.1 77.0 115.6
200 29.9 24.3 36.5





#42

2,4,5-Trichlorophenol

Concen: 37.058 ng/ul

RT: 12.857 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38

Instrument :

BNA_N

ClientSampleId :

E0737MS





Tgt Ion:196 Resp: 714686

Ion Ratio Lower Upper

196 100

198 97.3 80.4 120.6

200 31.5 25.1 37.7



#43

1,1'-Biphenyl

Concen: 36.662 ng/ul

RT: 13.204 min Scan# 1720

Delta R.T. 0.000 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38




Tgt Ion:154 Resp: 2747762


Ion Ratio Lower Upper

154 100

153 40.0 32.3 48.5

76 14.0 11.2 16.8









#44
2-Chloronaphthalene
Concen: 36.725 ng/ul
RT: 13.240 min Scan# 1726
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:162 Resp: 2173042
Ion Ratio Lower Upper
162 100
164 32.5 26.0 39.0
127 39.1 32.3 48.5



#45
2-Nitroaniline
Concen: 43.364 ng/ul
RT: 13.446 min Scan# 1761
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion: 65 Resp: 670976
Ion Ratio Lower Upper
65 100
92 73.4 58.9 88.3
138 112.9 89.5 134.3





#47

Dimethylphthalate

Concen: 38.811 ng/ul

RT: 13.840 min Scan# 1828

Delta R.T. 0.000 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38

Instrument :

BNA_N

ClientSampleId :

E0737MS





Tgt Ion:163 Resp: 2896048

Ion Ratio Lower Upper

163 100

194 4.6 3.3 4.9

164 10.3 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 44.944 ng/ul

RT: 13.951 min Scan# 1847

Delta R.T. 0.000 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38




Tgt Ion:165 Resp: 576333


Ion Ratio Lower Upper

165 100

89 59.1 47.5 71.3

121 22.8 18.6 28.0









#50
Acenaphthylene
Concen: 37.215 ng/ul
RT: 14.093 min Scan# 1871
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Instrument :
BNA_N
ClientSampleId :
E0737MS

Tgt Ion:152 Resp: 3677004




Ion	Ratio	Lower	Upper
152	100		
151	19.6	16.2	24.4
153	13.0	10.5	15.7




#51
3-Nitroaniline
Concen: 39.618 ng/ul
RT: 14.275 min Scan# 1902
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:138 Resp: 585934

Ion	Ratio	Lower	Upper
138	100		
108	13.6	12.4	18.6
92	115.6	95.5	143.3









#52
Acenaphthene
Concen: 37.092 ng/ul
RT: 14.434 min Scan# 1929
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:153 Resp: 2392680
Ion Ratio Lower Upper
153 100
152 48.7 38.1 57.1
154 88.9 68.0 102.0



#53
2,4-Dinitrophenol
Concen: 33.127 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:184 Resp: 190826
Ion Ratio Lower Upper
184 100
63 69.5 56.1 84.1
154 62.3 49.8 74.6









#55
4-Nitrophenol
Concen: 9.341 ng/ul
RT: 14.569 min Scan# 1952
Delta R.T. -0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:109 Resp: 100262
Ion Ratio Lower Upper
109 100
139 141.2 113.9 170.9
65 129.7 103.7 155.5



#56
Dibenzofuran
Concen: 37.552 ng/ul
RT: 14.769 min Scan# 1986
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:168 Resp: 3271679
Ion Ratio Lower Upper
168 100
139 38.1 30.6 46.0









#57
2,4-Dinitrotoluene
Concen: 45.497 ng/ul
RT: 14.734 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:165 Resp: 847635
Ion Ratio Lower Upper
165 100
63 40.7 35.4 53.0
182 2.8 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 37.427 ng/ul
RT: 14.993 min Scan# 2024
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:232 Resp: 593315
Ion Ratio Lower Upper
232 100
131 45.5 37.6 56.4
130 2.5 1.8 2.8
166 31.9 25.8 38.6









#59
Diethylphthalate
Concen: 39.495 ng/ul
RT: 15.210 min Scan# 2061
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:149 Resp: 2991604
Ion Ratio Lower Upper
149 100
177 21.8 17.5 26.3
150 12.5 9.8 14.8



#61
Fluorene
Concen: 38.750 ng/ul
RT: 15.422 min Scan# 2097
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:166 Resp: 2716167
Ion Ratio Lower Upper
166 100
165 94.4 75.4 113.0
167 13.3 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 39.379 ng/ul
RT: 15.422 min Scan# 2097
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:204 Resp: 1325074
Ion Ratio Lower Upper
204 100
206 33.4 26.5 39.7
141 61.1 50.9 76.3



#63
4-Nitroaniline
Concen: 46.674 ng/ul
RT: 15.440 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:138 Resp: 665018
Ion Ratio Lower Upper
138 100
92 51.4 42.6 64.0
108 75.7 63.4 95.0









#66
4,6-Dinitro-2-methylphenol
Concen: 39.804 ng/ul
RT: 15.493 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:198 Resp: 391841
Ion Ratio Lower Upper
198 100
182 4.0 3.6 5.4
77 28.3 21.2 31.8



#67
N-Nitrosodiphenylamine
Concen: 40.390 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:169 Resp: 2407233
Ion Ratio Lower Upper
169 100
168 68.0 50.2 75.2
167 36.6 28.3 42.5









#68
4-Bromophenyl-phenylether
Concen: 43.554 ng/ul
RT: 16.316 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:248 Resp: 855374
Ion Ratio Lower Upper
248 100
250 98.6 78.6 117.8
141 76.0 61.5 92.3



#69
Hexachlorobenzene
Concen: 42.198 ng/ul
RT: 16.416 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:284 Resp: 992531
Ion Ratio Lower Upper
284 100
142 34.2 28.5 42.7
249 30.5 22.1 33.1





#70

Atrazine

Concen: 19.043 ng/ul

RT: 16.592 min Scan# 2296

Delta R.T. 0.000 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38

Instrument :

BNA_N

ClientSampleId :

E0737MS





Tgt Ion:200 Resp: 364972

Ion Ratio Lower Upper

200 100

173 26.9 21.9 32.9

215 48.3 37.7 56.5



#71

Pentachlorophenol

Concen: 37.064 ng/ul

RT: 16.757 min Scan# 2324

Delta R.T. 0.000 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38




Tgt Ion:266 Resp: 537217


Ion Ratio Lower Upper

266 100

264 64.3 50.7 76.1

268 63.4 52.2 78.2









#72
Phenanthrene
Concen: 41.772 ng/u1
RT: 17.157 min Scan# 2392
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:178 Resp: 4535894
Ion Ratio Lower Upper
178 100
179 15.7 12.2 18.2
176 19.5 14.8 22.2



#74
Anthracene
Concen: 40.801 ng/u1
RT: 17.251 min Scan# 2408
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:178 Resp: 4497612
Ion Ratio Lower Upper
178 100
179 15.5 11.6 17.4
176 18.0 14.1 21.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 39.280 ng/uL
RT: 13.157 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:216 Resp: 1053026
Ion Ratio Lower Upper
216 100
214 77.4 61.1 91.7
218 47.1 39.0 58.4



#76
Pentachlorobenzene
Concen: 39.848 ng/uL
RT: 14.681 min Scan# 1971
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:250 Resp: 1051442
Ion Ratio Lower Upper
250 100
248 65.2 50.1 75.1
252 64.8 51.3 76.9









#77
 Carbazole
 Concen: 43.964 ng/ul
 RT: 17.522 min Scan# 2453
 Delta R.T. 0.000 min
 Lab File: BN030445.D
 Acq: 21 Mar 2024 13:38

Instrument :
 BNA_N
 ClientSampleId :
 E0737MS

Tgt Ion:167 Resp: 4301203




Ion	Ratio	Lower	Upper
167	100		
166	22.0	16.3	24.5
139	13.2	10.2	15.4




#78
 Di-n-butylphthalate
 Concen: 43.325 ng/ul
 RT: 18.104 min Scan# 2553
 Delta R.T. -0.006 min
 Lab File: BN030445.D
 Acq: 21 Mar 2024 13:38

Tgt Ion:149 Resp: 5330301

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.2	10.8
104	5.0	4.1	6.1









#80
Fluoranthene
Concen: 40.458 ng/u1
RT: 19.181 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:202 Resp: 5250380
Ion Ratio Lower Upper
202 100
101 12.3 10.2 15.2
100 9.4 7.7 11.5



#82
Pyrene
Concen: 40.378 ng/u1
RT: 19.545 min Scan# 2798
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:202 Resp: 5414445
Ion Ratio Lower Upper
202 100
101 14.7 11.5 17.3
100 12.2 9.3 13.9









#83
Butylbenzylphthalate
Concen: 44.145 ng/ul
RT: 20.463 min Scan# 2954
Delta R.T. -0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:149 Resp: 2278073
Ion Ratio Lower Upper
149 100
91 75.5 57.4 86.0
206 20.2 16.3 24.5



#84
3,3'-Dichlorobenzidine
Concen: 40.900 ng/ul
RT: 21.269 min Scan# 3091
Delta R.T. -0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:252 Resp: 1619449
Ion Ratio Lower Upper
252 100
254 63.6 53.9 80.9
126 13.5 10.1 15.1





#85

Benzo(a)anthracene

Concen: 39.967 ng/ul

RT: 21.345 min Scan# 3104

Delta R.T. 0.000 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38

Instrument :

BNA_N

ClientSampleId :

E0737MS





Tgt Ion:228 Resp: 5229795

Ion Ratio Lower Upper

228 100

229 18.5 14.7 22.1

226 27.8 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 42.728 ng/ul

RT: 21.298 min Scan# 3096

Delta R.T. -0.006 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38




Tgt Ion:149 Resp: 3378306


Ion Ratio Lower Upper

149 100

167 28.6 21.3 31.9

279 4.2 3.4 5.2





#87

Chrysene

Concen: 40.865 ng/ul

RT: 21.404 min Scan# 3114

Delta R.T. 0.000 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38

Instrument :

BNA_N

ClientSampleId :

E0737MS





Tgt Ion:228 Resp: 4941285

Ion Ratio Lower Upper

228 100

226 30.5 22.3 33.5

229 20.0 15.2 22.8



#89

Di-n-octyl phthalate

Concen: 42.679 ng/ul




RT: 22.439 min Scan# 3290


Delta R.T. -0.006 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38

Tgt Ion:149 Resp: 5565933









#90
Benzo(b)fluoranthene
Concen: 44.328 ng/ul
RT: 23.392 min Scan# 3452
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38




Instrument :
BNA_N
ClientSampleId :
E0737MS


Tgt Ion:252 Resp: 5228206
Ion Ratio Lower Upper
252 100
253 21.9 17.2 25.8
125 11.4 9.0 13.4



#91
Benzo(k)fluoranthene
Concen: 43.335 ng/ul
RT: 23.457 min Scan# 3463
Delta R.T. -0.006 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:252 Resp: 5162736
Ion Ratio Lower Upper
252 100
253 20.5 16.9 25.3
125 11.4 9.5 14.3





#93

Benzo(a)pyrene

Concen: 42.924 ng/ul

RT: 24.192 min Scan# 3588

Delta R.T. -0.006 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38

Instrument :

BNA_N

ClientSampleId :

E0737MS





Tgt Ion:252 Resp: 4737360

Ion Ratio Lower Upper

252 100

253 20.7 17.8 26.8

125 11.9 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 43.692 ng/ul

RT: 27.680 min Scan# 4181

Delta R.T. -0.006 min

Lab File: BN030445.D

Acq: 21 Mar 2024 13:38




Tgt Ion:276 Resp: 5191984


Ion Ratio Lower Upper

276 100

138 22.4 17.3 25.9

277 23.8 19.3 28.9









#95
Dibenzo(a,h)anthracene
Concen: 43.875 ng/ul
RT: 27.751 min Scan# 4193
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Instrument :
BNA_N
ClientSampleId :
E0737MS

Tgt Ion:278 Resp: 4356253

Ion	Ratio	Lower	Upper
278	100		
139	18.0	15.0	22.4
279	22.0	19.2	28.8



#96
Benzo(g,h,i)perylene
Concen: 43.087 ng/ul
RT: 28.721 min Scan# 4358
Delta R.T. 0.000 min
Lab File: BN030445.D
Acq: 21 Mar 2024 13:38

Tgt Ion:276 Resp: 4040370

Ion	Ratio	Lower	Upper
276	100		
138	23.6	19.4	29.2
277	23.9	17.7	26.5

