

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032125\
 Data File : BN036704.D
 Acq On : 21 Mar 2025 13:51
 Operator : RC/JU
 Sample : PB167209BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK209

Quant Time: Mar 21 14:28:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN022825.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 20 13:16:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	87145	20.000	ng/u1	0.00
20) Naphthalene-d8	10.487	136	356629	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.345	164	242145	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.092	188	495016	20.000	ng/u1	0.00
79) Chrysene-d12	21.327	240	520331	20.000	ng/u1	0.00
88) Perylene-d12	24.280	264	532581	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	16692	7.447	ng/uL	0.00
4) Pyridine-d5	3.599	84	243055	35.474	ng/u1	0.00
7) Phenol-d5	6.875	99	286313	34.399	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.040	67	188540	31.839	ng/u1	0.00
11) 2-Chlorophenol-d4	7.234	132	210998	36.052	ng/u1	0.00
15) 4-Methylphenol-d8	8.410	113	224082	32.606	ng/u1	0.00
21) Nitrobenzene-d5	8.857	128	99498	35.295	ng/u1	0.00
24) 2-Nitrophenol-d4	9.575	143	104120	34.959	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.110	165	193186	35.762	ng/u1	0.00
31) 4-Chloroaniline-d4	10.628	131	273901	32.834	ng/u1	0.00
46) Dimethylphthalate-d6	13.757	166	658311	36.150	ng/u1	0.00
49) Acenaphthylene-d8	14.039	160	689274	33.884	ng/u1	0.00
54) 4-Nitrophenol-d4	14.551	143	109215	30.277	ng/u1	0.00
60) Fluorene-d10	15.339	176	529723	34.488	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.463	200	99815	29.467	ng/u1	0.00
73) Anthracene-d10	17.192	188	861099	35.350	ng/u1	0.00
81) Pyrene-d10	19.492	212	1036331	36.458	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.086	264	971567	36.612	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032125\
Data File : BN036704.D
Acq On : 21 Mar 2025 13:51
Operator : RC/JU
Sample : PB167209BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK209

Quant Time: Mar 21 14:28:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN022825.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 20 13:16:23 2025
Response via : Initial Calibration

