

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_N\DATA\BN032218\
 Data File : BN000553.D
 Acq On : 22 Mar 2018 22:45
 Operator : JU/SJ
 Sample : J1905-10DL 50X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DAM45DL

Quant Time: Mar 23 08:24:11 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 23 02:22:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.41	152	163570	20.00	ng/ul	0.00
18) Naphthalene-d8	11.25	136	708410	20.00	ng/ul	0.00
35) Acenaphthene-d10	15.03	164	400450	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.76	188	860678	20.00	ng/ul	0.00
75) Chrysene-d12	21.88	240	853984	20.00	ng/ul	-0.01
83) Perylene-d12	24.36	264	810179	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	7.56	99	2135	0.14	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.72	67	3226	0.37	ng/ul	0.00
9) 2-Chlorophenol-d4	7.94	132	2362	0.20	ng/ul	0.00
13) 4-Methylphenol-d8	9.13	113	2100	0.18	ng/ul	0.00
19) Nitrobenzene-d5	9.60	128	1542	0.27	ng/ul	0.00
22) 2-Nitrophenol-d4	10.33	143	1188	0.20	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.87	165	2379	0.22	ng/ul	0.00
29) 4-Chloroaniline-d4	11.35	131	4375	0.40	ng/ul	-0.04
43) Dimethylphthalate-d6	14.42	166	8771	0.28	ng/ul	0.00
46) Acenaphthylene-d8	14.73	160	11088	0.27	ng/ul	0.00
51) 4-Nitrophenol-d4	15.23	143	2476	0.42	ng/ul	0.00
57) Fluorene-d10	16.01	176	10391	0.39	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	16.14	200	3545	0.72	ng/ul	0.00
70) Anthracene-d10	17.85	188	13641	0.33	ng/ul	0.00
76) Pyrene-d10	20.11	212	12563	0.27	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.20	264	10476	0.27	ng/ul	-0.01

Target Compounds

					Qvalue	
34) 2-Methylnaphthalene	12.88	142	26618	1.072	ng/ul	100
55) 2,3,4,6-Tetrachlorophenol	15.65	232	130019	20.010	ng/ul#	92
68) Pentachlorophenol	17.42	266	1564015	319.819	ng/ul	97
69) Phenanthrene	17.80	178	59027	1.190	ng/ul	97
74) Fluoranthene	19.78	202	59052	1.199	ng/ul#	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_N\DATA\BN032218\
Data File : BN000553.D
Acq On : 22 Mar 2018 22:45
Operator : JU/SJ
Sample : J1905-10DL 50X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DAM45DL

Quant Time: Mar 23 08:24:11 2018
Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 23 02:22:02 2018
Response via : Initial Calibration

