

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_N\DATA\BN032218\
 Data File : BN000563.D
 Acq On : 23 Mar 2018 04:28
 Operator : JU/SJ
 Sample : J1905-03DL2 200X
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DAM40DL2

Quant Time: Mar 23 08:47:06 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 23 06:10:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.41	152	174608	20.00	ng/ul	0.00
18) Naphthalene-d8	11.25	136	741265	20.00	ng/ul	0.00
35) Acenaphthene-d10	15.03	164	416684	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.76	188	901417	20.00	ng/ul	0.00
75) Chrysene-d12	21.88	240	879549	20.00	ng/ul	-0.01
83) Perylene-d12	24.36	264	763590	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	0.00	99	0d	0.00	ng/ul	
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0d	0.00	ng/ul	
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul	
13) 4-Methylphenol-d8	0.00	113	0d	0.00	ng/ul	
19) Nitrobenzene-d5	9.63	128	798	0.13	ng/ul	0.03
22) 2-Nitrophenol-d4	10.35	143	1499	0.24	ng/ul	0.03
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul	
29) 4-Chloroaniline-d4	11.35	131	7588	0.66	ng/ul	-0.04
43) Dimethylphthalate-d6	14.42	166	339	0.01	ng/ul	0.00
46) Acenaphthylene-d8	14.74	160	525	0.01	ng/ul	0.00
51) 4-Nitrophenol-d4	15.22	143	2036	0.33	ng/ul	0.00
57) Fluorene-d10	16.01	176	3478	0.12	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	16.15	200	817	0.16	ng/ul	0.01
70) Anthracene-d10	17.85	188	3009	0.07	ng/ul	0.00
76) Pyrene-d10	20.11	212	946	0.02	ng/ul	0.00
87) Benzo(a)pyrene-d12	0.00	264	0d	0.00	ng/ul	

Target Compounds

					Qvalue
34) 2-Methylnaphthalene	12.88	142	59352	2.284 ng/ul	99
55) 2,3,4,6-Tetrachlorophenol	15.65	232	63401	9.377 ng/ul#	91
68) Pentachlorophenol	17.41	266	685676	133.874 ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_N\DATA\BN032218\
Data File : BN000563.D
Acq On : 23 Mar 2018 04:28
Operator : JU/SJ
Sample : J1905-03DL2 200X
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DAM40DL2

Quant Time: Mar 23 08:47:06 2018
Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 23 06:10:38 2018
Response via : Initial Calibration

