

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_N\Data\BN032218\
 Data File : BN000566.D
 Acq On : 23 Mar 2018 06:11
 Operator : JU/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02029

Quant Time: Mar 23 08:38:28 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 23 06:10:38 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.603	0.563	6.6	85	0.00
3 S	1,4-Dioxane-d8	0.560	0.533	4.8	86	0.00
4	Benzaldehyde	0.761	0.798	-4.9	81	0.00
5 S	Phenol-d5	1.866	1.886	-1.1	84	0.00
6	Phenol	1.904	1.920	-0.8	83	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.076	1.076	0.0	83	0.00
8	Bis(2-Chloroethyl)ether	1.511	1.513	-0.1	83	0.00
9 S	2-Chlorophenol-d4	1.411	1.402	0.6	84	0.00
10	2-Chlorophenol	1.440	1.425	1.0	84	0.00
11	2-Methylphenol	1.396	1.419	-1.6	83	0.00
12	2,2'-oxybis(1-Chloropropane	1.628	1.649	-1.3	83	0.00
13 S	4-Methylphenol-d8	1.453	1.489	-2.5	83	0.00
14	Acetophenone	2.160	2.267	-5.0	83	0.00
15 P	N-Nitroso-di-n-propylamine	1.130	1.129	0.1	81	0.00
16	4-Methylphenol	1.495	1.533	-2.5	83	0.00
17	Hexachloroethane	0.601	0.584	2.8	84	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
19 S	Nitrobenzene-d5	0.162	0.155	4.3	82	0.00
20	Nitrobenzene	0.401	0.393	2.0	84	0.00
21	Isophorone	0.764	0.747	2.2	81	0.00
22 S	2-Nitrophenol-d4	0.168	0.163	3.0	82	0.00
23 C	2-Nitrophenol	0.181	0.178	1.7	83	0.00
24	2,4-Dimethylphenol	0.390	0.383	1.8	82	0.00
25	Bis(2-Chloroethoxy)methane	0.483	0.472	2.3	83	0.00
26 S	2,4-Dichlorophenol-d3	0.302	0.302	0.0	84	0.00
27 C	2,4-Dichlorophenol	0.293	0.295	-0.7	84	0.00
28	Naphthalene	1.050	1.032	1.7	84	0.00
29 S	4-Chloroaniline-d4	0.310	0.390	-25.8#	81	0.00
30	4-Chloroaniline	0.314	0.396	-26.1#	81	0.00
31 C	Hexachlorobutadiene	0.162	0.162	0.0	86	0.00
32	Caprolactam	0.107	0.110	-2.8	79	0.00
33 C	4-Chloro-3-methylphenol	0.344	0.358	-4.1	84	0.00
34	2-Methylnaphthalene	0.701	0.704	-0.4	84	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	83	0.00
36	1,2,4,5-Tetrachlorobenzene	0.601	0.580	3.5	85	0.00
37	Hexachlorocyclopentadiene	0.313	0.327	-4.5	101	0.00
38 C	2,4,6-Trichlorophenol	0.408	0.395	3.2	83	0.00
39	2,4,5-Trichlorophenol	0.425	0.411	3.3	83	0.00
40	1,1'-Biphenyl	1.703	1.652	3.0	84	0.00
41	2-Chloronaphthalene	1.312	1.271	3.1	85	0.00
42	2-Nitroaniline	0.398	0.393	1.3	83	0.00
43 S	Dimethylphthalate-d6	1.580	1.582	-0.1	84	0.00
44	Dimethylphthalate	1.563	1.564	-0.1	83	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_N\Data\BN032218\
 Data File : BN000566.D
 Acq On : 23 Mar 2018 06:11
 Operator : JU/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02029

Quant Time: Mar 23 08:38:28 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 23 06:10:38 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.315	0.319	-1.3	83	0.00
46 S	Acenaphthylene-d8	2.084	2.033	2.4	83	0.00
47	Acenaphthylene	2.042	1.977	3.2	83	0.00
48	3-Nitroaniline	0.304	0.336	-10.5	82	0.00
49 C	Acenaphthene	1.392	1.360	2.3	83	0.00
50	2,4-Dinitrophenol	0.149	0.151	-1.3	93	0.00
51 S	4-Nitrophenol-d4	0.293	0.274	6.5	76	0.00
52	4-Nitrophenol	0.227	0.222	2.2	79	0.00
53	Dibenzofuran	1.891	1.865	1.4	84	0.00
54	2,4-Dinitrotoluene	0.447	0.463	-3.6	84	0.00
55	2,3,4,6-Tetrachlorophenol	0.325	0.323	0.6	82	0.00
56	Diethylphthalate	1.595	1.613	-1.1	83	0.00
57 S	Fluorene-d10	1.346	1.344	0.1	84	0.00
58	Fluorene	1.503	1.495	0.5	84	0.00
59	4-Chlorophenyl-phenylether	0.694	0.699	-0.7	85	0.00
60	4-Nitroaniline	0.342	0.347	-1.5	85	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.115	0.093	19.1	71	0.00
63	4,6-Dinitro-2-methylphenol	0.121	0.098	19.0	71	0.00
64	N-Nitrosodiphenylamine	0.659	0.625	5.2	83	0.00
65	4-Bromophenyl-phenylether	0.204	0.199	2.5	86	0.00
66	Hexachlorobenzene	0.215	0.215	0.0	88	0.00
67	Atrazine	0.199	0.215	-8.0	84	0.00
68 C	Pentachlorophenol	0.114	0.111	2.6	86	0.00
69	Phenanthrene	1.152	1.116	3.1	84	0.00
70 S	Anthracene-d10	0.961	0.931	3.1	84	0.00
71	Anthracene	1.181	1.136	3.8	83	0.00
72	Carbazole	1.013	1.026	-1.3	83	0.00
73	Di-n-butylphthalate	1.378	1.335	3.1	82	0.00
74 C	Fluoranthene	1.145	1.201	-4.9	84	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	78	0.00
76 S	Pyrene-d10	1.104	1.138	-3.1	84	0.00
77	Pyrene	1.461	1.459	0.1	83	0.00
78	Butylbenzylphthalate	0.726	0.727	-0.1	81	0.00
79	3,3'-Dichlorobenzidine	0.391	0.341	12.8	65	0.00
80	Benzo(a)anthracene	1.333	1.256	5.8	78	0.00
81	Bis(2-ethylhexyl)phthalate	1.053	1.032	2.0	79	0.00
82	Chrysene	1.249	1.163	6.9	77	0.00
83 I	Perylene-d12	1.000	1.000	0.0	64	0.00
84	Di-n-octyl phthalate	1.683	2.071	-23.1#	75	0.00
85	Benzo(b)fluoranthene	1.264	1.327	-5.0	72	0.00
86	Benzo(k)fluoranthene	1.220	1.207	1.1	65	0.00
87 S	Benzo(a)pyrene-d12	0.944	0.945	-0.1	66	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_N\Data\BN032218\
Data File : BN000566.D
Acq On : 23 Mar 2018 06:11
Operator : JU/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTD02029

Quant Time: Mar 23 08:38:28 2018
Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 23 06:10:38 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.189	1.165	2.0	66	0.00
89	Indeno(1,2,3-cd)pyrene	1.282	1.182	7.8	65	0.00
90	Dibenzo(a,h)anthracene	1.068	0.983	8.0	65	0.00
91	Benzo(g,h,i)perylene	1.066	0.973	8.7	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0