

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019072.D
 Acq On : 23 Mar 2022 01:32
 Operator : CG/JU
 Sample : PB143519BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB143519BL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/23/2022
 Supervised By : Jagrut Upadhyay 03/23/2022

Quant Time: Mar 23 04:52:01 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	4138	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	9857	0.400	ng	# 0.00
13) Acenaphthene-d10	14.495	164	5414	0.400	ng	0.00
19) Phenanthrene-d10	17.236	188	9917	0.400	ng	0.00
29) Chrysene-d12	21.422	240	8316	0.400	ng	# 0.00
35) Perylene-d12	23.785	264	7395	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	3124	0.337	ng	0.00
5) Phenol-d6	7.017	99	2377	0.244	ng	0.00
8) Nitrobenzene-d5	9.014	82	2046m	0.288	ng	0.00
11) 2-Methylnaphthalene-d10	12.253	152	4858	0.297	ng	0.00
14) 2,4,6-Tribromophenol	15.995	330	459	0.176	ng	0.01
15) 2-Fluorobiphenyl	13.127	172	7093	0.327	ng	0.00
27) Fluoranthene-d10	19.265	212	10895	0.370	ng	0.00
31) Terphenyl-d14	19.860	244	6511	0.319	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
Data File : BN019072.D
Acq On : 23 Mar 2022 01:32
Operator : CG/JU
Sample : PB143519BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB143519BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/23/2022
Supervised By :Jagrut Upadhyay 03/23/2022

Quant Time: Mar 23 04:52:01 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

