

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019074.D
 Acq On : 23 Mar 2022 02:44
 Operator : CG/JU
 Sample : N2041-12
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D1-20220317

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/23/2022
 Supervised By : Jagrut Upadhyay 03/23/2022

Quant Time: Mar 23 04:52:11 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3847	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	10386	0.400	ng	# 0.00
13) Acenaphthene-d10	14.484	164	6223	0.400	ng	-0.01
19) Phenanthrene-d10	17.225	188	13222	0.400	ng	-0.01
29) Chrysene-d12	21.413	240	10294	0.400	ng	# 0.00
35) Perylene-d12	23.785	264	8220	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	1076	0.125	ng	0.00
5) Phenol-d6	7.017	99	561	0.062	ng	0.00
8) Nitrobenzene-d5	9.014	82	1498m	0.200	ng	0.00
11) 2-Methylnaphthalene-d10	12.246	152	5060	0.294	ng	0.00
14) 2,4,6-Tribromophenol	15.974	330	1004	0.334	ng	-0.01
15) 2-Fluorobiphenyl	13.116	172	7875	0.316	ng	-0.01
27) Fluoranthene-d10	19.261	212	14819	0.378	ng	0.00
31) Terphenyl-d14	19.855	244	8596	0.340	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	30648	6.446	ng	98
34) Bis(2-ethylhexyl)phtha...	21.324	149	9084	0.432	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
Data File : BN019074.D
Acq On : 23 Mar 2022 02:44
Operator : CG/JU
Sample : N2041-12
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D1-20220317

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 03/23/2022
Supervised By : Jagrut Upadhyay 03/23/2022

Quant Time: Mar 23 04:52:11 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

