

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019075.D
 Acq On : 23 Mar 2022 03:20
 Operator : CG/JU
 Sample : N2041-13
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE115D1-20220317

Quant Time: Mar 23 04:52:15 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/23/2022
 Supervised By :Jagrut Upadhyay 03/23/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3855	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	9741	0.400	ng	# 0.00
13) Acenaphthene-d10	14.485	164	5890	0.400	ng	-0.01
19) Phenanthrene-d10	17.236	188	11563	0.400	ng	0.00
29) Chrysene-d12	21.413	240	9418	0.400	ng	# 0.00
35) Perylene-d12	23.785	264	7795	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	982	0.114	ng	0.00
5) Phenol-d6	7.024	99	542	0.060	ng	0.01
8) Nitrobenzene-d5	9.014	82	1399m	0.199	ng	0.00
11) 2-Methylnaphthalene-d10	12.250	152	4004	0.248	ng	0.00
14) 2,4,6-Tribromophenol	15.985	330	635	0.223	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	6092	0.258	ng	-0.01
27) Fluoranthene-d10	19.261	212	13031	0.380	ng	0.00
31) Terphenyl-d14	19.856	244	7810	0.338	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	18061	3.791	ng	97
6) bis(2-Chloroethyl)ether	7.277	93	455	0.042	ng	97
25) Phenanthrene	17.277	178	890	0.025	ng	98
34) Bis(2-ethylhexyl)phtha...	21.324	149	3104	0.161	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
Data File : BN019075.D
Acq On : 23 Mar 2022 03:20
Operator : CG/JU
Sample : N2041-13
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE115D1-20220317

Quant Time: Mar 23 04:52:15 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/23/2022
Supervised By :Jagrut Upadhyay 03/23/2022

