

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019078.D
 Acq On : 23 Mar 2022 05:07
 Operator : CG/JU
 Sample : N2041-16
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE103D1-20220317

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/23/2022
 Supervised By : Jagrut Upadhyay 03/23/2022

Quant Time: Mar 23 06:26:47 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.855	152	3981	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	9851	0.400	ng	# 0.00
13) Acenaphthene-d10	14.485	164	5691	0.400	ng	-0.01
19) Phenanthrene-d10	17.236	188	11036	0.400	ng	0.00
29) Chrysene-d12	21.422	240	9019	0.400	ng	# 0.00
35) Perylene-d12	23.788	264	7335	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	1001	0.112	ng	0.00
5) Phenol-d6	7.024	99	466	0.050	ng	0.01
8) Nitrobenzene-d5	9.014	82	1737m	0.244	ng	0.00
11) 2-Methylnaphthalene-d10	12.250	152	4752	0.291	ng	0.00
14) 2,4,6-Tribromophenol	15.985	330	629	0.229	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	6905	0.303	ng	-0.01
27) Fluoranthene-d10	19.261	212	14071	0.430	ng	0.00
31) Terphenyl-d14	19.860	244	8215	0.371	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.247	88	28361	5.764	ng	96
34) Bis(2-ethylhexyl)phtha...	21.333	149	2955	0.160	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
Data File : BN019078.D
Acq On : 23 Mar 2022 05:07
Operator : CG/JU
Sample : N2041-16
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE103D1-20220317

Quant Time: Mar 23 06:26:47 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/23/2022
Supervised By :Jagrut Upadhyay 03/23/2022

