

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019080.D
 Acq On : 23 Mar 2022 06:18
 Operator : CG/JU
 Sample : N2041-18
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP02-20220317

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/23/2022
 Supervised By : Jagrut Upadhyay 03/23/2022

Quant Time: Mar 23 07:14:55 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3923	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	9740	0.400	ng	# 0.00
13) Acenaphthene-d10	14.484	164	5894	0.400	ng	-0.01
19) Phenanthrene-d10	17.236	188	11410	0.400	ng	0.00
29) Chrysene-d12	21.422	240	9379	0.400	ng	# 0.00
35) Perylene-d12	23.787	264	6688	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	943	0.107	ng	0.00
5) Phenol-d6	7.024	99	441	0.048	ng	0.01
8) Nitrobenzene-d5	9.014	82	1469m	0.209	ng	0.00
11) 2-Methylnaphthalene-d10	12.250	152	4121	0.255	ng	0.00
14) 2,4,6-Tribromophenol	15.984	330	607	0.213	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	6045	0.256	ng	-0.01
27) Fluoranthene-d10	19.261	212	12296	0.363	ng	0.00
31) Terphenyl-d14	19.855	244	7665	0.333	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	18431	3.801	ng	98
6) bis(2-Chloroethyl)ether	7.269	93	497	0.045	ng	94
34) Bis(2-ethylhexyl)phtha...	21.324	149	31572	1.648	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
Data File : BN019080.D
Acq On : 23 Mar 2022 06:18
Operator : CG/JU
Sample : N2041-18
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP02-20220317

Quant Time: Mar 23 07:14:55 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/23/2022
Supervised By :Jagrut Upadhyay 03/23/2022

