

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019082.D
 Acq On : 23 Mar 2022 07:29
 Operator : CG/JU
 Sample : N2041-20
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE114D2-20220318

Quant Time: Mar 23 08:12:04 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3863	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	9610	0.400	ng	# 0.00
13) Acenaphthene-d10	14.484	164	5774	0.400	ng	-0.01
19) Phenanthrene-d10	17.236	188	10765	0.400	ng	0.00
29) Chrysene-d12	21.422	240	9010	0.400	ng	# 0.00
35) Perylene-d12	23.785	264	6944	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	962	0.111	ng	0.00
5) Phenol-d6	7.024	99	458	0.050	ng	0.01
8) Nitrobenzene-d5	9.014	82	1421	0.205	ng	0.00
11) 2-Methylnaphthalene-d10	12.253	152	4273	0.268	ng	0.00
14) 2,4,6-Tribromophenol	15.984	330	526	0.189	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	5905	0.255	ng	-0.01
27) Fluoranthene-d10	19.261	212	12687	0.397	ng	0.00
31) Terphenyl-d14	19.860	244	7213	0.326	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	12176	2.550	ng	98
34) Bis(2-ethylhexyl)phtha...	21.324	149	2587	0.141	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
Data File : BN019082.D
Acq On : 23 Mar 2022 07:29
Operator : CG/JU
Sample : N2041-20
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE114D2-20220318

Quant Time: Mar 23 08:12:04 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

