

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019096.D
 Acq On : 23 Mar 2022 17:48
 Operator : CG/JU
 Sample : N2041-12DL 2X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D1-20220317DL

Quant Time: Mar 24 05:31:58 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/24/2022
 Supervised By :Jagrut Upadhyay 03/24/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	4554	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	12273	0.400	ng	# 0.00
13) Acenaphthene-d10	14.484	164	7304	0.400	ng	-0.01
19) Phenanthrene-d10	17.225	188	15256	0.400	ng	-0.01
29) Chrysene-d12	21.422	240	11237	0.400	ng	# 0.00
35) Perylene-d12	23.784	264	8913	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	503	0.049	ng	0.00
5) Phenol-d6	7.031	99	222	0.021	ng	0.02
8) Nitrobenzene-d5	9.025	82	653m	0.074	ng	0.01
11) 2-Methylnaphthalene-d10	12.246	152	2585	0.127	ng	0.00
14) 2,4,6-Tribromophenol	15.974	330	478	0.136	ng	-0.01
15) 2-Fluorobiphenyl	13.116	172	3994	0.136	ng	-0.01
27) Fluoranthene-d10	19.261	212	7140	0.158	ng	0.00
31) Terphenyl-d14	19.855	244	4205	0.152	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	15976	2.839	ng	98
34) Bis(2-ethylhexyl)phtha...	21.333	149	3648	0.159	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
Data File : BN019096.D
Acq On : 23 Mar 2022 17:48
Operator : CG/JU
Sample : N2041-12DL 2X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D1-20220317DL

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 03/24/2022
Supervised By : Jagrut Upadhyay 03/24/2022

Quant Time: Mar 24 05:31:58 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

