

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019099.D
 Acq On : 23 Mar 2022 19:36
 Operator : CG/JU
 Sample : N2074-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT180D1-20220318

Quant Time: Mar 24 05:32:12 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	4222	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	10862	0.400	ng	# 0.00
13) Acenaphthene-d10	14.484	164	6452	0.400	ng	-0.01
19) Phenanthrene-d10	17.225	188	13754	0.400	ng	-0.01
29) Chrysene-d12	21.422	240	9757	0.400	ng	# 0.00
35) Perylene-d12	23.784	264	7763	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.413	112	962	0.102	ng	0.00
5) Phenol-d6	7.009	99	561	0.057	ng	0.00
8) Nitrobenzene-d5	9.014	82	1522	0.194	ng	0.00
11) 2-Methylnaphthalene-d10	12.250	152	4521	0.251	ng	0.00
14) 2,4,6-Tribromophenol	15.984	330	686	0.220	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	6532	0.253	ng	-0.01
27) Fluoranthene-d10	19.261	212	16232	0.398	ng	0.00
31) Terphenyl-d14	19.855	244	9573	0.399	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.247	88	9924	1.902	ng	97
25) Phenanthrene	17.266	178	1416	0.033	ng	99
28) Fluoranthene	19.291	202	1908	0.038	ng	96
30) Pyrene	19.653	202	1195	0.025	ng	97
34) Bis(2-ethylhexyl)phtha...	21.324	149	1665	0.084	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
Data File : BN019099.D
Acq On : 23 Mar 2022 19:36
Operator : CG/JU
Sample : N2074-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT180D1-20220318

Quant Time: Mar 24 05:32:12 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

