

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019101.D
 Acq On : 23 Mar 2022 20:48
 Operator : CG/JU
 Sample : N2074-04
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT180D2-20220318

Quant Time: Mar 24 05:32:25 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	4287	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	10935	0.400	ng	# 0.00
13) Acenaphthene-d10	14.484	164	6565	0.400	ng	-0.01
19) Phenanthrene-d10	17.225	188	12801	0.400	ng	-0.01
29) Chrysene-d12	21.422	240	9950	0.400	ng	# 0.00
35) Perylene-d12	23.784	264	7888	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	982	0.102	ng	0.00
5) Phenol-d6	7.009	99	578	0.057	ng	0.00
8) Nitrobenzene-d5	9.014	82	1593	0.202	ng	0.00
11) 2-Methylnaphthalene-d10	12.249	152	4546	0.251	ng	0.00
14) 2,4,6-Tribromophenol	15.984	330	848	0.268	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	7043	0.268	ng	-0.01
27) Fluoranthene-d10	19.261	212	16074	0.423	ng	0.00
31) Terphenyl-d14	19.855	244	11160	0.457	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.246	88	16501	3.114	ng	98
6) bis(2-Chloroethyl)ether	7.269	93	378	0.032	ng	97
25) Phenanthrene	17.266	178	1091	0.027	ng	96
28) Fluoranthene	19.291	202	1280	0.027	ng	95
34) Bis(2-ethylhexyl)phtha...	21.324	149	1028	0.051	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
Data File : BN019101.D
Acq On : 23 Mar 2022 20:48
Operator : CG/JU
Sample : N2074-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT180D2-20220318

Quant Time: Mar 24 05:32:25 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

