

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019103.D
 Acq On : 23 Mar 2022 21:59
 Operator : CG/JU
 Sample : N2074-06
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE131D3-20220319

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/24/2022
 Supervised By : Jagrut Upadhyay 03/24/2022

Quant Time: Mar 24 05:32:38 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3941	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	9565	0.400	ng	# 0.00
13) Acenaphthene-d10	14.485	164	5732	0.400	ng	-0.01
19) Phenanthrene-d10	17.236	188	11232	0.400	ng	0.00
29) Chrysene-d12	21.422	240	8635	0.400	ng	# 0.00
35) Perylene-d12	23.788	264	6946	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	915	0.104	ng	0.00
5) Phenol-d6	7.024	99	377	0.041	ng	0.01
8) Nitrobenzene-d5	9.025	82	1369m	0.198	ng	0.01
11) 2-Methylnaphthalene-d10	12.254	152	4045	0.255	ng	0.00
14) 2,4,6-Tribromophenol	15.985	330	511	0.185	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	6139	0.267	ng	-0.01
27) Fluoranthene-d10	19.261	212	13579	0.407	ng	0.00
31) Terphenyl-d14	19.860	244	8137	0.384	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	6608	1.357	ng	96
34) Bis(2-ethylhexyl)phtha...	21.324	149	806	0.046	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
Data File : BN019103.D
Acq On : 23 Mar 2022 21:59
Operator : CG/JU
Sample : N2074-06
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE131D3-20220319

Quant Time: Mar 24 05:32:38 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 03/24/2022
Supervised By : Jagrut Upadhyay 03/24/2022

