

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019117.D
 Acq On : 24 Mar 2022 08:07
 Operator : CG/JU
 Sample : N2074-16
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BPOW1-7-20220321

Quant Time: Mar 24 08:58:13 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/24/2022
 Supervised By : Jagrut Upadhyay 03/24/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	4033	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	10471	0.400	ng	# 0.00
13) Acenaphthene-d10	14.485	164	6373	0.400	ng	-0.01
19) Phenanthrene-d10	17.226	188	13496	0.400	ng	-0.01
29) Chrysene-d12	21.422	240	9654	0.400	ng	# 0.00
35) Perylene-d12	23.785	264	7697	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	885	0.098	ng	0.00
5) Phenol-d6	7.009	99	553	0.058	ng	0.00
8) Nitrobenzene-d5	9.014	82	1583m	0.210	ng	0.00
11) 2-Methylnaphthalene-d10	12.246	152	4489	0.258	ng	0.00
14) 2,4,6-Tribromophenol	15.974	330	655	0.213	ng	-0.01
15) 2-Fluorobiphenyl	13.116	172	6687	0.262	ng	-0.01
27) Fluoranthene-d10	19.261	212	16415	0.410	ng	0.00
31) Terphenyl-d14	19.856	244	9232	0.389	ng	0.00
Target Compounds						
25) Phenanthrene	17.267	178	1773	0.042	ng	96
28) Fluoranthene	19.291	202	2160	0.044	ng	99
30) Pyrene	19.653	202	1255	0.026	ng	97
34) Bis(2-ethylhexyl)phtha...	21.324	149	2288	0.116	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
Data File : BN019117.D
Acq On : 24 Mar 2022 08:07
Operator : CG/JU
Sample : N2074-16
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BPOW1-7-20220321

Quant Time: Mar 24 08:58:13 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 03/24/2022
Supervised By : Jagrut Upadhyay 03/24/2022

