

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN032222\  
 Data File : BN019119.D  
 Acq On : 24 Mar 2022 09:19  
 Operator : CG/JU  
 Sample : N2074-19  
 Misc :  
 ALS Vial : 30 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 TT101D2-20220321

Quant Time: Mar 24 14:32:02 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN031722.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Mar 17 13:18:23 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| -----                         |        |      |          |       |       |          |
| Internal Standards            |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.854  | 152  | 3813     | 0.400 | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.658 | 136  | 9380     | 0.400 | ng    | # 0.00   |
| 13) Acenaphthene-d10          | 14.484 | 164  | 5514     | 0.400 | ng    | -0.01    |
| 19) Phenanthrene-d10          | 17.236 | 188  | 9823     | 0.400 | ng    | 0.00     |
| 29) Chrysene-d12              | 21.422 | 240  | 7796     | 0.400 | ng    | # 0.00   |
| 35) Perylene-d12              | 23.793 | 264  | 6171     | 0.400 | ng    | # 0.00   |
| System Monitoring Compounds   |        |      |          |       |       |          |
| 4) 2-Fluorophenol             | 5.406  | 112  | 802      | 0.094 | ng    | 0.00     |
| 5) Phenol-d6                  | 7.031  | 99   | 253      | 0.028 | ng    | 0.02     |
| 8) Nitrobenzene-d5            | 9.025  | 82   | 1320     | 0.195 | ng    | 0.01     |
| 11) 2-Methylnaphthalene-d10   | 12.257 | 152  | 3871     | 0.249 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 15.995 | 330  | 398      | 0.150 | ng    | 0.01     |
| 15) 2-Fluorobiphenyl          | 13.127 | 172  | 4891     | 0.221 | ng    | 0.00     |
| 27) Fluoranthene-d10          | 19.265 | 212  | 12514    | 0.429 | ng    | 0.00     |
| 31) Terphenyl-d14             | 19.860 | 244  | 8227     | 0.430 | ng    | 0.00     |
| Target Compounds              |        |      |          |       |       |          |
| 2) 1,4-Dioxane                | 3.247  | 88   | 11783    | 2.500 | ng    | # 93     |
| 6) bis(2-Chloroethyl)ether    | 7.277  | 93   | 230      | 0.022 | ng    | 94       |
| 34) Bis(2-ethylhexyl)phtha... | 21.333 | 149  | 1189     | 0.075 | ng    | # 99     |
| -----                         |        |      |          |       |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN032222\  
Data File : BN019119.D  
Acq On : 24 Mar 2022 09:19  
Operator : CG/JU  
Sample : N2074-19  
Misc :  
ALS Vial : 30 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
TT101D2-20220321

Quant Time: Mar 24 14:32:02 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN031722.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Mar 17 13:18:23 2022  
Response via : Initial Calibration

