

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032222\
 Data File : BN019089.D
 Acq On : 23 Mar 2022 13:32
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 16:45:34 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.494	0.510	-3.2	91	0.00
3	n-Nitrosodimethylamine	0.577	0.475	17.7	80	0.00
4 S	2-Fluorophenol	0.896	0.819	8.6	85	0.00
5 S	Phenol-d6	0.940	0.788	16.2	83	0.00
6	bis(2-Chloroethyl)ether	1.118	1.013	9.4	85	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.289	0.237	18.0	79	-0.01
9	Naphthalene	1.138	1.100	3.3	85	-0.01
10	Hexachlorobutadiene	0.278	0.263	5.4	82	-0.01
11 SURR	2-Methylnaphthalene-d10	0.664	0.577	13.1	76	0.00
12	2-Methylnaphthalene	0.774	0.691	10.7	79	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	75	-0.01
14 S	2,4,6-Tribromophenol	0.193	0.128	33.7#	55	0.00
15 S	2-Fluorobiphenyl	1.603	1.551	3.2	76	-0.01
16	Acenaphthylene	1.856	1.621	12.7	72	0.00
17	Acenaphthene	1.302	1.213	6.8	73	-0.01
18	Fluorene	1.641	1.505	8.3	70	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	71	-0.01
20	4,6-Dinitro-2-methylphenol	0.049	0.034	30.6#	54	0.01
21	4-Bromophenyl-phenylether	0.252	0.239	5.2	70	0.00
22	Hexachlorobenzene	0.325	0.337	-3.7	74	0.00
23	Atrazine	0.188	0.141	25.0#	59	-0.01
24	Pentachlorophenol	0.100	0.045	55.0#	37#	0.00
25	Phenanthrene	1.251	1.185	5.3	70	-0.01
26	Anthracene	1.057	0.886	16.2	65	-0.01
27 SURR	Fluoranthene-d10	1.187	1.139	4.0	69	0.00
28	Fluoranthene	1.468	1.420	3.3	70	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	72	0.00
30	Pyrene	1.993	1.925	3.4	70	0.00
31 S	Terphenyl-d14	0.982	0.831	15.4	63	0.00
32	Benzo(a)anthracene	1.233	1.060	14.0	69	0.00
33	Chrysene	1.662	1.649	0.8	71	0.00
34	Bis(2-ethylhexyl)phthalate	0.817	0.656	19.7	63	0.00
35 I	Perylene-d12	1.000	1.000	0.0	86	0.00
36	Indeno(1,2,3-cd)pyrene	1.443	1.522	-5.5	95	-0.01
37	Benzo(b)fluoranthene	1.459	1.362	6.6	88	0.00
38	Benzo(k)fluoranthene	2.000	1.856	7.2	80	0.00
39 C	Benzo(a)pyrene	1.328	1.278	3.8	85	0.00
40	Dibenzo(a,h)anthracene	1.124	1.090	3.0	87	-0.01
41	Benzo(g,h,i)perylene	1.417	1.524	-7.6	102	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4