

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032223\
 Data File : BN024486.D
 Acq On : 22 Mar 2023 19:02
 Operator : CG/JU
 Sample : 01949-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB937

Quant Time: Mar 23 03:10:57 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 22 14:57:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.030	152	9312	0.400	ng/ul	0.00
4) Naphthalene-d8	10.845	136	27171	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.666	164	16222	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.409	188	33354	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.592	240	26367	0.400	ng/ul	# 0.00
23) Perylene-d12	24.073	264	28321	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	51583	5.034	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.429	152	14367	0.445	ng/ul	0.00
18) Fluoranthene-d10	19.436	212	36075	0.514	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.431	88	15451	1.386	ng/ul#	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032223\
Data File : BN024486.D
Acq On : 22 Mar 2023 19:02
Operator : CG/JU
Sample : 01949-05
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
CB937

Quant Time: Mar 23 03:10:57 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 22 14:57:36 2023
Response via : Initial Calibration

