

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032223\
 Data File : BN024494.D
 Acq On : 23 Mar 2023 00:35
 Operator : CG/JU
 Sample : 01949-03
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB933

Quant Time: Mar 23 03:33:50 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 23 03:32:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.025	152	9524	0.400	ng/u1	0.00
4) Naphthalene-d8	10.839	136	27902	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.661	164	15575	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.409	188	33441	0.400	ng/u1	0.00
17) Chrysene-d12	21.589	240	26660	0.400	ng/u1	0.00
23) Perylene-d12	24.082	264	27075	0.400	ng/u1	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	49604	4.733	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.423	152	13463	0.406	ng/u1	0.00
18) Fluoranthene-d10	19.432	212	36597	0.516	ng/u1	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.431	88	565	0.050	ng/u1#	41


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032223\
Data File : BN024494.D
Acq On : 23 Mar 2023 00:35
Operator : CG/JU
Sample : 01949-03
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
CB933


Quant Time: Mar 23 03:33:50 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 23 03:32:22 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.050 ng/ul
 RT: 3.431 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BN024494.D
 Acq: 23 Mar 2023 00:35

Instrument :
 BNA_N
 ClientSampleId :
 CB933



Tgt Ion:	88	Resp:	565
Ion	Ratio	Lower	Upper
88	100		
43	127.2	32.1	48.1#
58	59.3	55.1	82.7

