

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032224\
 Data File : BN030457.D
 Acq On : 21 Mar 2024 21:29
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN032224

Quant Time: Mar 22 00:03:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN032224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 22 00:01:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	105	0.00
2	1,4-Dioxane	0.400	0.378	5.5	107	0.00
3	n-Nitrosodimethylamine	0.400	0.389	2.8	106	0.00
4 S	2-Fluorophenol	0.400	0.367	8.3	103	0.00
5 S	Phenol-d6	0.400	0.353	11.8	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.396	1.0	109	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	105	0.00
8 S	Nitrobenzene-d5	0.400	0.400	0.0	113	0.00
9	Naphthalene	0.400	0.388	3.0	107	0.00
10	Hexachlorobutadiene	0.400	0.372	7.0	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.396	1.0	111	0.00
12	2-Methylnaphthalene	0.400	0.387	3.3	107	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.400	0.314	21.5	96	0.00
15 S	2-Fluorobiphenyl	0.400	0.429	-7.2	114	0.00
16	Acenaphthylene	0.400	0.390	2.5	110	0.00
17	Acenaphthene	0.400	0.379	5.3	103	0.00
18	Fluorene	0.400	0.387	3.3	106	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.391	2.3	99	0.00
21	4-Bromophenyl-phenylether	0.400	0.372	7.0	105	0.00
22	Hexachlorobenzene	0.400	0.379	5.3	105	0.00
23	Atrazine	0.400	0.355	11.3	107	0.00
24	Pentachlorophenol	0.400	0.396	1.0	97	0.00
25	Phenanthrene	0.400	0.388	3.0	108	0.00
26	Anthracene	0.400	0.376	6.0	109	0.00
27 SURR	Fluoranthene-d10	0.400	0.385	3.8	109	0.00
28	Fluoranthene	0.400	0.369	7.8	105	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	104	0.00
30	Pyrene	0.400	0.383	4.3	104	0.00
31 S	Terphenyl-d14	0.400	0.434	-8.5	120	0.00
32	Benzo(a)anthracene	0.400	0.378	5.5	109	0.00
33	Chrysene	0.400	0.399	0.3	108	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.324	19.0	103	0.00
35 I	Perylene-d12	0.400	0.400	0.0	94	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.400	0.0	110	0.00
37	Benzo(b)fluoranthene	0.400	0.409	-2.2	105	0.00
38	Benzo(k)fluoranthene	0.400	0.407	-1.7	104	0.00
39 C	Benzo(a)pyrene	0.400	0.401	-0.3	109	0.00
40	Dibenzo(a,h)anthracene	0.400	0.394	1.5	108	-0.01
41	Benzo(g,h,i)perylene	0.400	0.382	4.5	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN032224