

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032224\
 Data File : BN030450.D
 Acq On : 21 Mar 2024 17:14
 Operator : MA/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Mar 21 23:54:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN032224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 21 23:54:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.731	152	10799	0.400	ng	0.00
7) Naphthalene-d8	10.509	136	29381	0.400	ng	0.00
13) Acenaphthene-d10	14.361	164	14980	0.400	ng	0.00
19) Phenanthrene-d10	17.106	188	34700	0.400	ng	0.00
29) Chrysene-d12	21.304	240	23694	0.400	ng	0.00
35) Perylene-d12	23.563	264	22770	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.312	112	2617	0.098	ng	0.00
5) Phenol-d6	6.886	99	3417	0.098	ng	0.00
8) Nitrobenzene-d5	8.875	82	2182	0.096	ng	0.00
11) 2-Methylnaphthalene-d10	12.106	152	4085	0.096	ng	0.00
14) 2,4,6-Tribromophenol	15.853	330	508	0.093	ng	0.00
15) 2-Fluorobiphenyl	12.993	172	6588	0.102	ng	0.00
27) Fluoranthene-d10	19.145	212	7797	0.090	ng	0.00
31) Terphenyl-d14	19.759	244	5636	0.106	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.261	88	1783	0.128	ng	# 84
3) n-Nitrosodimethylamine	3.564	42	1283	0.095	ng	# 94
6) bis(2-Chloroethyl)ether	7.161	93	3305	0.099	ng	99
9) Naphthalene	10.562	128	8332	0.098	ng	97
10) Hexachlorobutadiene	10.850	225	1568	0.099	ng	# 100
12) 2-Methylnaphthalene	12.181	142	5190	0.096	ng	98
16) Acenaphthylene	14.083	152	7384	0.096	ng	100
17) Acenaphthene	14.425	154	4881	0.095	ng	98
18) Fluorene	15.420	166	6519	0.095	ng	100
21) 4-Bromophenyl-phenylether	16.312	248	2193	0.098	ng	98
22) Hexachlorobenzene	16.411	284	2667	0.098	ng	99
23) Atrazine	16.585	200	1458	0.090	ng	95
25) Phenanthrene	17.156	178	10753	0.097	ng	100
26) Anthracene	17.243	178	8923	0.091	ng	99
28) Fluoranthene	19.178	202	11870	0.095	ng	99
30) Pyrene	19.540	202	12075	0.102	ng	100
32) Benzo(a)anthracene	21.295	228	9570	0.100	ng	98
33) Chrysene	21.340	228	10495	0.103	ng	97
34) Bis(2-ethylhexyl)phtha...	21.250	149	3852	0.095	ng	# 98
36) Indeno(1,2,3-cd)pyrene	25.849	276	8187	0.089	ng	99
37) Benzo(b)fluoranthene	22.885	252	9269	0.093	ng	# 89
38) Benzo(k)fluoranthene	22.929	252	9634	0.097	ng	# 88
39) Benzo(a)pyrene	23.467	252	6773	0.088	ng	# 85
40) Dibenzo(a,h)anthracene	25.873	278	6409	0.087	ng	# 90
41) Benzo(g,h,i)perylene	26.545	276	7524	0.092	ng	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032224\
Data File : BN030450.D
Acq On : 21 Mar 2024 17:14
Operator : MA/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Mar 21 23:54:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN032224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 21 23:54:14 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.731 min Scan# 64
 Delta R.T. -0.001 min
 Lab File: BN030450.D
 Acq: 21 Mar 2024 17:14




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 10799
 Ion Ratio Lower Upper
 152 100
 150 152.0 122.2 183.4
 115 60.7 48.2 72.4



#2
 1,4-Dioxane
 Concen: 0.128 ng
 RT: 3.261 min Scan# 25
 Delta R.T. 0.007 min
 Lab File: BN030450.D
 Acq: 21 Mar 2024 17:14

Tgt Ion: 88 Resp: 1783
 Ion Ratio Lower Upper
 88 100
 43 22.7 21.8 32.6
 58 52.2 53.9 80.9#









#3
n-Nitrosodimethylamine
Concen: 0.095 ng
RT: 3.564 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 42 Resp: 1283
Ion Ratio Lower Upper
42 100
74 146.0 112.6 169.0
44 0.0 9.9 14.9#



#4
2-Fluorophenol
Concen: 0.098 ng
RT: 5.312 min Scan# 309
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion: 112 Resp: 2617
Ion Ratio Lower Upper
112 100
64 56.2 44.2 66.2
63 28.2 23.0 34.4









#5
Phenol-d6
Concen: 0.098 ng
RT: 6.886 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 99 Resp: 3417
Ion Ratio Lower Upper
99 100
42 18.2 13.9 20.9
71 31.8 25.4 38.0



#6
bis(2-Chloroethyl)ether
Concen: 0.099 ng
RT: 7.161 min Scan# 565
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion: 93 Resp: 3305
Ion Ratio Lower Upper
93 100
63 72.8 58.6 87.8
95 33.4 26.3 39.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.509 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	136	Resp:	29381
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	9.5	7.4	11.0
68	8.1	6.5	9.7



#8
Nitrobenzene-d5
Concen: 0.096 ng
RT: 8.875 min Scan# 777
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:	82	Resp:	2182
Ion	Ratio	Lower	Upper
82	100		
128	41.3	31.8	47.8
54	52.1	38.8	58.2









#9
Naphthalene
Concen: 0.098 ng
RT: 10.562 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	128	Resp:	8332
Ion	Ratio	Lower	Upper
128	100		
129	12.1	9.1	13.7
127	14.4	10.4	15.6



#10
Hexachlorobutadiene
Concen: 0.099 ng
RT: 10.850 min Scan# 962
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:	225	Resp:	1568
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.9	51.8	77.8









#11
2-Methylnaphthalene-d10
Concen: 0.096 ng
RT: 12.106 min Scan# 1181
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 4085
Ion Ratio Lower Upper
152 100
151 21.1 17.0 25.6



#12
2-Methylnaphthalene
Concen: 0.096 ng
RT: 12.181 min Scan# 1201
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:142 Resp: 5190
Ion Ratio Lower Upper
142 100
141 88.4 70.5 105.7
115 39.8 29.2 43.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.361 min Scan# 1529
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:164 Resp: 14980
Ion Ratio Lower Upper
164 100
162 107.3 85.4 128.2
160 54.7 43.4 65.2



#14
2,4,6-Tribromophenol
Concen: 0.093 ng
RT: 15.853 min Scan# 1664
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:330 Resp: 508
Ion Ratio Lower Upper
330 100
332 93.1 77.4 116.0
141 46.5 36.2 54.4









#15
2-Fluorobiphenyl
Concen: 0.102 ng
RT: 12.993 min Scan# 1401
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:172 Resp: 6588
Ion Ratio Lower Upper
172 100
171 34.7 27.4 41.2
170 23.3 18.1 27.1



#16
Acenaphthylene
Concen: 0.096 ng
RT: 14.083 min Scan# 1503
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:152 Resp: 7384
Ion Ratio Lower Upper
152 100
151 19.5 15.5 23.3
153 13.0 10.3 15.5









#17
Acenaphthene
Concen: 0.095 ng
RT: 14.425 min Scan# 1535
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:154 Resp: 4881
Ion Ratio Lower Upper
154 100
153 114.1 89.3 133.9
152 60.7 47.4 71.2



#18
Fluorene
Concen: 0.095 ng
RT: 15.420 min Scan# 1628
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:166 Resp: 6519
Ion Ratio Lower Upper
166 100
165 96.8 77.8 116.6
167 13.7 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.106 min Scan# 1765
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:188 Resp: 34700
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 16.7 13.6 20.4



#21
4-Bromophenyl-phenylether
Concen: 0.098 ng
RT: 16.312 min Scan# 1701
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:248 Resp: 2193
Ion Ratio Lower Upper
248 100
250 96.0 77.0 115.6
141 64.8 49.2 73.8









#22
Hexachlorobenzene
Concen: 0.098 ng
RT: 16.411 min Scan# 1709
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:284 Resp: 2667
Ion Ratio Lower Upper
284 100
142 39.4 31.8 47.8
249 31.3 25.2 37.8



#23
Atrazine
Concen: 0.090 ng
RT: 16.585 min Scan# 1723
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:200 Resp: 1458
Ion Ratio Lower Upper
200 100
173 30.8 20.6 31.0
215 47.7 37.1 55.7









#25
Phenanthrene
Concen: 0.097 ng
RT: 17.156 min Scan# 1769
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:178 Resp: 10753
Ion Ratio Lower Upper
178 100
176 19.4 15.4 23.2
179 15.5 12.3 18.5



#26
Anthracene
Concen: 0.091 ng
RT: 17.243 min Scan# 1776
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:178 Resp: 8923
Ion Ratio Lower Upper
178 100
176 18.7 15.0 22.4
179 15.7 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.090 ng
RT: 19.145 min Scan# 2015
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:212 Resp: 7797
Ion Ratio Lower Upper
212 100
106 18.9 14.7 22.1
104 10.4 8.3 12.5



#28
Fluoranthene
Concen: 0.095 ng
RT: 19.178 min Scan# 2022
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:202 Resp: 11870
Ion Ratio Lower Upper
202 100
101 12.8 10.7 16.1
203 17.0 13.6 20.4









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.304 min Scan# 2449
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:240 Resp: 23694
Ion Ratio Lower Upper
240 100
120 19.4 15.9 23.9
236 29.5 23.8 35.6



#30
Pyrene
Concen: 0.102 ng
RT: 19.540 min Scan# 2100
Delta R.T. -0.000 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:202 Resp: 12075
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.8 14.1 21.1









#31
 Terphenyl-d14
 Concen: 0.106 ng
 RT: 19.759 min Scan# 2146
 Delta R.T. 0.004 min
 Lab File: BN030450.D
 Acq: 21 Mar 2024 17:14




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244 Resp: 5636
 Ion Ratio Lower Upper
 244 100
 212 10.5 7.8 11.6
 122 20.5 17.1 25.7



#32
 Benzo(a)anthracene
 Concen: 0.100 ng
 RT: 21.295 min Scan# 2448
 Delta R.T. -0.000 min
 Lab File: BN030450.D
 Acq: 21 Mar 2024 17:14

Tgt Ion:228 Resp: 9570
 Ion Ratio Lower Upper
 228 100
 226 27.8 21.5 32.3
 229 21.0 15.8 23.6





#33

Chrysene

Concen: 0.103 ng

RT: 21.340 min Scan# 2453

Delta R.T. -0.000 min

Lab File: BN030450.D

Acq: 21 Mar 2024 17:14

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:228 Resp: 10495

Ion Ratio Lower Upper

228 100

226 32.1 24.6 37.0

229 20.3 15.2 22.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.095 ng

RT: 21.250 min Scan# 2443

Delta R.T. -0.000 min

Lab File: BN030450.D

Acq: 21 Mar 2024 17:14




Tgt Ion:149 Resp: 3852


Ion Ratio Lower Upper

149 100

167 25.3 21.2 31.8

279 2.8 1.8 2.8#









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.563 min Scan# 2967
 Delta R.T. 0.003 min
 Lab File: BN030450.D
 Acq: 21 Mar 2024 17:14




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:264 Resp: 22770
 Ion Ratio Lower Upper
 264 100
 260 27.1 21.9 32.9
 265 39.6 32.6 48.8



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.089 ng
 RT: 25.849 min Scan# 3750
 Delta R.T. -0.003 min
 Lab File: BN030450.D
 Acq: 21 Mar 2024 17:14

Tgt Ion:276 Resp: 8187
 Ion Ratio Lower Upper
 276 100
 138 31.7 24.8 37.2
 277 23.2 19.1 28.7





#37

Benzo(b)fluoranthene

Concen: 0.093 ng

RT: 22.885 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN030450.D

Acq: 21 Mar 2024 17:14

Instrument :

BN030450.D

ClientSampleId :

SSTDICC0.1





Tgt Ion:252 Resp: 9269

Ion Ratio Lower Upper

252 100

253 26.4 18.2 27.2

125 21.8 12.0 18.0#



#38

Benzo(k)fluoranthene

Concen: 0.097 ng

RT: 22.929 min Scan# 2751

Delta R.T. -0.003 min

Lab File: BN030450.D

Acq: 21 Mar 2024 17:14




Tgt Ion:252 Resp: 9634


Ion Ratio Lower Upper

252 100

253 26.2 18.7 28.1

125 24.2 12.6 19.0#









#39
Benzo(a)pyrene
Concen: 0.088 ng
RT: 23.467 min Scan# 2934
Delta R.T. 0.003 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 6773
Ion Ratio Lower Upper
252 100
253 29.0 19.4 29.0
125 27.6 14.6 22.0#



#40
Dibenzo(a,h)anthracene
Concen: 0.087 ng
RT: 25.873 min Scan# 3758
Delta R.T. -0.006 min
Lab File: BN030450.D
Acq: 21 Mar 2024 17:14

Tgt Ion:278 Resp: 6409
Ion Ratio Lower Upper
278 100
139 27.7 18.2 27.2#
279 30.6 20.2 30.2#





#41
 Benzo(g,h,i)perylene
 Concen: 0.092 ng
 RT: 26.545 min Scan# 3988
 Delta R.T. -0.000 min
 Lab File: BN030450.D
 Acq: 21 Mar 2024 17:14

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	276	Resp:	7524
Ion	Ratio	Lower	Upper
276	100		
277	27.1	20.0	30.0
138	32.1	22.8	34.2

