


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_N\Data\BN032318\
 Data File : BN000567.D
 Acq On : 23 Mar 2018 08:48
 Operator : JU/SJ
 Sample : DFTPP85
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP85

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031918.M
 Title : SVOA CALIBRATION
 Last Update : Tue Mar 27 03:53:32 2018



AutoFind: Scans 2500, 2501, 2502; Background Corrected with Scan 2494

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.8	86704	PASS
68	69	0.00	2	1.9	1991	PASS
69	198	0.00	100	39.0	106597	PASS
70	69	0.00	2	0.7	701	PASS
127	198	10	80	49.2	134344	PASS
197	198	0.00	2	0.7	1971	PASS
198	198	100	100	100.0	272981	PASS
199	198	5	9	6.8	18429	PASS
275	198	10	60	24.7	67296	PASS
365	198	1	100	3.1	8451	PASS
441	443	0.01	100	82.8	33922	PASS
442	198	50	100	79.3	216405	PASS
443	442	15	24	18.9	40952	PASS

m/z	Abundance
37.00	390.0
38.00	1143.0
39.00	6744.0
40.00	581.0
41.00	453.0
44.00	375.0
49.10	901.0
50.10	26232.0
51.10	101664.0
52.10	5560.0
55.00	846.0
56.00	3176.0
57.00	7167.0
58.00	328.0
61.00	1447.0
62.00	1783.0
63.00	4406.0
64.00	717.0
65.00	2356.0
68.00	2372.0
69.00	122768.0
70.00	868.0
73.00	1239.0
74.00	13504.0
75.00	20208.0
76.10	7238.0
77.10	148992.0
78.10	10463.0
79.00	9506.0
80.00	7535.0
81.00	10307.0
82.00	2548.0
83.00	2478.0
84.00	428.0
85.00	2653.0
86.00	3010.0
87.00	1461.0
88.00	647.0
91.00	2365.0
92.00	2421.0
93.00	16392.0
94.00	1257.0
95.00	408.0
96.00	799.0
97.10	451.0
98.00	13136.0
99.00	10521.0
100.00	997.0
101.00	5557.0
102.00	326.0
103.00	2099.0
104.00	3665.0
105.00	3587.0
106.10	929.0
107.00	44304.0
108.00	7284.0
109.00	1423.0
110.00	82432.0
111.00	12101.0
112.00	1534.0
113.00	479.0
115.00	310.0
116.00	2402.0
117.00	36944.0
118.00	2925.0
119.00	389.0
120.00	459.0
122.00	2729.0
123.00	4789.0
124.00	2186.0
125.00	2037.0
126.10	822.0
127.00	147840.0
128.00	11354.0
129.00	56352.0
130.00	4583.0
131.00	1044.0
132.00	435.0
134.00	1546.0
135.00	3980.0
136.00	2015.0
137.00	2862.0
138.00	498.0
139.00	330.0
140.00	708.0
141.00	7372.0
142.00	2347.0
143.00	1552.0
144.00	540.0
145.00	384.0
146.00	1552.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

147.00	3621.0
148.00	9150.0
149.00	2208.0
150.10	558.0
151.00	1038.0
151.90	638.0
153.00	2513.0
154.00	1728.0
155.00	4022.0
156.00	5804.0
157.10	1160.0
157.90	1414.0
159.00	1102.0
160.00	2331.0
161.00	3205.0
162.00	993.0
163.90	315.0
165.00	2923.0
166.10	2055.0
167.00	14966.0
168.00	9459.0
169.10	1318.0
170.00	527.0
170.90	658.0
172.00	1317.0
173.00	1709.0
174.00	2915.0
175.10	5581.0
176.10	1675.0
177.00	2701.0
178.00	935.0
179.00	10168.0
180.00	7396.0
181.00	3135.0
182.10	605.0
184.00	930.0
185.00	5098.0
186.10	39024.0
187.10	11522.0
188.00	1407.0
189.00	2491.0
190.00	391.0
191.00	1245.0
192.00	3190.0
193.10	4090.0
194.00	855.0
195.00	495.0
196.00	8365.0
197.10	2184.0
198.00	284160.0
199.00	19456.0
200.00	1687.0
201.50	1222.0
203.00	1944.0
204.00	9746.0
205.10	16752.0
206.10	71936.0
207.10	10556.0
208.00	2639.0
209.00	949.0
210.00	1060.0
211.00	2692.0
213.00	302.0
215.00	846.0
216.00	1504.0
217.00	17848.0
218.00	2463.0
221.10	10941.0
221.80	2724.0
223.00	3930.0
224.10	38168.0
225.10	9571.0
226.10	1083.0
227.00	15614.0
228.00	2461.0
229.00	3274.0
230.00	547.0
231.10	1158.0
234.00	1101.0
235.00	1045.0
236.00	722.0
237.00	1446.0
239.00	711.0
240.00	513.0
241.00	814.0
242.10	1940.0
243.10	2230.0
244.10	29456.0
245.10	4038.0
246.00	5982.0
247.00	1280.0
249.00	1042.0
252.10	313.0
253.00	657.0
255.00	149248.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

256.10	22200.0
257.10	1777.0
258.00	8246.0
259.00	1396.0
260.00	350.0
265.00	3468.0
266.00	489.0
271.00	410.0
272.00	449.0
273.10	4402.0
274.00	11407.0
275.10	64328.0
276.10	8501.0
277.10	5500.0
278.00	1152.0
283.00	653.0
284.10	449.0
285.00	896.0
293.10	1312.0
294.10	326.0
296.00	18104.0
297.10	2525.0
302.00	473.0
303.10	2051.0
304.10	477.0
314.10	905.0
315.00	1941.0
316.10	1098.0
321.00	597.0
322.00	336.0
323.10	5501.0
324.00	871.0
327.00	1008.0
328.00	529.0
332.10	369.0
333.00	542.0
334.10	3578.0
335.10	1032.0
341.10	690.0
346.00	1403.0
352.10	1420.0
353.00	1202.0
354.10	1841.0
355.10	396.0
365.00	7554.0
366.00	1247.0
371.10	377.0
372.10	2851.0
373.00	643.0
383.00	692.0
390.10	335.0
391.00	308.0
402.10	1102.0
403.00	1783.0
404.00	649.0
421.00	1515.0
422.10	1379.0
423.10	10215.0
424.10	2644.0
425.00	322.0
441.10	27600.0
442.10	172480.0
443.10	32720.0
444.10	3026.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

m/z	Abundance
37.00	539.0
38.00	1245.0
39.10	6990.0
40.00	501.0
41.00	528.0
44.00	302.0
49.10	985.0
50.10	25640.0
51.10	98424.0
52.00	5182.0
55.10	676.0
56.00	3118.0
57.00	6812.0
58.00	418.0
61.00	1410.0
62.00	1809.0
63.00	4165.0
64.00	653.0
65.00	2307.0
67.00	316.0
68.00	2298.0
69.00	122504.0
70.00	753.0
73.00	1188.0
74.00	13008.0
75.00	20672.0
76.10	7121.0
77.10	148288.0
78.00	10444.0
79.00	9835.0
80.00	7295.0
81.00	10524.0
82.00	2482.0
83.00	2637.0
84.00	420.0
85.00	2897.0
86.00	3193.0
87.00	1376.0
88.00	659.0
89.10	317.0
91.00	2456.0
92.00	2562.0
93.00	16680.0
94.00	1024.0
95.00	406.0
96.10	1082.0
97.10	440.0
98.00	13547.0
99.00	10715.0
100.00	973.0
101.00	6214.0
102.10	345.0
103.00	2047.0
104.00	3775.0
105.00	3721.0
106.10	1230.0
107.00	46016.0
108.00	6947.0
109.00	1582.0
110.00	84624.0
111.00	12718.0
112.00	1550.0
113.00	530.0
116.00	2498.0
117.00	38640.0
118.00	3133.0
119.00	495.0
120.00	590.0
122.00	2812.0
123.00	4356.0
124.00	2098.0
125.00	2165.0
127.00	157056.0
128.00	11954.0
129.00	59360.0
130.00	4831.0
131.00	1016.0
132.00	484.0
133.00	398.0
134.00	1561.0
135.00	5030.0
136.00	2073.0
137.00	2689.0
138.10	538.0
139.00	338.0
140.00	864.0
141.00	7874.0
142.00	2530.0
143.00	1856.0
144.00	453.0
145.00	472.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

146.00	1398.0
147.00	4004.0
148.00	9629.0
149.00	1895.0
150.00	679.0
151.00	1032.0
151.80	556.0
153.00	2621.0
154.00	1764.0
155.00	4691.0
156.00	6715.0
157.10	1323.0
158.00	1513.0
159.00	1141.0
160.00	2389.0
161.10	3488.0
162.00	1214.0
163.00	371.0
163.90	359.0
165.00	2905.0
166.00	2426.0
167.00	16816.0
168.00	9093.0
169.00	1587.0
170.00	503.0
170.90	748.0
172.00	1408.0
173.00	1649.0
174.00	3478.0
175.00	6526.0
176.00	1853.0
177.10	3037.0
178.00	1063.0
179.00	12052.0
180.00	8384.0
181.00	3948.0
182.00	701.0
182.90	335.0
184.10	925.0
185.00	5709.0
186.00	43712.0
187.10	13058.0
188.10	1265.0
189.00	2756.0
190.00	538.0
191.00	1386.0
192.00	3834.0
193.10	4354.0
194.00	1036.0
195.00	404.0
196.10	9063.0
197.10	2218.0
198.00	319936.0
199.00	21808.0
200.00	1658.0
201.60	1464.0
203.00	2155.0
204.10	11074.0
205.10	19352.0
206.10	82048.0
207.10	11613.0
208.00	2775.0
209.00	1041.0
210.20	1320.0
211.00	3013.0
212.00	571.0
214.90	899.0
216.10	1798.0
217.00	21464.0
218.00	3136.0
219.10	310.0
221.10	12516.0
221.80	2837.0
223.00	4763.0
224.10	43672.0
225.10	11273.0
226.10	1435.0
227.00	18616.0
228.00	2824.0
229.00	3805.0
230.10	656.0
231.10	1645.0
232.00	341.0
233.10	367.0
234.00	1190.0
235.00	1361.0
236.10	999.0
237.00	1608.0
239.00	718.0
240.00	507.0
241.00	1017.0
242.00	2344.0
243.10	2670.0
244.10	35552.0
245.10	4397.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

246.00	6899.0
247.00	1352.0
248.00	349.0
249.00	1350.0
251.00	355.0
251.90	355.0
253.10	737.0
254.20	1721.0
255.10	181632.0
256.10	25872.0
257.00	2403.0
258.00	10029.0
259.10	1660.0
260.00	355.0
265.00	4073.0
266.00	522.0
271.00	337.0
272.00	504.0
273.00	5303.0
274.10	14233.0
275.10	80088.0
276.10	10977.0
277.00	7004.0
278.10	1127.0
283.00	680.0
284.00	672.0
285.10	1060.0
293.00	1543.0
294.00	408.0
295.20	618.0
296.10	22360.0
297.10	3345.0
301.10	315.0
302.00	367.0
303.10	2341.0
304.10	700.0
308.90	303.0
309.90	338.0
314.00	1039.0
315.00	2839.0
316.00	1498.0
317.00	307.0
321.00	815.0
322.10	437.0
323.10	7271.0
324.00	1038.0
327.00	1324.0
328.00	596.0
332.00	633.0
333.00	766.0
334.10	4850.0
335.10	1291.0
341.10	852.0
346.00	1560.0
347.00	335.0
352.10	2033.0
353.00	1742.0
354.10	2525.0
355.10	436.0
365.00	10475.0
366.00	1646.0
371.10	514.0
372.10	4135.0
373.10	979.0
383.00	1137.0
384.00	365.0
390.10	515.0
391.00	337.0
392.10	302.0
402.00	1661.0
403.00	2167.0
404.10	783.0
421.00	2411.0
422.00	2196.0
423.10	14751.0
424.10	3601.0
425.00	435.0
441.10	40664.0
442.10	262656.0
443.10	50808.0
444.10	4965.0
445.00	410.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

m/z	Abundance
37.00	383.0
38.00	748.0
39.10	4263.0
40.00	548.0
41.00	415.0
44.00	329.0
49.00	705.0
50.00	15845.0
51.00	60024.0
52.00	3114.0
55.00	788.0
56.00	1831.0
57.00	4393.0
61.00	868.0
62.00	977.0
63.00	2851.0
64.00	338.0
65.00	1594.0
67.00	304.0
68.00	1303.0
69.00	74520.0
70.00	483.0
73.00	898.0
74.00	8025.0
75.00	12072.0
76.10	4271.0
77.10	90968.0
78.00	6373.0
79.00	5983.0
80.00	4109.0
81.00	6329.0
82.00	1577.0
83.00	1606.0
84.00	317.0
85.00	1872.0
86.00	1793.0
87.00	776.0
88.00	441.0
91.00	1483.0
92.00	1699.0
93.00	10634.0
94.00	754.0
94.90	322.0
96.10	605.0
97.10	381.0
98.00	7881.0
99.00	6456.0
100.00	494.0
101.00	3511.0
102.90	1172.0
104.00	2201.0
105.00	2185.0
106.00	762.0
107.00	28016.0
108.00	4445.0
109.10	866.0
110.00	53056.0
111.00	7747.0
112.00	1075.0
113.00	303.0
116.00	1651.0
117.00	23912.0
118.00	1588.0
120.00	396.0
122.00	1836.0
123.00	3014.0
124.00	1345.0
125.00	1123.0
126.20	672.0
127.00	98136.0
128.00	7665.0
129.00	37104.0
130.00	3330.0
131.00	757.0
132.00	340.0
132.90	319.0
134.00	1193.0
135.00	3187.0
136.00	1207.0
137.00	1850.0
138.00	408.0
139.90	369.0
141.00	4947.0
142.00	1842.0
143.00	1052.0
143.90	381.0
145.10	348.0
146.00	991.0
147.00	2661.0
148.00	6361.0
149.00	1271.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

150.00	422.0
151.10	589.0
151.80	521.0
153.00	1639.0
154.00	1224.0
155.00	2793.0
156.00	4029.0
157.10	960.0
157.90	992.0
159.00	686.0
160.00	1562.0
161.10	2466.0
161.90	672.0
165.00	2082.0
166.00	1415.0
167.00	10859.0
168.00	6130.0
169.00	1021.0
170.00	386.0
171.00	472.0
172.00	1022.0
173.00	1299.0
174.10	2155.0
175.00	4143.0
176.00	1106.0
177.00	1911.0
177.90	616.0
179.00	7865.0
180.00	5588.0
181.00	2553.0
182.00	336.0
183.90	676.0
185.10	3630.0
186.00	29408.0
187.00	8405.0
188.00	863.0
189.00	1939.0
191.00	1088.0
192.00	2457.0
193.00	2710.0
194.10	611.0
195.00	319.0
196.00	5878.0
197.10	1511.0
198.00	214848.0
199.00	14025.0
200.00	994.0
201.50	836.0
203.00	1436.0
204.00	7583.0
205.00	12702.0
206.10	55376.0
207.10	7937.0
208.00	2076.0
209.00	716.0
210.10	885.0
211.00	2078.0
215.00	739.0
216.00	1262.0
217.00	14331.0
218.00	1989.0
221.00	8487.0
221.90	1951.0
223.00	3307.0
224.10	29560.0
225.10	7799.0
226.00	873.0
227.00	12927.0
228.00	1685.0
229.00	2709.0
230.00	425.0
231.10	1160.0
234.00	797.0
235.00	839.0
236.00	763.0
237.00	1024.0
239.00	561.0
240.10	423.0
241.00	621.0
242.00	1360.0
243.10	1664.0
244.10	24120.0
245.10	3058.0
246.00	4259.0
247.00	797.0
249.00	827.0
253.10	557.0
255.00	121288.0
256.10	16880.0
257.00	1329.0
258.00	7023.0
259.00	1052.0
265.00	2856.0
266.00	454.0
272.00	373.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

273.00	3559.0
274.10	10189.0
275.10	57472.0
276.00	7360.0
277.00	4940.0
278.00	941.0
283.10	465.0
285.10	985.0
293.10	1047.0
295.10	434.0
296.00	16608.0
297.10	2291.0
303.10	1795.0
304.10	537.0
308.00	311.0
314.10	740.0
315.00	1746.0
316.00	1034.0
321.00	591.0
323.10	5061.0
324.10	954.0
327.00	1113.0
328.00	536.0
332.00	387.0
333.10	510.0
334.10	3483.0
335.10	949.0
341.10	591.0
346.00	1353.0
352.00	1827.0
353.10	1017.0
354.10	1629.0
354.90	352.0
365.00	7324.0
366.00	1116.0
371.00	489.0
372.10	3277.0
373.10	707.0
383.00	848.0
390.00	612.0
391.10	351.0
402.00	1341.0
403.10	1884.0
404.00	605.0
421.00	1614.0
422.10	1507.0
423.10	12566.0
424.10	2593.0
425.00	398.0
441.10	33504.0
442.10	214080.0
443.10	39328.0
444.00	3823.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85

Scan 2494 (18.107 min): BN000567.D
DFTPP85

m/z	Abundance
40.00	462.0
55.00	343.0
207.10	928.0

Instrument :
BNA_N
ClientSampleId :
DFTPP85