

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032319\
 Data File : BN004919.D
 Acq On : 23 Mar 2019 13:08
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 24 01:30:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	66	0.00
2	1,4-Dioxane	0.400	0.562	-40.5#	92	-0.02
3	n-Nitrosodimethylamine	0.400	0.582	-45.5#	104	-0.02
4 S	2-Fluorophenol	0.400	0.512	-28.0#	83	0.00
5 S	Phenol-d6	0.400	0.514	-28.5#	85	0.00
6	bis(2-Chloroethyl)ether	0.400	0.457	-14.2	77	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	62	-0.01
8 S	Nitrobenzene-d5	0.400	0.518	-29.5#	81	-0.03
9	Naphthalene	0.400	0.435	-8.7	67	-0.03
10	Hexachlorobutadiene	0.400	0.410	-2.5	63	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.429	-7.2	67	-0.01
12	2-Methylnaphthalene	0.400	0.435	-8.7	67	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	68	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.478	-19.5	86	-0.01
15 S	2-Fluorobiphenyl	0.400	0.415	-3.7	69	-0.01
16	Acenaphthylene	0.400	0.434	-8.5	77	-0.01
17	Acenaphthene	0.400	0.411	-2.7	69	-0.01
18	Fluorene	0.400	0.393	1.8	65	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	60	-0.01
20	4-Bromophenyl-phenylether	0.400	0.422	-5.5	64	-0.01
21	Hexachlorobenzene	0.400	0.399	0.3	59	0.00
22	Pentachlorophenol	0.400	1.050	-162.5#	182	-0.01
23	Phenanthrene	0.400	0.413	-3.2	61	-0.01
24	Anthracene	0.400	0.430	-7.5	66	-0.01
25 SURR	Fluoranthene-d10	0.400	0.416	-4.0	63	-0.01
26	Fluoranthene	0.400	0.409	-2.2	61	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	60	-0.01
28	Pyrene	0.400	0.425	-6.2	62	-0.01
29 S	Terphenyl-d14	0.400	0.435	-8.7	64	-0.01
30	Benzo(a)anthracene	0.400	0.435	-8.7	66	-0.01
31	Chrysene	0.400	0.412	-3.0	59	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.641	-60.2#	104	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.494	-23.5	71	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	67	-0.01
35	Benzo(b)fluoranthene	0.400	0.382	4.5	63	-0.01
36	Benzo(k)fluoranthene	0.400	0.365	8.8	62	-0.01
37 C	Benzo(a)pyrene	0.400	0.400	0.0	68	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.427	-6.7	71	-0.02
39	Benzo(a,h,i)perylene	0.400	0.430	-7.5	71	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032319\
Data File : BN004919.D
Acq On : 23 Mar 2019 13:08
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 24 01:30:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				