

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032319\
 Data File : BN004919.D
 Acq On : 23 Mar 2019 13:08
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 24 01:30:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	66	0.00
2	1,4-Dioxane	0.455	0.638	-40.2#	92	-0.02
3	n-Nitrosodimethylamine	0.269	0.391	-45.4#	104	-0.02
4 S	2-Fluorophenol	1.087	1.391	-28.0#	83	0.00
5 S	Phenol-d6	1.354	1.741	-28.6#	85	0.00
6	bis(2-Chloroethyl)ether	1.221	1.396	-14.3	77	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	62	-0.01
8 S	Nitrobenzene-d5	0.265	0.343	-29.4#	81	-0.03
9	Naphthalene	1.122	1.221	-8.8	67	-0.03
10	Hexachlorobutadiene	0.214	0.220	-2.8	63	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.749	-7.3	67	-0.01
12	2-Methylnaphthalene	0.833	0.906	-8.8	67	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	68	-0.01
14 S	2,4,6-Tribromophenol	0.276	0.330	-19.6	86	-0.01
15 S	2-Fluorobiphenyl	1.653	1.715	-3.8	69	-0.01
16	Acenaphthylene	2.060	2.235	-8.5	77	-0.01
17	Acenaphthene	1.349	1.387	-2.8	69	-0.01
18	Fluorene	1.883	1.852	1.6	65	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	60	-0.01
20	4-Bromophenyl-phenylether	0.261	0.276	-5.7	64	-0.01
21	Hexachlorobenzene	0.299	0.299	0.0	59	0.00
22	Pentachlorophenol	0.058	0.153	-163.8#	182#	-0.01
23	Phenanthrene	1.292	1.333	-3.2	61	-0.01
24	Anthracene	1.154	1.240	-7.5	66	-0.01
25 SURR	Fluoranthene-d10	1.339	1.393	-4.0	63	-0.01
26	Fluoranthene	1.660	1.697	-2.2	61	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	60	-0.01
28	Pvrene	1.220	1.295	-6.1	62	-0.01
29 S	Terphenyl-d14	0.819	0.889	-8.5	64	-0.01
30	Benzo(a)anthracene	1.312	1.427	-8.8	66	-0.01
31	Chrysene	1.382	1.422	-2.9	59	-0.01
32	Bis(2-ethylhexyl)phthalate	0.524	0.840	-60.3#	104	-0.01
33	Indeno(1,2,3-cd)pyrene	1.999	2.467	-23.4	71	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	67	-0.01
35	Benzo(b)fluoranthene	1.447	1.381	4.6	63	-0.01
36	Benzo(k)fluoranthene	1.424	1.299	8.8	62	-0.01
37 C	Benzo(a)pyrene	1.327	1.327	0.0	68	-0.01
38	Dibenzo(a,h)anthracene	1.537	1.641	-6.8	71	-0.02
39	Benzo(a,h,i)perylene	1.566	1.685	-7.6	71	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032319\
Data File : BN004919.D
Acq On : 23 Mar 2019 13:08
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 24 01:30:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			