

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032321\
 Data File : BN014046.D
 Acq On : 23 Mar 2021 17:18
 Operator : CG/JU
 Sample : M1735-10
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D6-20210315

Quant Time: Mar 23 17:56:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 11:06:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	5374	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	20779	0.40	ng	#-0.01
13) Acenaphthene-d10	14.321	164	11980	0.40	ng	0.00
19) Phenanthrene-d10	17.067	188	26646	0.40	ng	# 0.00
29) Chrysene-d12	21.237	240	25108	0.40	ng	0.00
36) Perylene-d12	23.445	264	10135	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	1663	0.13	ng	0.00
5) Phenol-d6	6.863	99	1164	0.07	ng	0.00
8) Nitrobenzene-d5	8.845	82	4197	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.062	152	10653	0.33	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1198	0.31	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	14598	0.33	ng	0.00
27) Fluoranthene-d10	19.091	212	28930	0.39	ng	0.00
31) Terphenyl-d14	19.692	244	25320	0.46	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	29059	4.10	ng	# 88
6) bis(2-Chloroethyl)ether	7.128	93	597	0.04	ng	90
34) Bis(2-ethylhexyl)phtha...	21.162	149	2048	0.08	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032321\
Data File : BN014046.D
Acq On : 23 Mar 2021 17:18
Operator : CG/JU
Sample : M1735-10
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D6-20210315

Quant Time: Mar 23 17:56:45 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 11:06:27 2021
Response via : Initial Calibration

