

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032321\
 Data File : BN014048.D
 Acq On : 23 Mar 2021 18:30
 Operator : CG/JU
 Sample : M1735-08
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D5-20210315

Manual Integrations
 APPROVED

mohammad
 3/24/2021 1:55:35 PM

Quant Time: Mar 24 02:58:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 11:06:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	6187	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	24313	0.40	ng	#-0.01
13) Acenaphthene-d10	14.321	164	5215	0.40	ng	0.00
19) Phenanthrene-d10	17.067	188	23025	0.40	ng	# 0.00
29) Chrysene-d12	21.237	240	404	0.40	ng	# 0.00
36) Perylene-d12	23.449	264	4m	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	1666	0.12	ng	0.00
5) Phenol-d6	6.855	99	2843	0.16	ng	0.00
8) Nitrobenzene-d5	8.845	82	4452	0.28	ng	0.00
11) 2-Methylnaphthalene-d10	12.062	152	9373	0.25	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1544	0.91	ng	0.00
15) 2-Fluorobiphenyl	12.951	172	14328	0.74	ng	0.00
27) Fluoranthene-d10	19.091	212	8206	0.13	ng	0.00
31) Terphenyl-d14	19.692	244	19241	21.68	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	41202	5.05	ng	Qvalue 97
6) bis(2-Chloroethyl)ether	7.129	93	720	0.05	ng	# 77
25) Phenanthrene	17.104	178	1504	0.02	ng	# 84
32) Benzo(a)anthracene	21.258	228	89m	0.08	ng	
33) Chrysene	21.269	228	38m	0.03	ng	
34) Bis(2-ethylhexyl)phtha...	21.162	149	600	1.40	ng	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032321\
Data File : BN014048.D
Acq On : 23 Mar 2021 18:30
Operator : CG/JU
Sample : M1735-08
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D5-20210315

Manual Integrations
APPROVED

mohammad
3/24/2021 1:55:35 PM

Quant Time: Mar 24 02:58:17 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 11:06:27 2021
Response via : Initial Calibration

