

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032321\
 Data File : BN014050.D
 Acq On : 23 Mar 2021 19:42
 Operator : CG/JU
 Sample : M1735-18DL 2X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE131D1-20210317DL

Quant Time: Mar 24 02:58:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 11:06:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	6007	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	22696	0.40	ng	#-0.01
13) Acenaphthene-d10	14.321	164	13379	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	30963	0.40	ng	-0.01
29) Chrysene-d12	21.239	240	29388	0.40	ng	# 0.00
36) Perylene-d12	23.445	264	27511	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	815	0.06	ng	0.00
5) Phenol-d6	6.863	99	599	0.03	ng	0.00
8) Nitrobenzene-d5	8.845	82	1969	0.13	ng	0.00
11) 2-Methylnaphthalene-d10	12.062	152	4843	0.14	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	521	0.12	ng	0.00
15) 2-Fluorobiphenyl	12.951	172	6639	0.13	ng	0.00
27) Fluoranthene-d10	19.089	212	15502	0.18	ng	0.00
31) Terphenyl-d14	19.690	244	12956	0.20	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	23182	2.93	ng	96
34) Bis(2-ethylhexyl)phtha...	21.154	149	4465	0.14	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032321\
Data File : BN014050.D
Acq On : 23 Mar 2021 19:42
Operator : CG/JU
Sample : M1735-18DL 2X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE131D1-20210317DL

Quant Time: Mar 24 02:58:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 11:06:27 2021
Response via : Initial Calibration

