

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032321\
 Data File : BN014051.D
 Acq On : 23 Mar 2021 20:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 24 02:58:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 11:06:27 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.409	-2.2	96	0.00
3	n-Nitrosodimethylamine	0.400	0.397	0.8	107	0.00
4 S	2-Fluorophenol	0.400	0.495	-23.7	112	0.00
5 S	Phenol-d6	0.400	0.496	-24.0	113	0.00
6	bis(2-Chloroethyl)ether	0.400	0.467	-16.8	106	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	-0.01
8 S	Nitrobenzene-d5	0.400	0.425	-6.2	106	0.00
9	Naphthalene	0.400	0.453	-13.2	110	0.00
10	Hexachlorobutadiene	0.400	0.444	-11.0	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.442	-10.5	109	0.00
12	2-Methylnaphthalene	0.400	0.455	-13.7	112	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.400	0.416	-4.0	117	0.00
15 S	2-Fluorobiphenyl	0.400	0.437	-9.2	114	0.00
16	Acenaphthylene	0.400	0.408	-2.0	115	0.00
17	Acenaphthene	0.400	0.424	-6.0	113	0.00
18	Fluorene	0.400	0.433	-8.2	115	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	105	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.388	3.0	111	0.00
21	4-Bromophenyl-phenylether	0.400	0.443	-10.7	112	0.00
22	Hexachlorobenzene	0.400	0.438	-9.5	111	0.00
23	Atrazine	0.400	0.400	0.0	109	0.00
24	Pentachlorophenol	0.400	0.339	15.3	102	-0.01
25	Phenanthrene	0.400	0.446	-11.5	115	0.00
26	Anthracene	0.400	0.437	-9.2	117	0.00
27 SURR	Fluoranthene-d10	0.400	0.434	-8.5	113	0.00
28	Fluoranthene	0.400	0.458	-14.5	120	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	106	0.00
30	Pyrene	0.400	0.464	-16.0	118	0.00
31 S	Terphenyl-d14	0.400	0.489	-22.2	123	0.00
32	Benzo(a)anthracene	0.400	0.443	-10.7	114	0.00
33	Chrysene	0.400	0.471	-17.7	120	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.395	1.3	110	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.487	-21.7	125	0.00
36 I	Perylene-d12	0.400	0.400	0.0	114	0.00
37	Benzo(b)fluoranthene	0.400	0.456	-14.0	128	0.00
38	Benzo(k)fluoranthene	0.400	0.444	-11.0	122	0.00
39 C	Benzo(a)pyrene	0.400	0.463	-15.8	130	0.00
40	Dibenzo(a,h)anthracene	0.400	0.441	-10.2	124	0.00
41	Benzo(g,h,i)perylene	0.400	0.441	-10.2	123	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4