

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032321\
 Data File : BN014053.D
 Acq On : 23 Mar 2021 22:42
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 24 02:59:00 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 11:06:27 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	103	0.00
2	1,4-Dioxane	0.400	0.417	-4.2	105	0.00
3	n-Nitrosodimethylamine	0.400	0.354	11.5	102	-0.02
4 S	2-Fluorophenol	0.400	0.464	-16.0	112	0.00
5 S	Phenol-d6	0.400	0.466	-16.5	113	0.00
6	bis(2-Chloroethyl)ether	0.400	0.459	-14.8	112	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	-0.01
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	108	0.00
9	Naphthalene	0.400	0.445	-11.2	116	0.00
10	Hexachlorobutadiene	0.400	0.439	-9.7	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.442	-10.5	117	0.00
12	2-Methylnaphthalene	0.400	0.445	-11.2	118	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.400	0.373	6.8	110	0.00
15 S	2-Fluorobiphenyl	0.400	0.436	-9.0	119	0.00
16	Acenaphthylene	0.400	0.403	-0.8	118	0.00
17	Acenaphthene	0.400	0.418	-4.5	117	0.00
18	Fluorene	0.400	0.421	-5.2	116	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.370	7.5	110	0.00
21	4-Bromophenyl-phenylether	0.400	0.432	-8.0	114	0.00
22	Hexachlorobenzene	0.400	0.431	-7.7	114	0.00
23	Atrazine	0.400	0.372	7.0	105	0.00
24	Pentachlorophenol	0.400	0.377	5.8	118	-0.01
25	Phenanthrene	0.400	0.435	-8.7	116	0.00
26	Anthracene	0.400	0.422	-5.5	117	0.00
27 SURR	Fluoranthene-d10	0.400	0.432	-8.0	117	0.00
28	Fluoranthene	0.400	0.434	-8.5	118	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
30	Pyrene	0.400	0.459	-14.8	116	0.00
31 S	Terphenyl-d14	0.400	0.473	-18.2	119	0.00
32	Benzo(a)anthracene	0.400	0.429	-7.2	110	0.00
33	Chrysene	0.400	0.467	-16.8	118	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.345	13.8	96	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.483	-20.7	123	0.00
36 I	Perylene-d12	0.400	0.400	0.0	115	0.00
37	Benzo(b)fluoranthene	0.400	0.438	-9.5	124	0.00
38	Benzo(k)fluoranthene	0.400	0.435	-8.7	121	0.00
39 C	Benzo(a)pyrene	0.400	0.453	-13.2	129	0.00
40	Dibenzo(a,h)anthracene	0.400	0.433	-8.2	122	0.00
41	Benzo(g,h,i)perylene	0.400	0.433	-8.2	122	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4