

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032323\
 Data File : BN024509.D
 Acq On : 23 Mar 2023 12:07
 Operator : CG/JU
 Sample : 01903-03
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE103D2-20230310

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/24/2023
 Supervised By :Jagrut Upadhyay 03/24/2023

Quant Time: Mar 24 02:36:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 24 02:34:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.026	152	9894	0.400	ng	0.00
7) Naphthalene-d8	10.845	136	28101	0.400	ng	0.01
13) Acenaphthene-d10	14.654	164	14309	0.400	ng	0.00
19) Phenanthrene-d10	17.400	188	30144	0.400	ng	# 0.00
29) Chrysene-d12	21.592	240	19068	0.400	ng	# 0.00
35) Perylene-d12	24.097	264	16474	0.400	ng	# 0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.577	112	3681	0.168	ng	0.00
5) Phenol-d6	7.188	99	2512	0.089	ng	0.00
8) Nitrobenzene-d5	9.190	82	8635m	0.460	ng	0.00
11) 2-Methylnaphthalene-d10	12.422	152	14786	0.428	ng	0.00
14) 2,4,6-Tribromophenol	16.147	330	2819	0.385	ng	0.00
15) 2-Fluorobiphenyl	13.286	172	25600	0.474	ng	0.00
27) Fluoranthene-d10	19.433	212	30484	0.477	ng	0.00
31) Terphenyl-d14	20.025	244	19840	0.617	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.425	88	4911	0.456	ng	# 89
34) Bis(2-ethylhexyl)phtha...	21.493	149	3461	0.115	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032323\
Data File : BN024509.D
Acq On : 23 Mar 2023 12:07
Operator : CG/JU
Sample : 01903-03
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE103D2-20230310

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 03/24/2023
Supervised By : Jagrut Upadhyay 03/24/2023

Quant Time: Mar 24 02:36:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 24 02:34:14 2023
Response via : Initial Calibration

