

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032323\
 Data File : BN024512.D
 Acq On : 23 Mar 2023 14:34
 Operator : CG/JU
 Sample : 01941-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT180D1-20230313

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/24/2023
 Supervised By :Jagrut Upadhyay 03/24/2023

Quant Time: Mar 24 02:36:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 24 02:34:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.025	152	9275	0.400	ng	0.00
7) Naphthalene-d8	10.844	136	26426	0.400	ng	0.01
13) Acenaphthene-d10	14.654	164	13478	0.400	ng	0.00
19) Phenanthrene-d10	17.400	188	28590	0.400	ng	# 0.00
29) Chrysene-d12	21.600	240	19037	0.400	ng	0.00
35) Perylene-d12	24.085	264	16561	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.569	112	3080	0.150	ng	0.00
5) Phenol-d6	7.187	99	2015	0.076	ng	0.00
8) Nitrobenzene-d5	9.189	82	7539m	0.427	ng	0.00
11) 2-Methylnaphthalene-d10	12.421	152	13119	0.404	ng	0.00
14) 2,4,6-Tribromophenol	16.146	330	2502	0.363	ng	0.00
15) 2-Fluorobiphenyl	13.285	172	19993	0.393	ng	0.00
27) Fluoranthene-d10	19.433	212	29862	0.493	ng	0.00
31) Terphenyl-d14	20.024	244	18554	0.578	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.425	88	28549	2.829	ng	97
34) Bis(2-ethylhexyl)phtha...	21.493	149	4469	0.149	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032323\
Data File : BN024512.D
Acq On : 23 Mar 2023 14:34
Operator : CG/JU
Sample : 01941-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT180D1-20230313

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 03/24/2023
Supervised By : Jagrut Upadhyay 03/24/2023

Quant Time: Mar 24 02:36:51 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 24 02:34:14 2023
Response via : Initial Calibration

