

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032323\
 Data File : BN024531.D
 Acq On : 24 Mar 2023 05:18
 Operator : CG/JU
 Sample : 01941-03
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT180D-20230313

Quant Time: Mar 24 05:48:29 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 24 03:05:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.025	152	8300	0.400	ng	0.00
7) Naphthalene-d8	10.834	136	23812	0.400	ng	# 0.00
13) Acenaphthene-d10	14.654	164	11945	0.400	ng	0.00
19) Phenanthrene-d10	17.400	188	25158	0.400	ng	0.00
29) Chrysene-d12	21.592	240	16054	0.400	ng	0.00
35) Perylene-d12	24.079	264	13418	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.570	112	3030	0.165	ng	0.00
5) Phenol-d6	7.180	99	1977	0.084	ng	0.00
8) Nitrobenzene-d5	9.190	82	7785	0.489	ng	0.00
11) 2-Methylnaphthalene-d10	12.418	152	13371	0.457	ng	0.00
14) 2,4,6-Tribromophenol	16.147	330	2411	0.395	ng	0.00
15) 2-Fluorobiphenyl	13.285	172	21242	0.471	ng	0.00
27) Fluoranthene-d10	19.429	212	27481	0.515	ng	0.00
31) Terphenyl-d14	20.016	244	16282	0.602	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.425	88	1755	0.194	ng	# 66
34) Bis(2-ethylhexyl)phtha...	21.484	149	3721	0.147	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032323\
Data File : BN024531.D
Acq On : 24 Mar 2023 05:18
Operator : CG/JU
Sample : 01941-03
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT180D-20230313

Quant Time: Mar 24 05:48:29 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 24 03:05:57 2023
Response via : Initial Calibration

