

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032323\
 Data File : BN024533.D
 Acq On : 24 Mar 2023 06:30
 Operator : CG/JU
 Sample : 01941-16
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT162S1-20230314

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/24/2023
 Supervised By :Jagrut Upadhyay 03/24/2023

Quant Time: Mar 24 06:59:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 24 03:05:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.018	152	9851	0.400	ng	0.00
7) Naphthalene-d8	10.834	136	28228	0.400	ng	# 0.00
13) Acenaphthene-d10	14.654	164	14337	0.400	ng	0.00
19) Phenanthrene-d10	17.400	188	30426	0.400	ng	0.00
29) Chrysene-d12	21.592	240	19471	0.400	ng	0.00
35) Perylene-d12	24.097	264	15858	0.400	ng	# 0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.570	112	3234	0.148	ng	0.00
5) Phenol-d6	7.180	99	2313	0.083	ng	0.00
8) Nitrobenzene-d5	9.190	82	9082m	0.482	ng	0.00
11) 2-Methylnaphthalene-d10	12.418	152	15704	0.453	ng	0.00
14) 2,4,6-Tribromophenol	16.147	330	2210	0.302	ng	0.00
15) 2-Fluorobiphenyl	13.285	172	23629	0.436	ng	0.00
27) Fluoranthene-d10	19.429	212	32511	0.504	ng	0.00
31) Terphenyl-d14	20.016	244	18651	0.568	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.425	88	3265	0.305	ng	# 85
34) Bis(2-ethylhexyl)phtha...	21.484	149	4571	0.149	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032323\
Data File : BN024533.D
Acq On : 24 Mar 2023 06:30
Operator : CG/JU
Sample : 01941-16
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT162S1-20230314

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 03/24/2023
Supervised By : Jagrut Upadhyay 03/24/2023

Quant Time: Mar 24 06:59:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 24 03:05:57 2023
Response via : Initial Calibration

