

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN032323\  
 Data File : BN024537.D  
 Acq On : 24 Mar 2023 08:56  
 Operator : CG/JU  
 Sample : 01941-09  
 Misc :  
 ALS Vial : 32 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 TT101D2-20230314

Quant Time: Mar 24 09:25:59 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN031323.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Mar 24 03:05:57 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| -----                         |        |      |          |       |       |          |
| Internal Standards            |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 8.018  | 152  | 12557    | 0.400 | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.834 | 136  | 35330    | 0.400 | ng    | # 0.00   |
| 13) Acenaphthene-d10          | 14.654 | 164  | 17544    | 0.400 | ng    | 0.00     |
| 19) Phenanthrene-d10          | 17.400 | 188  | 37071    | 0.400 | ng    | 0.00     |
| 29) Chrysene-d12              | 21.592 | 240  | 24246    | 0.400 | ng    | 0.00     |
| 35) Perylene-d12              | 24.091 | 264  | 19210    | 0.400 | ng    | # 0.01   |
| System Monitoring Compounds   |        |      |          |       |       |          |
| 4) 2-Fluorophenol             | 5.570  | 112  | 4490     | 0.161 | ng    | 0.00     |
| 5) Phenol-d6                  | 7.180  | 99   | 3022     | 0.085 | ng    | 0.00     |
| 8) Nitrobenzene-d5            | 9.190  | 82   | 11987    | 0.508 | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10   | 12.418 | 152  | 20342    | 0.468 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 16.147 | 330  | 3491     | 0.389 | ng    | 0.00     |
| 15) 2-Fluorobiphenyl          | 13.285 | 172  | 30857    | 0.466 | ng    | 0.00     |
| 27) Fluoranthene-d10          | 19.429 | 212  | 39648    | 0.504 | ng    | 0.00     |
| 31) Terphenyl-d14             | 20.016 | 244  | 23199    | 0.567 | ng    | 0.00     |
| Target Compounds              |        |      |          |       |       |          |
| 2) 1,4-Dioxane                | 3.425  | 88   | 66228    | 4.847 | ng    | 96       |
| 6) bis(2-Chloroethyl)ether    | 7.440  | 93   | 1463     | 0.041 | ng    | 96       |
| 34) Bis(2-ethylhexyl)phtha... | 21.493 | 149  | 38938    | 1.022 | ng    | 100      |
| -----                         |        |      |          |       |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN032323\  
Data File : BN024537.D  
Acq On : 24 Mar 2023 08:56  
Operator : CG/JU  
Sample : 01941-09  
Misc :  
ALS Vial : 32 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
TT101D2-20230314

Quant Time: Mar 24 09:25:59 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN031323.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Mar 24 03:05:57 2023  
Response via : Initial Calibration

