

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032323\
 Data File : BN024522.D
 Acq On : 23 Mar 2023 23:12
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 24 03:07:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 24 03:05:57 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	141	0.00
2	1,4-Dioxane	0.400	0.386	3.5	130	0.00
3	n-Nitrosodimethylamine	0.400	0.396	1.0	137	0.00
4 S	2-Fluorophenol	0.400	0.422	-5.5	148	0.00
5 S	Phenol-d6	0.400	0.400	0.0	144	0.00
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	136	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	144	0.00
8 S	Nitrobenzene-d5	0.400	0.460	-15.0	175	0.00
9	Naphthalene	0.400	0.399	0.3	145	0.00
10	Hexachlorobutadiene	0.400	0.415	-3.7	148	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.399	0.3	147	0.00
12	2-Methylnaphthalene	0.400	0.390	2.5	144	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	140	0.00
14 S	2,4,6-Tribromophenol	0.400	0.313	21.8	120	0.00
15 S	2-Fluorobiphenyl	0.400	0.400	0.0	141	0.00
16	Acenaphthylene	0.400	0.361	9.8	132	0.00
17	Acenaphthene	0.400	0.377	5.8	134	0.00
18	Fluorene	0.400	0.378	5.5	136	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	135	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.301	24.8	73	0.00
21	4-Bromophenyl-phenylether	0.400	0.381	4.8	133	0.00
22	Hexachlorobenzene	0.400	0.381	4.8	132	0.00
23	Atrazine	0.400	0.365	8.8	139	0.00
24	Pentachlorophenol	0.400	0.341	14.8	109	0.00
25	Phenanthrene	0.400	0.378	5.5	131	0.00
26	Anthracene	0.400	0.353	11.8	127	0.00
27 SURR	Fluoranthene-d10	0.400	0.382	4.5	134	0.00
28	Fluoranthene	0.400	0.370	7.5	130	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	120	0.00
30	Pyrene	0.400	0.439	-9.7	131	0.00
31 S	Terphenyl-d14	0.400	0.414	-3.5	124	0.00
32	Benzo(a)anthracene	0.400	0.373	6.8	117	0.00
33	Chrysene	0.400	0.389	2.8	117	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.431	-7.7	141	0.00
35 I	Perylene-d12	0.400	0.400	0.0	114	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.378	5.5	113	0.00
37	Benzo(b)fluoranthene	0.400	0.396	1.0	120	0.00
38	Benzo(k)fluoranthene	0.400	0.371	7.3	112	0.00
39 C	Benzo(a)pyrene	0.400	0.385	3.8	117	0.00
40	Dibenzo(a,h)anthracene	0.400	0.373	6.8	113	0.00
41	Benzo(g,h,i)perylene	0.400	0.380	5.0	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4