

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN032419\
 Data File : BN004935.D
 Acq On : 24 Mar 2019 12:54
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 25 03:41:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	205#	-0.02
2	1,4-Dioxane	0.455	0.570	-25.3#	255#	-0.02
3	n-Nitrosodimethylamine	0.269	0.369	-37.2#	305#	-0.02
4 S	2-Fluorophenol	1.087	1.319	-21.3	245#	-0.02
5 S	Phenol-d6	1.354	1.538	-13.6	235#	-0.02
6	bis(2-Chloroethyl)ether	1.221	1.365	-11.8	236#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	183#	-0.03
8 S	Nitrobenzene-d5	0.265	0.335	-26.4#	234#	-0.03
9	Naphthalene	1.122	1.232	-9.8	199#	-0.03
10	Hexachlorobutadiene	0.214	0.226	-5.6	192#	-0.03
11 SURR	2-Methylnaphthalene-d10	0.698	0.687	1.6	181#	-0.02
12	2-Methylnaphthalene	0.833	0.834	-0.1	182#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	155#	-0.02
14 S	2,4,6-Tribromophenol	0.276	0.282	-2.2	168#	-0.02
15 S	2-Fluorobiphenyl	1.653	1.909	-15.5	176#	-0.02
16	Acenaphthylene	2.060	2.251	-9.3	178#	-0.02
17	Acenaphthene	1.349	1.433	-6.2	163#	-0.01
18	Fluorene	1.883	1.860	1.2	150#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	132	-0.01
20	4-Bromophenyl-phenylether	0.261	0.289	-10.7	147	-0.01
21	Hexachlorobenzene	0.299	0.304	-1.7	133	-0.01
22	Pentachlorophenol	0.058	0.084	-44.8#	220#	-0.02
23	Phenanthrene	1.292	1.352	-4.6	137	-0.01
24	Anthracene	1.154	1.222	-5.9	143	-0.02
25 SURR	Fluoranthene-d10	1.339	1.314	1.9	130	-0.02
26	Fluoranthene	1.660	1.643	1.0	130	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	121	-0.01
28	Pyrene	1.220	1.319	-8.1	129	-0.02
29 S	Terphenyl-d14	0.819	0.943	-15.1	138	-0.02
30	Benzo(a)anthracene	1.312	1.357	-3.4	127	-0.02
31	Chrysene	1.382	1.378	0.3	117	-0.02
32	Bis(2-ethylhexyl)phthalate	0.524	0.959	-83.0#	240#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.999	1.774	11.3	104	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	109	-0.02
35	Benzo(b)fluoranthene	1.447	1.554	-7.4	115	-0.02
36	Benzo(k)fluoranthene	1.424	1.449	-1.8	112	-0.02
37 C	Benzo(a)pyrene	1.327	1.338	-0.8	110	-0.02
38	Dibenzo(a,h)anthracene	1.537	1.483	3.5	104	-0.03
39	Benzo(g,h,i)perylene	1.566	1.477	5.7	101	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN032419\
Data File : BN004935.D
Acq On : 24 Mar 2019 12:54
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 25 03:41:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			