

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN032419\
 Data File : BN004952.D
 Acq On : 24 Mar 2019 22:53
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 25 04:13:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	261#	-0.02
2	1,4-Dioxane	0.455	0.484	-6.4	275#	-0.02
3	n-Nitrosodimethylamine	0.269	0.328	-21.9	344#	-0.02
4 S	2-Fluorophenol	1.087	1.316	-21.1	311#	0.00
5 S	Phenol-d6	1.354	1.657	-22.4	322#	0.00
6	bis(2-Chloroethyl)ether	1.221	1.375	-12.6	302#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	254#	-0.03
8 S	Nitrobenzene-d5	0.265	0.330	-24.5	320#	-0.03
9	Naphthalene	1.122	1.222	-8.9	273#	-0.03
10	Hexachlorobutadiene	0.214	0.220	-2.8	259#	-0.03
11 SURR	2-Methylnaphthalene-d10	0.698	0.694	0.6	253#	-0.02
12	2-Methylnaphthalene	0.833	0.840	-0.8	254#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	206#	-0.02
14 S	2,4,6-Tribromophenol	0.276	0.275	0.4	218#	-0.01
15 S	2-Fluorobiphenyl	1.653	1.954	-18.2	239#	-0.02
16	Acenaphthylene	2.060	2.515	-22.1	264#	-0.02
17	Acenaphthene	1.349	1.486	-10.2	225#	-0.01
18	Fluorene	1.883	1.905	-1.2	204#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	166#	-0.01
20	4-Bromophenyl-phenylether	0.261	0.299	-14.6	191#	-0.01
21	Hexachlorobenzene	0.299	0.309	-3.3	169#	-0.01
22	Pentachlorophenol	0.058	0.078	-34.5#	256#	-0.01
23	Phenanthrene	1.292	1.339	-3.6	171#	-0.01
24	Anthracene	1.154	1.281	-11.0	188#	-0.01
25 SURR	Fluoranthene-d10	1.339	1.234	7.8	154#	-0.02
26	Fluoranthene	1.660	1.507	9.2	150	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	123	-0.01
28	Pyrene	1.220	1.468	-20.3	146	-0.02
29 S	Terphenyl-d14	0.819	1.042	-27.2#	155#	-0.02
30	Benzo(a)anthracene	1.312	1.481	-12.9	141	-0.02
31	Chrysene	1.382	1.401	-1.4	121	-0.02
32	Bis(2-ethylhexyl)phthalate	0.524	1.028	-96.2#	263#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.999	1.856	7.2	111	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	118	-0.02
35	Benzo(b)fluoranthene	1.447	1.512	-4.5	121	-0.01
36	Benzo(k)fluoranthene	1.424	1.403	1.5	117	-0.01
37 C	Benzo(a)pyrene	1.327	1.341	-1.1	119	-0.02
38	Dibenzo(a,h)anthracene	1.537	1.446	5.9	109	-0.02
39	Benzo(g,h,i)perylene	1.566	1.475	5.8	109	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN032419\
Data File : BN004952.D
Acq On : 24 Mar 2019 22:53
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 25 04:13:05 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			