

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032422\
 Data File : BN019128.D
 Acq On : 24 Mar 2022 16:20
 Operator : CG/JU
 Sample : N2074-09
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW8-MW01S-20220320

Quant Time: Mar 25 04:29:21 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/25/2022
 Supervised By :Jagrut Upadhyay 03/25/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3709	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	9109	0.400	ng	# 0.00
13) Acenaphthene-d10	14.484	164	5471	0.400	ng	-0.01
19) Phenanthrene-d10	17.236	188	10346	0.400	ng	0.00
29) Chrysene-d12	21.422	240	7828	0.400	ng	# 0.00
35) Perylene-d12	23.790	264	6175	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.413	112	975	0.117	ng	0.00
5) Phenol-d6	7.031	99	374	0.043	ng	0.02
8) Nitrobenzene-d5	9.025	82	1463m	0.223	ng	0.01
11) 2-Methylnaphthalene-d10	12.257	152	4219	0.279	ng	0.00
14) 2,4,6-Tribromophenol	15.984	330	462	0.175	ng	0.00
15) 2-Fluorobiphenyl	13.127	172	7196	0.328	ng	0.00
27) Fluoranthene-d10	19.265	212	13005	0.424	ng	0.00
31) Terphenyl-d14	19.860	244	7573	0.394	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	2811	0.613	ng	# 93
34) Bis(2-ethylhexyl)phtha...	21.324	149	1471	0.092	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032422\
Data File : BN019128.D
Acq On : 24 Mar 2022 16:20
Operator : CG/JU
Sample : N2074-09
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW8-MW01S-20220320

Quant Time: Mar 25 04:29:21 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/25/2022
Supervised By :Jagrut Upadhyay 03/25/2022

