

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032422\
 Data File : BN019136.D
 Acq On : 24 Mar 2022 21:08
 Operator : CG/JU
 Sample : N2075-05
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE109D2-20220322

Quant Time: Mar 25 04:29:53 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/25/2022
 Supervised By :Jagrut Upadhyay 03/25/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	4287	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	11141	0.400	ng	# 0.00
13) Acenaphthene-d10	14.485	164	6749	0.400	ng	-0.01
19) Phenanthrene-d10	17.236	188	13601	0.400	ng	# 0.00
29) Chrysene-d12	21.413	240	10260	0.400	ng	# 0.00
35) Perylene-d12	23.785	264	8032	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	909	0.095	ng	0.00
5) Phenol-d6	7.017	99	444	0.044	ng	0.00
8) Nitrobenzene-d5	9.014	82	1582m	0.197	ng	0.00
11) 2-Methylnaphthalene-d10	12.250	152	4862	0.263	ng	0.00
14) 2,4,6-Tribromophenol	15.985	330	618	0.190	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	6973	0.258	ng	-0.01
27) Fluoranthene-d10	19.261	212	16121	0.399	ng	0.00
31) Terphenyl-d14	19.856	244	8728	0.346	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.247	88	12783	2.413	ng	97
34) Bis(2-ethylhexyl)phtha...	21.324	149	1804	0.086	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032422\
Data File : BN019136.D
Acq On : 24 Mar 2022 21:08
Operator : CG/JU
Sample : N2075-05
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE109D2-20220322

Quant Time: Mar 25 04:29:53 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 03/25/2022
Supervised By : Jagrut Upadhyay 03/25/2022

