

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032422\
 Data File : BN019138.D
 Acq On : 24 Mar 2022 22:20
 Operator : CG/JU
 Sample : N2075-07
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE109D3-20220322

Quant Time: Mar 25 04:30:01 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/25/2022
 Supervised By : Jagrut Upadhyay 03/25/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	4266	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	11144	0.400	ng	# 0.00
13) Acenaphthene-d10	14.485	164	6777	0.400	ng	-0.01
19) Phenanthrene-d10	17.226	188	13503	0.400	ng	-0.01
29) Chrysene-d12	21.422	240	9778	0.400	ng	# 0.00
35) Perylene-d12	23.785	264	7449	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	938	0.098	ng	0.00
5) Phenol-d6	7.017	99	530	0.053	ng	0.00
8) Nitrobenzene-d5	9.014	82	1631m	0.203	ng	0.00
11) 2-Methylnaphthalene-d10	12.250	152	4693	0.254	ng	0.00
14) 2,4,6-Tribromophenol	15.985	330	590	0.180	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	7032	0.259	ng	-0.01
27) Fluoranthene-d10	19.261	212	14839	0.370	ng	0.00
31) Terphenyl-d14	19.856	244	8035	0.335	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.247	88	12801	2.428	ng	98
25) Phenanthrene	17.277	178	953	0.023	ng	95
28) Fluoranthene	19.291	202	1258	0.025	ng	97
34) Bis(2-ethylhexyl)phtha...	21.324	149	1647	0.082	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032422\
Data File : BN019138.D
Acq On : 24 Mar 2022 22:20
Operator : CG/JU
Sample : N2075-07
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE109D3-20220322

Quant Time: Mar 25 04:30:01 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/25/2022
Supervised By :Jagrut Upadhyay 03/25/2022

