

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032422\
 Data File : BN019141.D
 Acq On : 25 Mar 2022 00:08
 Operator : CG/JU
 Sample : N2075-10
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-178S-20220322

Quant Time: Mar 25 04:30:14 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	4344	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	10905	0.400	ng	# 0.00
13) Acenaphthene-d10	14.484	164	6588	0.400	ng	-0.01
19) Phenanthrene-d10	17.236	188	12684	0.400	ng	0.00
29) Chrysene-d12	21.422	240	9936	0.400	ng	# 0.00
35) Perylene-d12	23.782	264	7902	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	925	0.095	ng	0.00
5) Phenol-d6	7.031	99	385	0.038	ng	0.02
8) Nitrobenzene-d5	9.025	82	1739	0.221	ng	0.01
11) 2-Methylnaphthalene-d10	12.253	152	5180	0.286	ng	0.00
14) 2,4,6-Tribromophenol	15.984	330	620	0.195	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	7597	0.288	ng	-0.01
27) Fluoranthene-d10	19.261	212	16635	0.442	ng	0.00
31) Terphenyl-d14	19.855	244	9012	0.369	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.246	88	598	0.111	ng	# 52
34) Bis(2-ethylhexyl)phtha...	21.324	149	1277	0.063	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032422\
Data File : BN019141.D
Acq On : 25 Mar 2022 00:08
Operator : CG/JU
Sample : N2075-10
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-178S-20220322

Quant Time: Mar 25 04:30:14 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 17 13:18:23 2022
Response via : Initial Calibration

